



## Scope of work for annual calibration

Calibration Company quoting for this work should be SANAS accredited for the respective disciplines listed and in the ranges required.

### Scope of Work for annual on-site calibration.

- 1: APPA digital thermometers & probe (x2) @3 points across the range 20 to 40degC.
- 2: Desco barometer
- 3: Inclined manometer (water) (x2) (ranges 0-250mm & 0-200mm [negative, i.e measuring a slight vacuum])
- 4: Weather station; ambient pressure, ambient temperature, ambient RH  
(ranges: 990-1020mbar, 20-30 degC, 35-55%RH)
- 5: CVS CELL 2; pressure and temperature  
(ranges 960-1020mbar, 20-40degC)
- 6: Analogue pressure gauge (tyre pressure), range 1.8-5,0 bar