

Technical Evaluation **Loop Calibrator**. Data sheet to be provided **(If no data sheet is supplied or scoring below 85 points will be regarded as technically non-compliant)**

Nr	Criteria	Weighting (%)	Returnable with the RFQ/Tender Yes / No / Partial compliance	Returnable Datasheet to be provided (Yes/No)	Score
1	440 mA TRUE-RMS DIGITAL MULTIMETER	20	Yes – 20 Partial – 10 No - 0		
2.	1000 VOLT IEC 1010 CAT III STANDARDS	20	Yes – 20 Partial – 10 No - 0		
3.	PERFORMS FREQUENCY MEASUREMENTS UP TO 20 KHZ	20	Yes – 20 Partial – 10 No - 0		
4.	RANGE 0-30 mA	10	Yes – 10 Partial – 5 No - 0		
5.	DRIVE CAPABILITY 500 OHM AT 24mA	10	Yes – 10 Partial –5 No - 0		
6	VOLTAGE MEASUREMENT RANGE 0-1000 VAC OR DC	10	Yes – 10 Partial –5 No - 0		
7	RESISTANCE MEASUREMENT TO 40 MEGAOHM	10	Yes – 10 Partial –5 No - 0		
	Total Score	100			
	Threshold	85			