

Clarification No. 1 (28.05.2026)

TENDER E2947GXMPKR: Service, Calibration, Repairs and Supply of Instrument and Spare Parts for Laboratory Equipment and Consumables at Kriel Power Station for a Period of 5 Years, on As and When Required Basis.

1. Question:

We have noted your requirement for a calibration to be conducted by a company accredited by SANAS.

██████████ is accredited by the American Association for Laboratory Accreditation (A2LA), rather than SANAS, for the ISO 17025-accredited calibrations we provide for mass balances.

Please find our certificate attached, along with an explanation outlining the relationship between SANAS and A2LA.

Kindly let us know whether this is acceptable, or if you would prefer your calibrations to be carried out by a SANAS-accredited supplier.

Response: The certificate is sufficient as long as accreditation is for ISO 17025. However, your company is only accredited for mass only; we also require temperature, time, and humidity.

A reminder that the communication platform is the Eskom e-tender portal and of National Treasury and the responsibility of the tenderer to visit regularly for possible updates. Any revised documents after site meeting, clarification information and minutes will be uploaded as such. Information asper the latest communication will apply