

Scope of work/supply

Supply and delivery of 5 x Zero air generator.

Specification of Product or Goods

Zero air flow: 0-9l/min at 200KPA (std for 8301 P) 8301

Zero air pressure: 200KPA (nominal for portable version)

H2O: Dew-Point -15°C (nominal)*

Power: 220V AC 50Hz

Air scrubbers: Air Clean-Up Canister for removal of: NOx , NO, NO2, O3, SO2, H2S,

TECHNICAL EVALUATION OF TENDERS

The process applied for conducting technical evaluation for the supply and delivery zero air generator (Technical Score 100%)

Specifications of equipment as per procurement request 50%

Guarantee 15%

Documented proof of guarantee provided

Supply and delivery timelines 10%

References letters or purchase order of the previous supply of zero air generator within the period of 5 years 25%

The qualifying threshold for all the suppliers is 80%.

Technical Criteria	Rating
Specifications of equipment as per procurement request and submission of the brochure and data sheet • Meet full specification including submission of the brochure and data sheet- 50% • Submission of brochure and data sheet only – 30 • Submission of brochure only or data sheet only – 10% • Not meeting the specification - 0%	50%
Guarantee Documented proof of guarantee provided • 3-4 years = 15% • 1-2 years = 10% • 0-1 years = 5% • <1 year = 0%	15%
Supply and delivery timelines. Supplier to submit a letter in their company letterhead indicating the delivery time. • Delivery within 3 months 10% • Delivery within 4 months 5% • Delivery withing 5 months -0%	10%
References letters or purchase order of the previous supply of zero air generator within the period of 5 years. • >3 references = 25% • 1 to 2 references = 15% • 0 to 1 references = 10% • 0 references = 0%	25%
Minimum Threshold	80%
TOTAL Score	100%