

INTRODUCTION:

Appointment of a Service Provider for the calibration, repair, and maintenance of alcohol detection machines for PRASA Rail Western Cape Region in accordance with the specification, for a period of 24 months on an as-and-when required basis.

Doc No. & Version:	562455-v1	Page 1 of 5
Author:	Nombasa Matya	
Approved:	Elizabeth Letsoalo	Last Edited:
Template Name & Version:	Template : Blank Document (Portrait) (Office 2003) v5 #169069	

WESTERN CAPE

PRASA RAIL

SPECIFICATION FOR BREATHALYSER CALIBRATION & REPAIRS:

MODEL: MARK X

BACKGROUND:

PRASA (Metrorail) is committed to providing a healthy and safe environment for its employees, as well as for all individuals who may be affected by its operations. To uphold this commitment, PRASA(Metrorail) maintains a zero- tolerant policy toward alcohol and related substances.

Accordingly, breathalyser testing is mandatory for all individuals accessing its buildings and operational areas.

PURPOSE:

To ensure accuracy and reliability of the breathalysers. Basically, the process of calibrating repairing and maintaining the breathalysers is essential to guarantee that alcohol testing machines remain accurate, credible and able to withstand legal scrutiny when required.

SCOPE:

This specification applies to the appointed service provider responsible for calibrating repairing, maintaining the MARK X breathalysers for PRASA (Metrorail) Western Cape for a period of 24 months on an as and when required basis.

SCOPE OF WORK:

The scope of work for this bid entails that the appointed service provider calibrates, repairs, and maintains the breathalysers for PRASA (Metrorail)Western Cape.

Calibration Services: Supplier would perform calibration services for breathalysers or alcohol measuring devices, in accordance with the manufacturer's specifications and relevant standards. This would involve the measurement of the device's accuracy, precision, and linearity.

Maintenance and Repair: Supplier would provide maintenance and repair services for all breathalysers or alcohol measuring devices. This would include fault finding, troubleshooting, replacement of faulty components, and any necessary software updates.

Doc No. & Version:	562455-v1	Page 2 of 5
Author:	Nombasa Matya	
Approved:	Elizabeth Letsoalo	Last Edited:
Template Name & Version:	Template : Blank Document (Portrait) (Office 2003) v5 #169069	

WESTERN CAPE

PRASA RAIL

Traceability: Supplier would ensure that the calibration of the breathalysers or alcohol measuring devices is traceable to national and international standards. This would involve maintaining accurate calibration measurements, procedures, and equipment records.

Certificate of Calibration: Supplier would provide a certificate of calibration for each device calibrated. The certificate would include information such as the device's identification number, calibration date, and calibration results

SPECIFICATION:

Service provider to calibrate, on an as- and when basis, eighty-seven (87) Alco Visor Mark X breathalyser machines per year, and to perform repairs where necessary.

The internal memory of instruments shall be set after each calibration to indicate the correct date of calibration.

The date on the calibration certificate should correspond with the date on the machine.

The date and time of the machine must be checked and re-set if required.

Ensure that the machines are programmed to save all the confirmatory tests conducted in mouthpiece test mode in their correct numerical sequence. The machine to be set to only save positive tests.

A calibration certificate to be issued stating the following:

1. Instrument serial number.
2. Calibration date.
3. Next calibration date and
4. Calibration procedure must be clearly declared with traceability.

Calibration certificate to be signed by a competent and authorized person.

Doc No. & Version:	562455-v1	Page 3 of 5
Author:	Nombasa Matya	
Approved:	Elizabeth Letsoalo	Last Edited:
Template Name & Version:	Template : Blank Document (Portrait) (Office 2003) v5 #169069	

WESTERN CAPE**PRASA RAIL**

CATEGORY:	PICTORIAL MODEL :
Breathalysers/Handheld alcohol testing and measuring device	

Doc No. & Version:	562455-v1	Page 4 of 5
Author:	Nombasa Matya	
Approved:	Elizabeth Letsoalo	Last Edited:
Template Name & Version:	Template : Blank Document (Portrait) (Office 2003) v5 #169069	