

NEC3 Term Service Contract (TSC3)

**Between ESKOM HOLDINGS SOC Ltd
(Reg No. 2002/015527/30)**

**and [Insert at award stage]
(Reg No. _____)**

for the Provision of service & maintenance annually as well as supply of spares, repairs, and consumables, for Metrohm analytical instruments to all Petroleum and Flow Sciences (P&FS) laboratories on an as and when required basis for a period of 36 months (3 years)..

Contents:	No of pages
Part C1 Agreements & Contract Data	[•]
Part C2 Pricing Data	[•]
Part C3 Scope of Work	[•]

CONTRACT No. [Insert at award stage]

PART C1: AGREEMENTS & CONTRACT DATA

Contents:	No of pages
C1.1 Form of Offer and Acceptance	[•]
[to be inserted from Returnable Documents at award stage]	
C1.2a Contract Data provided by the <i>Employer</i>	[•]
C1.2b Contract Data provided by the <i>Contractor</i>	[•]
[to be inserted from Returnable Documents at award stage]	
C1.3 Proforma Guarantees	[•]

C1.1 Form of Offer & Acceptance

Offer

The Employer, identified in the Acceptance signature block, has solicited offers to enter into a contract for the procurement of:

For the Provision of service & maintenance annually as well as supply of spares, repairs, and consumables, for Metrohm analytical instruments to all Petroleum and Flow Sciences (P&FS) laboratories on an as and when required basis for a period of 36 months (3 years).

Options A	The offered total of the Prices exclusive of VAT is	R [●]
	Sub total	R [●]
	Value Added Tax @ 15% is	R [●]
	The offered total of the amount due inclusive of VAT is ¹	R [●]
	(in words) [●]	

Signature(s)			
Name(s)			
Capacity			
For the tenderer:			
Name & signature of witness	(Insert name and address of organisation)	Date	
Tenderer's CIDB registration number:			
CIDB Requirements			

¹ This total is required by the *Employer* for budgeting purposes only. Actual amounts due will be assessed in terms of the *conditions of contract*.

Acceptance

By signing this part of this Form of Offer and Acceptance, the Employer identified below accepts the tenderer's Offer. In consideration thereof, the Employer shall pay the Contractor the amount due in accordance with the *conditions of contract* identified in the Contract Data. Acceptance of the tenderer's Offer shall form an agreement between the Employer and the tenderer upon the terms and conditions contained in this agreement and in the contract that is the subject of this agreement.

The terms of the contract, are contained in:

Part C1	Agreements and Contract Data, (which includes this Form of Offer and Acceptance)
Part C2	Pricing Data
Part C3	Scope of Work: Service Information

and drawings and documents (or parts thereof), which may be incorporated by reference into the above listed Parts.

Deviations from and amendments to the documents listed in the Tender Data and any addenda thereto listed in the Returnable Schedules as well as any changes to the terms of the Offer agreed by the tenderer and the Employer during this process of offer and acceptance, are contained in the Schedule of Deviations attached to and forming part of this Form of Offer and Acceptance. No amendments to or deviations from said documents are valid unless contained in this Schedule.

The tenderer shall within two weeks of receiving a completed copy of this agreement, including the Schedule of Deviations (if any), contact the Employer's agent (whose details are given in the Contract Data) to arrange the delivery of any securities, bonds, guarantees, proof of insurance and any other documentation to be provided in terms of the *conditions of contract* identified in the Contract Data at, or just after, the date this agreement comes into effect. Failure to fulfil any of these obligations in accordance with those terms shall constitute a repudiation of this agreement.

Notwithstanding anything contained herein, this agreement comes into effect on the date when the tenderer receives one fully completed and signed original copy of this document, including the Schedule of Deviations (if any).

Signature(s)			
Name(s)			
Capacity			
for the Employer			
Name & signature of witness	(Insert name and address of organisation)	Date	

Note: If a tenderer wishes to submit alternative tenders, use another copy of this Form of Offer and Acceptance.

INVITATION TO TENDER FOR THE PROVISION OF SERVICE & MAINTENANCE ANNUALLY AS WELL AS SUPPLY OF SPARES, REPAIRS, AND CONSUMABLES, FOR METROHM ANALYTICAL INSTRUMENTS TO ALL PETROLEUM AND FLOW SCIENCES (P&FS) LABORATORIES ON AN AS AND WHEN REQUIRED BASIS FOR A PERIOD OF 36 MONTHS (3 YEARS).

Schedule of Deviations to be completed by the *Employer* prior to contract award

Note:

1. This part of the Offer & Acceptance would not be required if the contract has been developed by negotiation between the Parties and is not the result of a process of competitive tendering.
2. The extent of deviations from the tender documents issued by the Employer prior to the tender closing date is limited to those permitted in terms of the Conditions of Tender.
3. A tenderer's covering letter must not be included in the final contract document. Should any matter in such letter, which constitutes a deviation as aforesaid be the subject of agreement reached during the process of Offer and Acceptance, the outcome of such agreement shall be recorded here and the final draft of the contract documents shall be revised to incorporate the effect of it.

No.	Subject	Details
1	[•]	[•]
2	[•]	[•]
3	[•]	[•]
4	[•]	[•]
5	[•]	[•]
6	[•]	[•]
7	[•]	[•]

By the duly authorised representatives signing this Schedule of Deviations below, the Employer and the tenderer agree to and accept this Schedule of Deviations as the only deviations from and amendments to the documents listed in the Tender Data and any addenda thereto listed in the Tender Schedules, as well as any confirmation, clarification or changes to the terms of the Offer agreed by the tenderer and the Employer during this process of Offer and Acceptance.

It is expressly agreed that no other matter whether in writing, oral communication or implied during the period between the issue of the tender documents and the receipt by the tenderer of a completed signed copy of this Form shall have any meaning or effect in the contract between the parties arising from this Agreement.

	For the tenderer:		For the Employer
Signature			
Name			
Capacity			
On behalf of	(Insert name and address of organisation)		(Insert name and address of organisation)
Name & signature of witness			
Date			

C1.2 TSC3 Contract Data

Part one - Data provided by the *Employer*

[Instructions to the contract compiler: (delete these two notes in the final draft of a contract)]

1. Please read the relevant clauses in the conditions of contract before you enter data. The number of the clause which requires the data is shown in the left hand column for each statement however other clauses may also use the same data.
2. Some TSC3 options are always selected by Eskom Holdings SOC Ltd. The remaining TSC3 options are identified by shading in the left hand column. In the event that the option is not required select and delete the whole row. Where the following symbol is used "■" - data is required to be inserted relevant to the specific option selected.]

Completion of this data in full, according to the Options chosen, is essential to create a complete contract.

Clause	Statement	Data
1	General	
	The <i>conditions of contract</i> are the core clauses and the clauses for main Option:	
		A: Priced contract with price list
	dispute resolution Option	W1: Dispute resolution procedure
	and secondary Options	
		X1: Price adjustment for inflation
		X2 Changes in the law
		X17: Low service damages
		X18: Limitation of liability
		X19: Task Order
		Z: Additional conditions of contract
	of the NEC3 Term Service Contract April 2013 ² (TSC3)	
10.1	The <i>Employer</i> is (name):	Eskom Holdings SOC Ltd (reg no: 2002/015527/30), a state owned company incorporated in terms of the company laws of the Republic of South Africa
	Address	Registered office at Megawatt Park, Maxwell Drive, Sandton, Johannesburg
	Tel No.	011 800 4247

² Available from Engineering Contract Strategies Tel 011 803 3008 Fax 086 539 1902 www.ecs.co.za

	Fax No.	086 6627848
10.1	The <i>Service Manager</i> is (name):	Cherol Tabane
	Address	Megawatt Park, Maxwell Drive, Sandton, Johannesburg
	Tel	011 800 4247
	Fax	086 6627848
	e-mail	tshabac@eskom.co.za
11.2(2)	The Affected Property is	Eskom stations / sites
11.2(13)	The <i>service</i> is	For the Provision of service & maintenance annually as well as supply of spares, repairs, and consumables, for Metrohm analytical instruments to all Petroleum and Flow Sciences (P&FS) laboratories on an as and when required basis for a period of 36 months (3 years).
11.2(14)	The following matters will be included in the Risk Register	Eskom Sites / Stations
11.2(15)	The Service Information is in	Part 3: Scope of Work and all documents and drawings to which it makes reference.
12.2	The <i>law of the contract</i> is the law of	the Republic of South Africa
13.1	The <i>language of this contract</i> is	English
13.3	The <i>period for reply</i> is	[1] weeks
2	The Contractor's main responsibilities	Data required by this section of the core clauses is also provided by the Contractor in Part 2 and terms in italics used in this section are identified elsewhere in this Contract Data
21.1	The <i>Contractor</i> submits a first plan for acceptance within	[1] weeks of the Contract Date
3	Time	
30.1	The <i>starting date</i> is.	01 July 2022
30.1	The <i>service period</i> is	24 Months
4	Testing and defects	There is no reference to Contract Data in this section of the core clauses and terms in italics used in this section are identified elsewhere in this Contract Data
5	Payment	
50.1	The <i>assessment interval</i> is	between the [30] day of each successive month.
51.1	The <i>currency of this contract</i> is the	South African Rand
51.2	The period within which payments are made is	[4] weeks.

51.4	The <i>interest rate</i> is	<p>the publicly quoted prime rate of interest (calculated on a 365 day year) charged by from time to time by the Standard Bank of South Africa Limited (as certified, in the event of any dispute, by any manager of such bank, whose appointment it shall not be necessary to prove) for amounts due in Rands and</p> <p>(ii) the LIBOR rate applicable at the time for amounts due in other currencies. LIBOR is the 6 month London Interbank Offered Rate quoted under the caption "Money Rates" in The Wall Street Journal for the applicable currency or if no rate is quoted for the currency in question then the rate for United States Dollars, and if no such rate appears in The Wall Street Journal then the rate as quoted by the Reuters Monitor Money Rates Service (or such service as may replace the Reuters Monitor Money Rates Service) on the due date for the payment in question, adjusted <i>mutatis mutandis</i> every 6 months thereafter (and as certified, in the event of any dispute, by any manager employed in the foreign exchange department of The Standard Bank of South Africa Limited, whose appointment it shall not be necessary to prove.</p>
6	Compensation events	There is no reference to Contract Data in this section of the core clauses and terms in italics used in this section are identified elsewhere in this Contract Data
7	Use of Equipment Plant and Materials	There is no reference to Contract Data in this section of the core clauses and terms in italics used in this section are identified elsewhere in this Contract Data
8	Risks and insurance	
80.1	These are additional <i>Employer's</i> risks	
9	Termination	There is no reference to Contract Data in this section of the core clauses and terms in italics used in this section are identified elsewhere in this Contract Data.
10	Data for main Option clause	
A	Priced contract with price list	Is applicable
20.5	The <i>Contractor</i> prepares forecasts of the final total of the Prices for the whole of the <i>service</i> at intervals no longer than	[1] week.
11	Data for Option W1	
W1.1	The <i>Adjudicator</i>	the person selected from the ICE-SA Division (or its successor body) of the South African Institution of Civil Engineering Panel of Adjudicators by the Party intending to refer a

		dispute to him. (see www.ice-sa.org.za). If the Parties do not agree on an Adjudicator the Adjudicator will be appointed by the Arbitration Foundation of Southern Africa (AFSA).		
W1.2(3)	The <i>Adjudicator nominating body</i> is:	the Chairman of ICE-SA a joint Division of the South African Institution of Civil Engineering and the Institution of Civil Engineers (London) (see www.ice-sa.org.za) or its successor body.		
W1.4(2)	The <i>tribunal</i> is:	arbitration		
W1.4(5)	The <i>arbitration procedure</i> is	the latest edition of Rules for the Conduct of Arbitrations published by The Association of Arbitrators (Southern Africa) or its successor body.		
	The place where arbitration is to be held is	[•] South Africa		
	The person or organisation who will choose an arbitrator - if the Parties cannot agree a choice or - if the arbitration procedure does not state who selects an arbitrator, is	the Chairman for the time being or his nominee of the Association of Arbitrators (Southern Africa) or its successor body.		
12	Data for secondary Option clauses			
X1	Price adjustment for inflation			
X1.1	The <i>base date</i> for indices is			
	The proportions used to calculate the Price Adjustment Factor are:	proportion	linked to index for	Index prepared by
		0.	[•]	[•]
X2	Changes in the law	There is no reference to Contract Data in this Option and terms in italics are identified elsewhere in this Contract Data.		
X3	Multiple currencies	Not Applicable		
X4	Parent company guarantee	Not Applicable		
X12	Partnering	Not Applicable		
X13	Performance bond	Not Applicable		
X17	Low service damages			
X17.1	The <i>service level table</i> is in	[•]		
X18	Limitation of liability			
X18.1	The <i>Contractor's</i> liability to the <i>Employer</i> for indirect or consequential loss is limited to	R0.0 (zero Rand)		
X18.5	The <i>end of liability date</i> is	[6] months after the end of the <i>service period</i> .		
X19	Task Order			

X19.5	The <i>Contractor</i> submits a Task Order programme to the <i>Service Manager</i> within	[3] days of receiving the Task Order
X20	Key Performance Indicators (not used when Option X12 applies)	Not Applicable
Z	The <i>additional conditions of contract</i> are	Z1 to Z14 always apply.
Z1	Cession delegation and assignment	
Z1.1	The <i>Contractor</i> does not cede, delegate or assign any of its rights or obligations to any person without the written consent of the <i>Employer</i> .	
Z1.2	Notwithstanding the above, the <i>Employer</i> may on written notice to the <i>Contractor</i> cede and delegate its rights and obligations under this contract to any of its subsidiaries or any of its present divisions or operations which may be converted into separate legal entities as a result of the restructuring of the Electricity Supply Industry.	
Z2	Joint ventures	
Z2.1	If the <i>Contractor</i> constitutes a joint venture, consortium or other unincorporated grouping of two or more persons or organisations then these persons or organisations are deemed to be jointly and severally liable to the <i>Employer</i> for the performance of this contract.	
Z2.2	Unless already notified to the <i>Employer</i> , the persons or organisations notify the <i>Service Manager</i> within two weeks of the Contract Date of the key person who has the authority to bind the <i>Contractor</i> on their behalf.	
Z2.3	The <i>Contractor</i> does not alter the composition of the joint venture, consortium or other unincorporated grouping of two or more persons without the consent of the <i>Employer</i> having been given to the <i>Contractor</i> in writing.	
Z3	Change of Broad Based Black Economic Empowerment (B-BBEE) status	
Z3.1	Where a change in the <i>Contractor's</i> legal status, ownership or any other change to his business composition or business dealings results in a change to the <i>Contractor's</i> B-BBEE status, the <i>Contractor</i> notifies the <i>Employer</i> within seven days of the change.	
Z3.2	The <i>Contractor</i> is required to submit an updated verification certificate and necessary supporting documentation confirming the change in his B-BBEE status to the <i>Service Manager</i> within thirty days of the notification or as otherwise instructed by the <i>Service Manager</i> .	
Z3.3	Where, as a result, the <i>Contractor's</i> B-BBEE status has decreased since the Contract Date the <i>Employer</i> may either re-negotiate this contract or alternatively, terminate the <i>Contractor's</i> obligation to Provide the Service.	
Z3.4	Failure by the <i>Contractor</i> to notify the <i>Employer</i> of a change in its B-BBEE status may constitute a reason for termination. If the <i>Employer</i> terminates in terms of this clause, the procedures on termination are P1, P2 and P4 as stated in clause 92, and the amount due is A1 and A3 as stated in clause 93.	
Z4	Confidentiality	
Z4.1	The <i>Contractor</i> does not disclose or make any information arising from or in connection with	

	this contract available to Others. This undertaking does not, however, apply to information which at the time of disclosure or thereafter, without default on the part of the <i>Contractor</i> , enters the public domain or to information which was already in the possession of the <i>Contractor</i> at the time of disclosure (evidenced by written records in existence at that time). Should the <i>Contractor</i> disclose information to Others in terms of clause 25.1, the <i>Contractor</i> ensures that the provisions of this clause are complied with by the recipient.
Z4.2	If the <i>Contractor</i> is uncertain about whether any such information is confidential, it is to be regarded as such until notified otherwise by the <i>Service Manager</i> .
Z4.3	In the event that the <i>Contractor</i> is, at any time, required by law to disclose any such information which is required to be kept confidential, the <i>Contractor</i> , to the extent permitted by law prior to disclosure, notifies the <i>Employer</i> so that an appropriate protection order and/or any other action can be taken if possible, prior to any disclosure. In the event that such protective order is not, or cannot, be obtained, then the <i>Contractor</i> may disclose that portion of the information which it is required to be disclosed by law and uses reasonable efforts to obtain assurances that confidential treatment will be afforded to the information so disclosed.
Z4.4	The taking of images (whether photographs, video footage or otherwise) of the Affected Property or any portion thereof, in the course of Providing the Service and after the end of the <i>service period</i> , requires the prior written consent of the <i>Service Manager</i> . All rights in and to all such images vests exclusively in the <i>Employer</i> .
Z4.5	The <i>Contractor</i> ensures that all his subcontractors abide by the undertakings in this clause.
Z5	Waiver and estoppel: Add to core clause 12.3:
Z5.1	Any extension, concession, waiver or relaxation of any action stated in this contract by the Parties, the <i>Service Manager</i> or the <i>Adjudicator</i> does not constitute a waiver of rights, and does not give rise to an estoppel unless the Parties agree otherwise and confirm such agreement in writing.
Z6	Health, safety and the environment: Add to core clause 27.4
Z6.1	<p>The <i>Contractor</i> undertakes to take all reasonable precautions to maintain the health and safety of persons in and about the execution of the <i>service</i>. Without limitation the <i>Contractor</i>:</p> <ul style="list-style-type: none"> • accepts that the <i>Employer</i> may appoint him as the "Principal Contractor" (as defined and provided for under the Construction Regulations 2014 (promulgated under the Occupational Health & Safety Act 85 of 1993) ("the Construction Regulations") for the Affected Property; • warrants that the total of the Prices as at the Contract Date includes a sufficient amount for proper compliance with the Construction Regulations, all applicable health & safety laws and regulations and the health and safety rules, guidelines and procedures provided for in this contract and generally for the proper maintenance of health & safety in and about the execution of the <i>service</i>; and • undertakes, in and about the execution of the <i>service</i>, to comply with the Construction Regulations and with all applicable health & safety laws and regulations and rules, guidelines and procedures otherwise provided for under this contract and ensures that his Subcontractors, employees and others under the <i>Contractor's</i> direction and control, likewise observe and comply with the foregoing.
Z6.2	The <i>Contractor</i> , in and about the execution of the <i>service</i> , complies with all applicable environmental laws and regulations and rules, guidelines and procedures otherwise provided for under this contract and ensures that his Subcontractors, employees and others under the <i>Contractor's</i> direction and control, likewise observe and comply with the foregoing.

Z7	Provision of a Tax Invoice and interest. Add to core clause 51
Z7.1	Within one week of receiving a payment certificate from the <i>Service Manager</i> in terms of core clause 51.1, the <i>Contractor</i> provides the <i>Employer</i> with a tax invoice in accordance with the <i>Employer's</i> procedures stated in the Service Information, showing the amount due for payment equal to that stated in the payment certificate.
Z7.2	If the <i>Contractor</i> does not provide a tax invoice in the form and by the time required by this contract, the time by when the <i>Employer</i> is to make a payment is extended by a period equal in time to the delayed submission of the correct tax invoice. Interest due by the <i>Employer</i> in terms of core clause 51.2 is then calculated from the delayed date by when payment is to be made.
Z7.3	The <i>Contractor</i> (if registered in South Africa in terms of the companies Act) is required to comply with the requirements of the Value Added Tax Act, no 89 of 1991 (as amended) and to include the <i>Employer's</i> VAT number 4740101508 on each invoice he submits for payment.
Z8	Notifying compensation events
Z8.1	Delete the last paragraph of core clause 61.3 and replace with: If the <i>Contractor</i> does not notify a compensation event within eight weeks of becoming aware of the event, he is not entitled to a change in the Prices.
Z9	<i>Employer's</i> limitation of liability
Z9.1	The <i>Employer's</i> liability to the <i>Contractor</i> for the <i>Contractor's</i> indirect or consequential loss is limited to R0.00 (zero Rand)
Z9.2	The <i>Contractor's</i> entitlement under the indemnity in 82.1 is provided for in 60.1(12) and the <i>Employer's</i> liability under the indemnity is limited to compensation as provided for in core clause 63 and X19.11 if Option X19 Task Order applies to this contract.
Z10	Termination: Add to core clause 91.1, at the second main bullet point, fourth sub-bullet point, after the words "against it":
Z10.1	or had a business rescue order granted against it.
Z11	Ethics
For the purposes of this Z-clause, the following definitions apply:	
Affected Party	means, as the context requires, any party, irrespective of whether it is the <i>Contractor</i> or a third party, such party's employees, agents, or Subcontractors or Subcontractor's employees, or any one or more of all of these parties' relatives or friends,
Coercive Action	means to harm or threaten to harm, directly or indirectly, an Affected Party or the property of an Affected Party, or to otherwise influence or attempt to influence an Affected Party to act unlawfully or illegally,
Collusive Action	means where two or more parties co-operate to achieve an unlawful or illegal purpose, including to influence an Affected Party to act unlawfully or illegally,
Committing Party	means, as the context requires, the <i>Contractor</i> , or any member thereof in the case of a joint venture, or its employees, agents, or Subcontractors or the Subcontractor's employees,

Corrupt Action	means the offering, giving, taking, or soliciting, directly or indirectly, of a good or service to unlawfully or illegally influence the actions of an Affected Party,
Fraudulent Action	means any unlawfully or illegally intentional act or omission that misleads, or attempts to mislead, an Affected Party, in order to obtain a financial or other benefit or to avoid an obligation or incurring an obligation,
Obstructive Action	means a Committing Party unlawfully or illegally destroying, falsifying, altering or concealing information or making false statements to materially impede an investigation into allegations of Prohibited Action, and
Prohibited Action	means any one or more of a Coercive Action, Collusive Action Corrupt Action, Fraudulent Action or Obstructive Action.
Z11.1	A Committing Party may not take any Prohibited Action during the course of the procurement of this contract or in execution thereof.
Z11.2	The <i>Employer</i> may terminate the <i>Contractor's</i> obligation to Provide the Services if a Committing Party has taken such Prohibited Action and the <i>Contractor</i> did not take timely and appropriate action to prevent or remedy the situation, without limiting any other rights or remedies the <i>Employer</i> has. It is not required that the Committing Party had to have been found guilty, in court or in any other similar process, of such Prohibited Action before the <i>Employer</i> can terminate the <i>Contractor's</i> obligation to Provide the Services for this reason.
Z11.3	If the <i>Employer</i> terminates the <i>Contractor's</i> obligation to Provide the Services for this reason, the amounts due on termination are those intended in core clauses 92.1 and 92.2.
Z11.4	A Committing Party co-operates fully with any investigation pursuant to alleged Prohibited Action. Where the <i>Employer</i> does not have a contractual bond with the Committing Party, the <i>Contractor</i> ensures that the Committing Party co-operates fully with an investigation.

Z12 Insurance**Z 12 .1 Replace core clause 83 with the following:**

Insurance cover	83									
	83.1	When requested by a Party, the other Party provides certificates from his insurer or broker stating that the insurances required by this contract are in force.								
	83.2	<p>The <i>Contractor</i> provides the insurances stated in the Insurance Table A from the <i>starting date</i> until the earlier of Completion and the date of the termination certificate.</p> <table><tr><th colspan="2">INSURANCE TABLE A</th></tr><tr><th>Insurance against</th><th>Minimum amount of cover or minimum limit of indemnity</th></tr><tr><td>Loss of or damage caused by the <i>Contractor</i> to the <i>Employer's</i> property</td><td><p>The replacement cost where not covered by the <i>Employer's</i> insurance.</p><p>The <i>Employer's</i> policy deductible as at Contract Date, where covered by the <i>Employer's</i> insurance.</p></td></tr><tr><td>Loss of or damage to Plant and Materials</td><td>The replacement cost where not covered by the <i>Employer's</i> insurance.</td></tr></table>	INSURANCE TABLE A		Insurance against	Minimum amount of cover or minimum limit of indemnity	Loss of or damage caused by the <i>Contractor</i> to the <i>Employer's</i> property	<p>The replacement cost where not covered by the <i>Employer's</i> insurance.</p> <p>The <i>Employer's</i> policy deductible as at Contract Date, where covered by the <i>Employer's</i> insurance.</p>	Loss of or damage to Plant and Materials	The replacement cost where not covered by the <i>Employer's</i> insurance.
INSURANCE TABLE A										
Insurance against	Minimum amount of cover or minimum limit of indemnity									
Loss of or damage caused by the <i>Contractor</i> to the <i>Employer's</i> property	<p>The replacement cost where not covered by the <i>Employer's</i> insurance.</p> <p>The <i>Employer's</i> policy deductible as at Contract Date, where covered by the <i>Employer's</i> insurance.</p>									
Loss of or damage to Plant and Materials	The replacement cost where not covered by the <i>Employer's</i> insurance.									

			The <i>Employer's</i> policy deductible as at Contract Date, where covered by the <i>Employer's</i> insurance.
		Loss of or damage to Equipment	The replacement cost where not covered by the <i>Employer's</i> insurance. The <i>Employer's</i> policy deductible as at Contract Date, where covered by the <i>Employer's</i> insurance.
		The <i>Contractor's</i> liability for loss of or damage to property (except the <i>Employer's</i> property, Plant and Materials and Equipment) and liability for bodily injury to or death of a person (not an employee of the <i>Contractor</i>) arising from or in connection with the <i>Contractor's</i> Providing the Service	<u>Loss of or damage to property</u> The replacement cost <u>Bodily injury to or death of a person</u> The amount required by the applicable law.
		Liability for death of or bodily injury to employees of the <i>Contractor</i> arising out of and in the course of their employment in connection with this contract	The amount required by the applicable law

Z 12.2 Replace core clause 86 with the following:

Insurance by the Employer	86		
	86.1	The Employer provides the insurances stated in the Insurance Table B	
		INSURANCE TABLE B	
		Insurance against or name of policy	Minimum amount of cover or minimum limit of indemnity
		Assets All Risk	Per the insurance policy document
		Contract Works insurance	Per the insurance policy document
		Environmental Liability	Per the insurance policy document
		General and Public Liability	Per the insurance policy document
		Transportation (Marine)	Per the insurance policy document

		Motor Fleet and Mobile Plant	Per the insurance policy document	
		Terrorism	Per the insurance policy document	
		Cyber Liability	Per the insurance policy document	
		Nuclear Material Damage and Business Interruption	Per the insurance policy document	
		Nuclear Material Damage Terrorism	Per the insurance policy document	
Z13	Nuclear Liability			
Z13.1	The <i>Employer</i> is the operator of the Koeberg Nuclear Power Station (KNPS), a nuclear installation, as designated by the National Nuclear Regulator of the Republic of South Africa, and is the holder of a nuclear licence in respect of the KNPS.			
Z13.2	The <i>Employer</i> is solely responsible for and indemnifies the <i>Contractor</i> or any other person against any and all liabilities which the <i>Contractor</i> or any person may incur arising out of or resulting from nuclear damage, as defined in Act 44 of 1999, save to the extent that any liabilities are incurred due to the unlawful intent of the <i>Contractor</i> or any other person or the presence of the <i>Contractor</i> or that person or any property of the <i>Contractor</i> or such person at or in the KNPS or on the KNPS site, without the permission of the <i>Employer</i> or of a person acting on behalf of the <i>Employer</i> .			
Z13.3	Subject to clause Z13.4 below, the <i>Employer</i> waives all rights of recourse, arising from the aforesaid, save to the extent that any claims arise or liability is incurred due or attributable to the unlawful intent of the <i>Contractor</i> or any other person, or the presence of the <i>Contractor</i> or that person or any property of the <i>Contractor</i> or such person at or in the KNPS or on the KNPS site, without the permission of the <i>Employer</i> or of a person acting on behalf of the <i>Employer</i> .			
Z13.4	The <i>Employer</i> does not waive its rights provided for in section 30 (7) of Act 44 of 1999, or any replacement section dealing with the same subject matter.			
Z13.5	The protection afforded by the provisions hereof shall be in effect until the KNPS is decommissioned.			
Z14	Asbestos			
For the purposes of this Z-clause, the following definitions apply:				
AAIA	means approved asbestos inspection authority.			
ACM	means asbestos containing materials.			
AL	means action level, i.e. a level of 50% of the OEL, i.e. 0.1 regulated asbestos fibres per ml of air measured over a 4 hour period. The value at which proactive actions is required in order to control asbestos exposure to prevent exceeding the OEL.			
Ambient Air	means breathable air in area of work with specific reference to breathing zone, which is defined to be a virtual area within a radius of approximately 30cm from the nose inlet.			
Compliance Monitoring	means compliance sampling used to assess whether or not the personal exposure of workers to regulated asbestos fibres is in compliance with the Standard's requirements for safe processing, handling, storing, disposal and phase-out of asbestos and asbestos containing material, equipment and articles.			

OEL	means occupational exposure limit.
Parallel Measurements	means measurements performed in parallel, yet separately, to existing measurements to verify validity of results.
Safe Levels	means airborne asbestos exposure levels conforming to the Standard's requirements for safe processing, handling, storing, disposal and phase-out of asbestos and asbestos containing material, equipment and articles.
Standard	means the <i>Employer's</i> Asbestos Standard 32-303: Requirements for Safe Processing, Handling, Storing, Disposal and Phase-out of Asbestos and Asbestos Containing Material, Equipment and Articles.
SANAS	means the South African National Accreditation System.
TWA	means the average exposure, within a given workplace, to airborne asbestos fibres, normalised to the baseline of a 4 hour continuous period, also applicable to short term exposures, i.e. 10-minute TWA.
Z14.1	The <i>Employer</i> ensures that the Ambient Air in the area where the <i>Contractor</i> will Provide the Services conforms to the acceptable prescribed South African standard for asbestos, as per the regulations published in GNR 155 of 10 February 2002, under the Occupational Health and Safety Act, 1993 (Act 85 of 1993) ("Asbestos Regulations"). The OEL for asbestos is 0.2 regulated asbestos fibres per millilitre of air as a 4-hour TWA, averaged over any continuous period of four hours, and the short term exposure limit of 0.6 regulated asbestos fibres per millilitre of air as a 10-minute TWA, averaged over any 10 minutes, measured in accordance with HSG248 and monitored according to HSG173 and OESSM.
Z14.2	Upon written request by the <i>Contractor</i> , the <i>Employer</i> certifies that these conditions prevail. All measurements and reporting are effected by an independent, competent, and certified occupational hygiene inspection body, i.e. a SANAS accredited and Department of Employment and Labour approved AAIA. The <i>Contractor</i> may perform Parallel Measurements and related control measures at the <i>Contractor's</i> expense. For the purposes of compliance the results generated from Parallel Measurements are evaluated only against South African statutory limits as detailed in clause Z14.1. Control measures conform to the requirements stipulated in the AAIA-approved asbestos work plan.
Z14.3	The <i>Employer</i> manages asbestos and ACM according to the Standard.
Z14.4	In the event that any asbestos is identified while Providing the Services, a risk assessment is conducted and if so required, with reference to possible exposure to an airborne concentration of above the AL for asbestos, immediate control measures are implemented and relevant air monitoring conducted in order to declare the area safe.
Z14.5	The <i>Contractor's</i> personnel are entitled to stop working and leave the contaminated area forthwith until such time that the area of concern is declared safe by either Compliance Monitoring or an AAIA approved control measure intervention, for example, per the emergency asbestos work plan, if applicable.
Z14.6	The <i>Contractor</i> continues to Provide the Services, without additional control measures presented, on presentation of Safe Levels. The contractually agreed dates to Provide the Services, including the Completion Date, are adjusted accordingly. The contractually agreed dates are extended by the notification periods required by regulations 3 and 21 of the Asbestos Regulations, 2001.
Z14.7	Any removal and disposal of asbestos, asbestos containing materials and waste, is done by a registered asbestos contractor, instructed by the <i>Employer</i> at the <i>Employer's</i> expense, and conducted in line with South African legislation.

C1.2 Contract Data

Part two - Data provided by the Contractor

Notes to a tendering contractor:

1. Please read both the both the NEC3 Term Service Contract April 2013 and the relevant parts of its Guidance Notes (TSC3-GN)³ in order to understand the implications of this Data which the tenderer is required to complete.
2. The number of the clause which requires the data is shown in the left hand column for each statement however other clauses may also use the same data.
3. Where a form field like this [] appears, data is required to be inserted relevant to the option selected. Click on the form field **once** and type in the data. Otherwise complete by hand and in ink.

Completion of the data in full, according to Options chosen, is essential to create a complete contract.

Clause	Statement	Data
10.1	The <i>Contractor</i> is (Name):	
	Address	
	Tel No.	
	Fax No.	
11.2(8)	The <i>direct fee percentage</i> is	%
	The <i>subcontracted fee percentage</i> is	%
11.2(14)	The following matters will be included in the Risk Register	
11.2(15)	The Service Information for the <i>Contractor's</i> plan is in:	
21.1	The plan identified in the Contract Data is contained in:	
24.1	The key people are:	
	1 Name:	
	Job:	
	Responsibilities:	
	Qualifications:	
	Experience:	
	2 Name:	
	Job	
	Responsibilities:	
	Qualifications:	

³ Available from Engineering Contract Strategies Tel 011 803 3008 Fax 086 5391902 or www.ecs.co.za

INVITATION TO TENDER FOR THE PROVISION OF SERVICE & MAINTENANCE ANNUALLY AS WELL AS SUPPLY OF SPARES, REPAIRS, AND CONSUMABLES, FOR METROHM ANALYTICAL INSTRUMENTS TO ALL PETROLEUM AND FLOW SCIENCES (P&FS) LABORATORIES ON AN AS AND WHEN REQUIRED BASIS FOR A PERIOD OF 36 MONTHS (3 YEARS).

	Experience:	
		CV's (and further key person's data including CVs) are in .
A	Priced contract with price list	
11.2(12)	The <i>price list</i> is in	
11.2(19)	The tendered total of the Prices is	R

PART 2: PRICING DATA

TSC3 Option A

Document reference	Title	No of pages
C2.1	Pricing assumptions: Option A	2
C2.2	The <i>price list</i>	[•]

C2.1 Pricing assumptions: Option A

How work is priced and assessed for payment

Clause 11 in NEC3 Term Service Contract (TSC3) core clauses and Option A states:

Identified and defined terms	11	
	11.2	(12) The Price List is the <i>price list</i> unless later changed in accordance with this contract.
		(17) The Price for Services Provided to Date is the total of <ul style="list-style-type: none"> the Price for each lump sum item in the Price List which the <i>Contractor</i> has completed and where a quantity is stated for an item in the Price List, an amount calculated by multiplying the quantity which the <i>Contractor</i> has completed by the rate.
		(19) The Prices are the amounts stated in the Price column of the Price List. Where a quantity is stated for an item in the Price List, the Price is calculated by multiplying the quantity by the rate.

This confirms that Option A is a priced contract where the Prices are derived from a list of items of service which can be priced as lump sums or as expected quantities of service multiplied by a rate or a mix of both.

Function of the Price List

Clause 54.1 in Option A states: "Information in the Price List is not Service Information". This confirms that instructions to do work or how it is to be done are not included in the Price List but in the Service Information. This is further confirmed by Clause 20.1 which states, "The *Contractor* Provides the Service in accordance with the Service Information". Hence the *Contractor* does **not** Provide the Service in accordance with the Price List. The Price List is only a pricing document.

Link to the *Contractor's* plan

Clause 21.4 states "The *Contractor* provides information which shows how each item description on the Price List relates to the operations on each plan which he submits for acceptance". Hence when compiling the *price list*, the tendering contractor needs to develop his first clause 21.2 plan in such a way that operations shown on it can be priced in the *price list* and result in a satisfactory cash flow in terms of clause 11.2(17).

Preparing the *price list*

Before preparing the *price list*, both the *Employer* and tendering contractors should read the TSC3 Guidance Notes pages 14 and 15. In an Option A contract, either Party may have entered items into the *price list* either as a process of offer and acceptance (tendering) or by negotiation depending on the nature of the *service* to be provided. Alternatively the *Employer*, in his Instructions to Tenderers or in a Tender Schedule, may have listed some items that he requires the *Contractor* to include in the *price list* to be prepared and priced by him.

It is assumed that in preparing or finalising the *price list* the *Contractor*:

- Has taken account of the guidance given in the TSC3 Guidance Notes relevant to Option A;
- Understands the function of the Price List and how work is priced and paid for;
- Is aware of the need to link operations shown in his plan to items shown in the Price List;
- Has listed and priced items in the *price list* which are inclusive of everything necessary and incidental to Providing the Service in accordance with the Service Information, as it was at the time of tender, as well as correct any Defects not caused by an *Employer's* risk;
- Has priced work he decides not to show as a separate item within the Prices or rates of other listed items in order to fulfil the obligation to complete the *service* for the tendered total of the Prices.
- Understands there is no adjustment to items priced as lump sums if the amount, or quantity, of work within that item later turns out to be different to that which the *Contractor* estimated at time of tender. The only basis for a change to the (lump sum) Prices is as a result of a compensation event.

Format of the *price list*

(From the example given in an Appendix within the TSC3 Guidance Notes)

Entries in the first four columns in the *price list* in section C2.2 are made either by the *Employer* or the tendering contractor.

If the *Contractor* is to be paid an amount for the item which is not adjusted if the quantity of work in the item changes, the tendering contractor enters the amount in the Price column only, the Unit, Expected Quantity and Rate columns being left blank.

If the *Contractor* is to be paid an amount for an item of work which is the rate for the work multiplied by the quantity completed, the tendering contractor enters the rate which is then multiplied by the Expected Quantity to produce the Price, which is also entered.

If the *Contractor* is to be paid a Price for an item proportional to the length of time for which a service is provided, a unit of time is stated in the Unit column and the expected length of time (as a quantity of the stated units of time) is stated in the Expected Quantity column.

C2.2 the *price list*

A. Service and maintenance per laboratory annually

Laboratory	Instrument Model	Serial number	Frequency	Units Prices	Total Prices
Seshego	Metrohm 899 Coulometer	1899001011102	3		
Bloemfontein	Metrohm 899 Coulometer	1899001002251	3		
Bloemfontein	Metrohm 899 Coulometer	1899001011116	3		
Brackenfell	Metrohm 899 Coulometer	1899001004173	3		
Witbank	Metrohm 831 Coulometer	1831001034239	3		
Witbank	Metrohm 684 KF Coulometer	IE9/297	3		
Pietermaritzburg	Metrohm 831 Coulometer	1831001016196	3		
Pietermaritzburg	Metrohm 899 Coulometer	1899001005201	3		
Rotek	Metrohm 831 Coulometer	1831001040389	3		
Seshego	Metrohm 785 DMP Titrimo	1785001011110	3		
Rosherville	Metrohm 899 Coulometer	1899001005206	3		

B. Spares and consumables as and when required

Item Number	Item	Description	Dimensions	Estimated quantities in a 3yr period	Units Prices	Total Prices
60229010	Solvotrode Easy clean	Metrohm glass electrode for non-aqueous titrations. pH Scale: 0-14	Type Length:125mm Diameter :12mm Measurement range: 0-14 Cable Length:1m Application : Acid/Base titration and pH measurements Specification : Glass pH probe with Standard:	3		
62320000	Electrolyte solution TEABr (tetraethylammonium bromide in ethylene glycol)	Electrolyte solution used to fill glass electrode used for Total acid number application.	Type : Trade name: Tetraethylammonium bromide in ethylene glycol Physical form : Liquid Colour : Clear/colourless Composition : Application : Non-aqueous Acid base titration Container : Plastic Container capacity :250mL Purity : Analytical grade	18		
63026220	Complete 20ml Exchange unit	Plastic unit that holds titrant solution and is mounted on the metrohm titrimo series of titrators.	Type : Metrohm type accessory Material : glass & PTFE Capacity :20mL Power range220-240v Power source : Powered by Metrohm titrimos	1		
61542020	Flat Stop Cock	Two channel devices for filling of burette and dispensing of the solution during titration.	Type : Stop cock Valve type : Switch type Valve size : Pressure rating : Dimensions : Material : Material grade : PTCFE Material specification : Application : Specification : Standard :	1		

INVITATION TO TENDER FOR THE PROVISION OF SERVICE & MAINTENANCE ANNUALLY AS WELL AS SUPPLY OF SPARES, REPAIRS, AND CONSUMABLES, FOR METROHM ANALYTICAL INSTRUMENTS TO ALL PETROLEUM AND FLOW SCIENCES (P&FS) LABORATORIES ON AN AS AND WHEN REQUIRED BASIS FOR A PERIOD OF 36 MONTHS (3 YEARS).

61543060	TIP/M6 thread tubing	Burette tip for dosing out titrant solution into titration beaker/cell.	Size: 400mm length 2mm inner diameter Connection: treaded ends Material: fluorinated ethylene propylene Grade: Not specified	1		
61805010	FEP tubing	Siphon tubing for delivery of titrant solution.	Size: 400mm length 2mm inner diameter Connection: treaded ends Material: fluorinated ethylene propylene Grade: Not specified	1		
61805080	FEP tubing	Siphon tubing for delivery of titrant solution.	Size: 130mm length 2mm inner diameter Connection: treaded ends Material: fluorinated ethylene propylene Grade: Not specified	1		
61805100	FEP tubing	Siphon tubing for delivery of titrant solution.	Size: 400mm length 2mm inner diameter Connection: treaded ends Material: fluorinated ethylene propylene Grade: Not specified	1		
2.802.0020	Overhead Rod Stirrer for 804	Device used to homogenise solutions	Type: Metrohm Stirringpositions: Clock-wise Power source: Metrohm sample changers/Titrinos/Titrados Application: Titration & sample measurement	1		
60344100	Gen Electrode with diaphragm WOC for 756/831	Generator electrode for coulometric Karl Fisher titration application for water determination.	Type: generator electrode Material: Glass Dimensions: length 108mm Diameter 24mm Temperature range: 0 0 100 DEGREE pHrange: 1-14 Application: Coulometric	66		
62104020	Electrode Cable for 60344100- Standard Metrohm probe cable	Standard metrohm probe/Electrode cable.	Type: Metrohm plug-in head G Diameter: Not specified Length: 1m Material: Metallic ends and rubber insulation and rubber Specification: Fits into Metrohm F sockets	33		
62104120	Electrode cable for 60344100	Standard metrohm probe/Electrode cable for generator electrodes.	Type: Metrohm plug-in head G Diameter: Not specified Length: 1m Material: Metallic ends and rubber Specification: Fits into Metrohm F sockets	20		
61464320	KF titration Vessel	Metrohm Coulometric titration assembly to slot in electrodes and seals.	Type: Metrohmtype Insidediameter: 50mm Outsidediameter: 50mm Material: Borosilicateglass Height: 114mm Capacity rating: 250mL	66		
60341100	Indicator Electrode	Indicator probe with two platinum pins, used in	Type: Generator electrode Material: Glass Dimensions: length 108mm Diameter 24mm Temperature range: 0-to-100-	66		

INVITATION TO TENDER FOR THE PROVISION OF SERVICE & MAINTENANCE ANNUALLY AS WELL AS SUPPLY OF SPARES, REPAIRS, AND CONSUMABLES, FOR METROHM ANALYTICAL INSTRUMENTS TO ALL PETROLEUM AND FLOW SCIENCES (P&FS) LABORATORIES ON AN AS AND WHEN REQUIRED BASIS FOR A PERIOD OF 36 MONTHS (3 YEARS).

		the Karl Fisher application for water determinatio n.	degree pH range: 1 -14 Application: Coulometric Karl Fischer			
T&S						
Item	Description			Total Prices	Units Prices	
1	Labour rate per hour (certified technician)					
2	Travel rate per km					
3	Travel time per hour					
4	Delivery charge					
5	Accommodation					

Tenders must specify labour, travel and subsistence or any other costs separately and must be based on laboratories listed above.

The total of the Prices Excluding VAT

TSC3 COVER PAGES

1. SCOPE OF WORK

For the provision of service & maintenance annually as well as supply of spares, repairs, and consumables, for Metrohm analytical instruments to all Petroleum and Flow Sciences (P&FS) laboratories on an as and when required basis for a period of 36 months (3 years).

Provide service, maintenance and calibration on Metrohm analytical instruments for all P&FS Laboratories annually, see list of instruments on table 1 below. Repairs on these instruments will be done as and when require.

Table 1 List of Instruments

Laboratory	Instrument Model	Serial number	Frequency
Seshego	Metrohm 899 Coulometer	1899001011102	3
Bloemfontein	Metrohm 899 Coulometer	1899001002251	3
Bloemfontein	Metrohm 899 Coulometer	1899001011116	3
Brackenfell	Metrohm 899 Coulometer	1899001004173	3
Witbank	Metrohm 831 Coulometer	1831001034239	3
Witbank	Metrohm 684 KF Coulometer	IE9/297	3
Pietermaritzburg	Metrohm 831 Coulometer	1831001016196	3
Pietermaritzburg	Metrohm 899 Coulometer	1899001005201	3
Rotek	Metrohm 831 Coulometer	1831001040389	3
Seshego	Metrohm 785 DMP Titrino	1785001011110	3
Rosherville	Metrohm 899 Coulometer	1899001005206	3
Klerksdorp	Metrohm 899 Coulometer	1899001002250	3

Requirements:

- Supplier to provide a report after each service/repairs done in each laboratory.
- Service/repair agent must respond to the request within 48hrs of purchase order being issued.
- Calibration certificate must contain minimum information as follows:
 - ✓ Name of the requester,
 - ✓ unique number for the certificate,
 - ✓ address of requester and supplier,
 - ✓ supplier company name,
 - ✓ accreditation number (if accredited),
 - ✓ Serial number of instrument being calibrated,
 - ✓ Summary of status,
 - ✓ Results,
 - ✓ Uncertainty of measurements,
 - ✓ Environmental conditions at the time of calibration,
 - ✓ Date calibrated,
 - ✓ Name and signature of the technician calibrated the instrument
- Supplier to provide proof that the instrument used to calibrate Metrohm equipment is calibrated, and calibration is traceable to NIST.
- If the supplier is not accredited according to ISO 17025 or any other equivalent standard, the following **documents as per SANAS TR 25 section 3.3 in house calibration must be provided:**
 - ✓ A documented and validated calibrated procedure or method;
 - ✓ A calibration certificate or report, or some other suitable method of recording the calibration and measurement data;
 - ✓ Appropriate educational and training records to demonstrate the competence of the personnel performing the calibration;

- ✓ Copies of certificates and reports to demonstrate traceability to the SI system of units or RMs;
- ✓ All reference standards and measuring instruments shall be calibrated at appropriate intervals, the facility shall have and apply a documented procedure for establishing these calibration intervals;
- ✓ A documented procedure for the estimation of the measurement uncertainty, which shall be applied for at least each type of calibration. The uncertainty of measurement shall be taken into account when making a statement of compliance;
- ✓ Records to demonstrate the assurance of the calibration results i.e. participation in Inter-Laboratory Comparisons or Proficiency Testing;
- ✓ Documented requirements for environmental conditions, and records where applicable; and
- ✓ Evidence of internal audits of all activities that could influence the quality of calibration results.

2. Supply of spares and consumables as and when required for a period of three (3) years.

Table 2 List of spares and consumables

Item Number	Item	Description	Dimensions	Estimated quantities in a 3yr period
60229010	Solvotrode Easy clean	Metrohm glass electrode for non-aqueous titrations. pH Scale: 0-14	Type Length:125mm Diameter :12mm Measurement range : 0-14 Cable Length :1m Application : Acid/Base titration and pH measurements Specification : Glass pH probe with Standard:	3
62320000	Electrolyte solution TEABr (tetraethylammonium bromide in ethylene glycol)	Electrolyte solution used to fill glass electrode used for Total acid number application.	Type : Trade name : Tetraethylammonium bromide in ethylene glycol Physical form : Liquid Colour : Clear/colourless Composition : Application : Non-aqueous Acid base titration Container : Plastic Container capacity :250mL Purity : Analytical grade	18
63026220	Complete 20ml Exchange unit	Plastic unit that holds titrant solution and is mounted on the metrohm titrino series of titrators.	Type : Metrohm type accessory Material : glass & PTFE Capacity :20mL Power range :220-240v Power source : Powered by Metrohm titrinos	1
61542020	Flat Stop Cock	Two channel devices for filling of burette and dispensing of the solution during titration.	Type : Stop cock Valve type : Switch type Valve size : Pressure rating : Dimensions : Material : Material grade : PTCFE Material specification : Application : Specification : Standard :	1
61543060	TIP/M6 thread tubing	Burette tip for dosing out titrant solution into titration beaker/cell.	Size : 400mm length 2mm inner diameter Connection : treaded ends Material : fluorinated ethylene propylene	1

			Grade: Not specified	
61805010	FEP tubing	Siphon tubing for delivery of titrant solution.	Size: 400mm length 2mm inner diameter Connection: treaded ends Material: fluorinated ethylene propylene Grade: Not specified	1
61805080	FEP tubing	Siphon tubing for delivery of titrant solution.	Size: 130mm length 2mm inner diameter Connection: treaded ends Material: fluorinated ethylene propylene Grade: Not specified	1
61805100	FEP tubing	Siphon tubing for delivery of titrant solution.	Size: 400mm length 2mm inner diameter Connection: treaded ends Material: fluorinated ethylene propylene Grade: Not specified	1
2.802.0020	Overhead Rod Stirrer for 804	Device used to homogenise solutions	Type: Metrohm Stirringpositions: Clock-wise Power source: Metrohm sample changers/Titrinos/Titrandos Application: Titration & sample measurement	1
60344100	Gen Electrode with diaphragm WOC for 756/831	Generator electrode for coulometric Karl Fisher titration application for water determination.	Type: generator electrode Material: Glass Dimensions: length 108mm Diameter 24mm Temperature range: 0 0 100 DEGREE pHrange: 1-14 Application: Coulometric	66
62104020	Electrode Cable for 60344100-Standard Metrohm probe cable	Standard metrohm probe/Electrode cable.	Type: Metrohm plug-in head G Diameter: Not specified Length: 1m Material: Metallic ends and rubber insulation and rubber Specification: Fits into Metrohm F sockets	33
62104120	Electrode cable for 60344100	Standard metrohm probe/Electrode cable for generator electrodes.	Type: Metrohm plug-in head G Diameter: Notspecified Length: 1m Material: Metallic ends and rubber Specification: Fits into Metrohm F sockets	20
61464320	KF titration Vessel	Metrohm Coulometric titration assembly to slot in electrodes and seals.	Type: Metrohmtype Inside diameter: 50mm Outsidediameter: 50mm Material: Borosilicate glass Height: 114mm Capacity rating: 250mL	66

60341100	Indicator Electrode	Indicator probe with two platinum pins, used in the Karl Fisher application for water determination.	Type: Generator electrode Material: Glass Dimensions: length 108mm Diameter 24mm Temperature range: 0-to-100-degree pH range: 1 -14 Application: Coulometric Karl Fischer	66
----------	---------------------	--	---	----

2.1 Requirements:

- Spares/consumables must be for Metrohm brand of equipment.

2.2 Other requirements

2.2.1 Supplier to include in the quotation:

- ✓ Labour rate per hour (certified technician)
- ✓ Travel rate per km
- ✓ Travel time per hour
- ✓ Delivery charge
- ✓ Accommodation

3 Instrument service/calibration/repairs/spares and consumables to be delivered to the following addresses:

3.1 Requirements:

- Supplier must be able to provide service to all the P&FS Laboratories.
- Service/repair agent must respond to the request within 48hrs of purchase order issue.

Annexure C: List of delivery addresses

Item	Eskom Sites	Address
1	Rosherville ERIC	P&FS laboratory Eskom Research & Innovation Centre Oil lab - Block B - Reception Lower Germiston Road Rosherville 2022
2	Rotek	P&FS laboratory Rotek Oil lab (Next to Site & Switchgear Workshop) 1 Lower Germiston Road Rosherville
3	Brackenfell	P&FS laboratory Eskom Oil Lab Eskom Road Brackenfell Cape Town 7535
4	Bloemfontein	P&FS laboratory C/o Transnet Depot Oil lab – Eskom Loco Street East End Bloemfontein 9001
5	Witbank	P&FS laboratory Eskom Oil lab Eskom Park (next to Electrical Workshop) Watermeyer Street Witbank 1035
6	Seshego Polokwane	- Eskom P&FS laboratory 53 Moletsi Street Seshego Polokwane
7	Klerksdorp	P&FS laboratory Oil lab - Eskom 22 Goueweg Townsland Klerksdorp 2571
8	Port Elizabeth	P&FS laboratory c/o Eastern Cape Training Centre Oil lab Workshop 2 Spondo Road Struandale Port Elizabeth 6001
9	Pietermaritzburg	P&FS laboratory Oil lab - Eskom1 Portland Road Mkondeni Pietermaritzburg 3021

C3.1: EMPLOYER'S SERVICE INFORMATION

Contents

Part 3: Scope of Work vi

C3.1: Employer's service Information x

1	Description of the <i>service</i>	xii
1.1	Executive overview	xii
1.2	<i>Employer's</i> requirements for the <i>service</i>	xii
1.3	Interpretation and terminology	xii
2	Management strategy and start up.	xii
2.1	The <i>Contractor's</i> plan for the <i>service</i>	xii
2.2	Management meetings	xiii
2.3	<i>Contractor's</i> management, supervision and key people	xiii
2.4	Provision of bonds and guarantees	xiv
2.5	Documentation control	xiv
2.6	Invoicing and payment	xiv
2.7	Contract change management	xiv
2.8	Records of Defined Cost to be kept by the <i>Contractor</i>	xiv
2.9	Insurance provided by the <i>Employer</i>	xiv
2.10	Training workshops and technology transfer	xv
2.11	Design and supply of Equipment	xv
2.12	Things provided at the end of the <i>service period</i> for the <i>Employer's</i> use	xv
2.12.1	Equipment	xv
2.12.2	Information and other things	xv
2.13	Management of work done by Task Order	xv
3	Health and safety, the environment and quality assurance	xvii
3.1	Health and safety risk management	xvii
3.2	Environmental constraints and management	xvii
3.3	Quality assurance requirements	xvii
4	Procurement	xvii
4.1	People	xvii
4.1.1	Minimum requirements of people employed	xviii
4.1.2	BBBEE and preferencing scheme	xviii
4.1.3	Accelerated Shared Growth Initiative – South Africa (ASGI-SA)	xviii
4.2	Subcontracting	xviii
4.2.1	Preferred subcontractors	xviii
4.2.2	Subcontract documentation, and assessment of subcontract tenders	xviii
4.2.3	Limitations on subcontracting	xviii
4.2.4	Attendance on subcontractors	xviii
4.3	Plant and Materials	xviii
4.3.1	Specifications	xviii
4.3.2	Correction of defects	xix
4.3.3	<i>Contractor's</i> procurement of Plant and Materials	xix
4.3.4	Tests and inspections before delivery	xix
4.3.5	Plant & Materials provided "free issue" by the <i>Employer</i>	xix
4.3.6	Cataloguing requirements	xix
5	Working on the Affected Property	xx
5.1	<i>Employer's</i> site entry and security control, permits, and site regulations	xx
5.2	People restrictions, hours of work, conduct and records	xx
5.3	Health and safety facilities on the Affected Property	xx
5.4	Environmental controls, fauna & flora	xx

5.5	Cooperating with and obtaining acceptance of Others.....	xx
5.6	Records of <i>Contractor's</i> Equipment.....	xx
5.7	Equipment provided by the <i>Employer</i>	xx
5.8	Site services and facilities.....	xxi
5.8.1	Provided by the <i>Employer</i>	xxi
5.8.2	Provided by the <i>Contractor</i>	xxi
5.9	Control of noise, dust, water and waste	xxi
5.10	Hook ups to existing works	xxi
5.11	Tests and inspections	xxi
5.11.1	Description of tests and inspections	xxi
5.11.2	Materials facilities and samples for tests and inspections	xxi
6	List of drawings	xxi
6.1	Drawings issued by the <i>Employer</i>	xxi
	Description of the <i>service</i>	

1.1 Executive overview

Put yourself in the position of the *Contractor's* senior management who need a high level overview of what is involved – the scope of work - so that they can decide when tendering whether they have the resources and capability to undertake the work. Include a description of the different services and disciplines of work involved within the *service* and the location where most of the work will be carried out.

1.2 Employer's requirements for the service

Either

Describe in detail what the *Employer* requires the *Contractor* to do and how he is to do it if the *Contractor* is mainly providing labour and tools to carry out the *Employer's* requirements

Or

Provide the *Employer's* operating philosophy / user requirement specification (URS) / performance specification giving deliverables and constraints for the *service* from which the *Contractor* is to plan in detail how he is to achieve the required deliverables.

Reference could be made to an Annexure for a detailed classification of services or to the Price List in the case of Option A or C and if the Price List descriptions are complete.

1.3 Interpretation and terminology

If required include here definitions additional to those used in the *conditions of contract* which are required only for the purpose of making the Service Information easier to draft and read. Also list abbreviations used and provide a full interpretation of each one, for example:

The following abbreviations are used in this Service Information:

Abbreviation	Meaning given to the abbreviation
OBL	Outside battery limits

2 Management strategy and start up.

2.1 The Contractor's plan for the service

In the TSC3 the *Contractor's* plan is his “design” for performing the *service* throughout the *service period*. Section 2 of the *conditions of contract* describes what the *Contractor* is to show in his plan both in the core clauses and some additional requirements in each of the main Options.

The extent of the *Contractor's* plan will depend on whether the *Contractor* is required to develop a plan in accordance with the *Employer's* broad outline of the *service* or whether the *Employer* has provided a plan for the *Contractor* to follow. Read the TSC3 Guidance Notes pages 21 and 22 for more information on the *Contractor's* plan.

Use this section to describe any particulars which must be taken into account by the *Contractor* in developing his plan as required by clause 21.2. For example information about the order and timing or method of carrying out particular items of work.

List technical reporting and scheduling requirements which are to be incorporated into the *Contractor's* plan.

2.2 Management meetings

The *conditions of contract* (e.g. Clause 16.2) and other sections of the Service Information (e.g. safety risk management) may require that a meeting shall be held. However the intention of all NEC contracts is that the Parties and their agents use the techniques of partnering to manage the contract by holding meetings designed to pro actively and jointly manage the administration of the contract with the objective of minimising the adverse effects of risks and surprises for both Parties.

Depending on the size and complexity of the *service*, it is probably beneficial for the *Service Manager* to hold a weekly risk register meeting (Clause 16.2). This could be used to discuss safety, compensation events, subcontracting, overall co-ordination and other matters of a general nature. Separate meetings for specialist activities such as planning and activities of a technical nature may also be warranted.

Describe here the general meetings and their purpose. Provide particulars of approximate times, days, location, and attendance requirements, stipulating that attendees shall have the necessary delegated authority to make decisions in respect of matters discussed at such meetings.

The following text could be used as a model for this section:

Regular meetings of a general nature may be convened and chaired by the *Supply Manager* as follows:

Title and purpose	Approximate time & interval	Location	Attendance by:
Risk register and compensation events	Weekly on _____ at _____		
Overall contract progress and feedback	Monthly on _____ at _____		<i>Employer, Contractor</i> and _____

Meetings of a specialist nature may be convened as specified elsewhere in this Service Information or if not so specified by persons and at times and locations to suit the Parties, the nature and the progress of the *service*. Records of these meetings shall be submitted to the *Service Manager* by the person convening the meeting within five days of the meeting.

All meetings shall be recorded using minutes or a register prepared and circulated by the person who convened the meeting. Such minutes or register shall not be used for the purpose of confirming actions or instructions under the contract as these shall be done separately by the person identified in the *conditions of contract* to carry out such actions or instructions.

2.3 Contractor's management, supervision and key people

State any additional constraining requirements on *Contractor's* supervision and key people that are not already stated in other sections such as for Health and Safety. This section could be used to solicit an organogramme from the *Contractor* showing his people and their lines of authority / communication. This would be essential if the *Contractor* is a Joint Venture.

2.4 Provision of bonds and guarantees

The form in which a bond or guarantee required by the *conditions of contract* (if any) is to be provided by the *Contractor* is given in Part 1 Agreements and Contract Data, document C1.3, Sureties.

The *Employer* may withhold payment of amounts due to the *Contractor* until the bond or guarantee required in terms of this contract has been received and accepted by the person notified to the *Contractor* by the *Service Manager* to receive and accept such bond or guarantee. Such withholding of payment due to the *Contractor* does not affect the *Employer's* right to termination stated in this contract.

2.5 Documentation control

Specify how documentation will be identified with an alpha numeric which indicates source, recipient, communication number etc. Provide details of any particular format or other constraints; for example that all contractual communications will be in the form of properly compiled letters or forms attached to e mails and not as a message in the e mail itself. State any particular routing requirements but note from TSC3 who issues what to whom.

2.6 Invoicing and payment

The Z clauses make reference to invoicing procedures stated here in this Service Information. Also include a list of information which is to be shown on an invoice.

Within one week of receiving a payment certificate from the *Service Manager* in terms of core clause 51.1, the *Contractor* provides the *Employer* with a tax invoice showing the amount due for payment equal to that stated in the *Service Manager's* payment certificate.

The *Contractor* shall address the tax invoice to

and include on each invoice the following information:

- Name and address of the *Contractor* and the *Service Manager*;
- The contract number and title;
- *Contractor's* VAT registration number;
- The *Employer's* VAT registration number 4740101508;
- Description of service provided for each item invoiced based on the Price List;
- Total amount invoiced excluding VAT, the VAT and the invoiced amount including VAT;
- (add other as required)

Add procedures for invoice submission and payment (e. g. electronic payment instructions)

2.7 Contract change management

This section is intended to deal with any additional requirements to the compensation event clauses in section 6 of the core clauses; such as the use of standard forms. Not the same thing as documentation control.

2.8 Records of Defined Cost to be kept by the *Contractor*

If Option C or E applies first read clause 52.2 and then state whether the *Contractor* is required to keep any other records. Include any other constraint which may be required in regard to format and filing of the records, and whether access for the *Service Manager* shall be provided in hard copy or electronically.

Could delete if Option A applies unless the *Employer* requires some form of control over the *Contractor's* record keeping for the purpose of compensation event management.

2.9 Insurance provided by the *Employer*

First read TSC3 Core Clause 86.1 and then add anything necessary for the management of insurance related issues such as a cross reference to where procedures for making claims can be found. Also provide contact details for persons capable of being able to answer any insurance related queries the *Contractor* may have, as well as to whom the information required by Marine Insurance (if any) may be addressed.

2.10 Training workshops and technology transfer

Describe type and frequency of any on job training workshops, as well as any obligation for technology transfer being included as part of the *service* or at the end of the *service period*.

2.11 Design and supply of Equipment

On some complex services (e. g. nuclear) it could be in the Parties best interests that some details of the design of Equipment are shared with the *Service Manager*, not necessarily for his acceptance but as an assurance that the Equipment will be able to allow the *Contractor* to Provide the Service efficiently and without delay. For example specialised handling Equipment for a particular maintenance operation. Clause 23.1 is always available to the *Service Manager* if this paragraph is not used.

Also the *Employer* may wish to exercise constraints or include witness and hold points during manufacture, assembly or delivery of such Equipment. Include these constraints here.

Draft in such a way that there is no doubt that the liability for such design supply and use of the Equipment remains with the *Contractor*.

2.12 Things provided at the end of the *service period* for the *Employer's* use

2.12.1 Equipment

The *Employer* may wish to use some of the Equipment used by the *Contractor* after the end of the *service period*. Clause 70.2 requires that details of such requirement be stated in the Service Information. Complete here or if not applicable retain the heading and state 'None'.

2.12.2 Information and other things

Clause 70.2 requires that information and other things which the *Contractor* is to provide at the end of the *service period* be stated in the Service Information.

2.13 Management of work done by Task Order

Only use this heading if Option X19 applies to this contract.

In some cases all work may be done in terms of Task Orders in which case it may be logical to move this section closer to the start of this part 2 of the Service Information. In some cases only parts of the *service* may require to be handled by Task Order, for example a major repair which has become necessary during a continuous maintenance service contract.

Please read Option X19 before drafting requirements here as much of the procedure for the administration of Task Orders is already provided in X19, for example X19.2 specifies what a task Order should include

A Task Order format could be provided in an Annexure to this Service Information.

Many considerations can apply to Task Orders, such as availability of resources, arrangements for emergency work, Task Order reporting (work carried out and service results), assessment of additional Prices for *service* not included in the Price List etc.

INVITATION TO TENDER FOR THE PROVISION OF SERVICE & MAINTENANCE ANNUALLY AS WELL AS SUPPLY OF SPARES, REPAIRS, AND CONSUMABLES, FOR METROHM ANALYTICAL INSTRUMENTS TO ALL PETROLEUM AND FLOW SCIENCES (P&FS) LABORATORIES ON AN AS AND WHEN REQUIRED BASIS FOR A PERIOD OF 36 MONTHS (3 YEARS).

Clause X19.6 requires information which should be included on a Task Order programme.
Further requirements for Task Orders include things to be provided by the *Employer* under a Task Order and the conditions under which the *Employer* or Others are to work.

3 Health and safety, the environment and quality assurance

3.1 Health and safety risk management

In addition to the requirements of the laws governing health and safety, Eskom may have some additional requirements particular to the *service* and the Affected Property for this contract. The text below provides for these being attached as an Annexure to this Service Information. PLEASE ALSO READ CORE CLAUSE 27.4 TOGETHER WITH Z7 IN THE ADDITIONAL CONDITIONS OF CONTRACT TO MAKE SURE THAT WHATSOEVER IS INCLUDED IN THE ANNEXURE FOLLOWS ON FROM THOSE CLAUSES.

The Divisional/Regional Safety Risk Manager or his representative having jurisdiction over the *service* must provide the relevant safety, health and environmental (SHE) criteria for incorporation into this Service Information. The SHE specification / scope must be signed off by the Divisional/Regional Safety Risk Manager or his representative confirming that the applicable safety criteria have been taken into account.

The Commodity Manager / Buyer must refer the tender to the Divisional/Regional Safety Risk Manager or his representative in order to evaluate against enquiry-specific safety criteria.

The Divisional Safety Risk Managers who will be responsible for the allocation of resources to assist P&SCM with the above processes are as follows:

- Generation: Roley McIntyre
- Transmission: Tony Patterson
- Distribution: Alex Stramrood
- Enterprises: Jace Naidoo
- Corporate: Kerser Pather

The *Contractor* shall comply with the health and safety requirements contained in Annexure _____ to this Service Information.

3.2 Environmental constraints and management

Describe or cross refer to environmental constraints applicable to the *Contractor's* plan and his activities on the Affected Property and how they should be managed. Include here or cross refer to an Annexure to the Service Information.

The *Contractor* shall comply with the environmental criteria and constraints stated in Annexure _____

3.3 Quality assurance requirements

Specify minimum requirements for the *Contractor's* Quality Plan and Work Procedures or provide the *Employer's* Quality Plan if that is to be used. Make sure witness and hold points are identified generally and describe any particular requirements for QA outside the Affected Property. Indicate how the *Contractor's* QA documentation is to be submitted for acceptance and any conditions that need to be imposed relating to acceptance. State whether ISO compliance is a condition and if so which ISO standard shall apply.

4 Procurement

There is a cross reference from the core clause 11.2(6) definition of Disallowed Cost to the Service Information regarding procurement procedures. This part of the Service Information MUST include any such procedures to be able to administer Disallowed Cost.

4.1 People

4.1.1 Minimum requirements of people employed

Specify any constraints relating to people employed to Provide the Service; for example permits for foreigners, training (other than H & S), use of labour from designated areas and industrial relations.

4.1.2 BBBEE and preferencing scheme

Specify constraints which *Contractor* must comply with after contract award in regard to any Broad Based Black Economic Empowerment (B-BBEE) or preferencing scheme measures.

4.1.3 Accelerated Shared Growth Initiative – South Africa (ASGI-SA)

If the ASGI-SA requirements are to be included in this contract specify constraints which *Contractor* must comply with after contract award in regard to any ASGI-SA requirements. The ASGI-SA Compliance Schedule completed in the returnable tender schedules is reproduced here. If ASGI-SA does not apply, delete this paragraph.

The *Contractor* complies with and fulfils the *Contractor's* obligations in respect of the Accelerated and Shared Growth Initiative - South Africa in accordance with and as provided for in the *Contractor's* ASGI-SA Compliance Schedule stated below

[Insert the agreed ASGI-SA Compliance Schedule here]

The *Contractor* shall keep accurate records and provide the *Service Manager* with reports on the *Contractor's* actual delivery against the above stated ASGI-SA criteria. [Elaborate on access to and format of records and frequency of submission etc.]

The *Contractor's* failure to comply with his ASGI-SA obligations constitutes substantial failure on the part of the *Contractor* to comply with his obligations under this contract.

4.2 Subcontracting

4.2.1 Preferred subcontractors

TSC3 does not make use of nominated subcontracting, but the *Employer* may list which subcontractors or suppliers the *Contractor* is required to enter into subcontracts with. This is usually only required where specialist services need to be obtained from a particular supplier or group of suppliers in order to comply with operational standards.

4.2.2 Subcontract documentation, and assessment of subcontract tenders

Specify any constraints on how the *Contractor* is to prepare subcontract documentation, whether use of the NEC system is compulsory or not (compulsory is recommended) and how subcontract tenders are to be issued, received, assessed (using a joint report?) and awarded.

4.2.3 Limitations on subcontracting

The *Employer* may require that the *Contractor* must subcontract certain specialised work, or that the *Contractor* shall not subcontract more than a specified proportion of the whole of the contract.

4.2.4 Attendance on subcontractors

State requirements for attendance on Subcontractors, if any

4.3 Plant and Materials

4.3.1 Specifications

Plant and Materials are defined as items intended to be included in the Affected Property. This will refer to replacement of worn or defective parts, routine replacement as part of regular preventative maintenance and supply of spare parts. Quality is usually designed in or specified in the technical specifications. However to cover circumstances where quality may not be prescribed, this sub-paragraph could also be used to state an overarching default requirement – fitness for purpose etc.

Either specify here or provide a list of the applicable specifications and attach them as Annexure or state where they can be obtained from.

4.3.2 Correction of defects

State any constraints when dealing with defective Plant and Materials such as how repairs are carried out - can the item be fixed up or must it be replaced by a new one.

4.3.3 Contractor's procurement of Plant and Materials

Specify any constraints on how the *Contractor* is to order, codify, expedite, freight, import, transport to the Affected Property and any other requirements for delivery and storage before installation. The *Employer* may require warranties from suppliers to be in favour of the *Employer* and not just to the *Contractor*. The *Employer* may also need schedules of vendor data for his own use after the end of the *service period*.

4.3.4 Tests and inspections before delivery

Core Clause 41.1 makes reference to the Service Information stating which Plant and Materials are to be inspected and tested before delivery. Specify any requirements particularly if such tests and inspections are to be carried out by agents of the *Employer* overseas.

4.3.5 Plant & Materials provided “free issue” by the Employer

List any Plant and Materials which are to be provided by the *Employer*.

State arrangements for collection by *Contractor* or delivery by others on behalf of the *Employer*, off loading, inspection, storage, care custody and control, return of unused Plant and Materials, etc. Always include a statement to the effect that ‘all other Plant and Materials are to be provided by the *Contractor*’.

4.3.6 Cataloguing requirements by the Contractor

State whether cataloguing is applicable, if it is, reference the requirements for cataloguing that need to be satisfied by the *Contractor* (consult Procurement Instruction Number 1 of 2018 – Incorporating Cataloguing into the Procurement Environment, Unique Identifier 240-1289988974).

5 Working on the Affected Property

This part of the Service Information addresses constraints, facilities, services and rules applicable to the *Contractor* whilst he is doing work on the Affected Property.

5.1 *Employer's* site entry and security control, permits, and site regulations

Sites such as Koeberg Nuclear Power Station have very strict entrance requirements which tendering contractors need to allow for in their prices, and the *Contractor* has to comply with. State these or similar requirements here.

In addition to the above there may be other restrictions once on the site, plus rules relating to roads, walkways and the provision of barricades

5.2 People restrictions, hours of work, conduct and records

Restrictions and hours of work may apply on some sites. It is very important that the *Contractor* keeps records of his people working on the Affected Property, including those of his Subcontractors. State that the *Service Manager* shall have access to them at any time. These records may be needed when assessing compensation events.

5.3 Health and safety facilities on the Affected Property

Section 3 deals with contractual H & S requirements in addition to those of the OHSA Act. This section allows the *Employer* to state what measures are to be taken on the Affected Property by describing where First Aid facilities provided by the *Employer* are located and any other emergency arrangements. Do not use if already addressed in 2.3.

5.4 Environmental controls, fauna & flora

This sub-paragraph may not be required in a service contract or if these matters are dealt with in the general environmental requirements referred to in section 3 above.

5.5 Cooperating with and obtaining acceptance of Others

This sub-paragraph could be used to deal with two issues.

- 1) The cross reference from core clause 25.1 about cooperation generally as well as details about Others with whom the *Contractor* may be required to share the Affected Property. See clause 11.2(9) for the definition of Others.
- 2) Requirements for liaison with and acceptance from statutory authorities or inspection agencies.

5.6 Records of *Contractor's* Equipment

This sub-paragraph is intended to address how records are to be kept of Equipment on Site including whether it is owned or hired. Include any constraints about scaffolding, rigs, heavy lifts and cranes, including removal from the Affected Property.

5.7 Equipment provided by the *Employer*

Provide details of equipment (e.g. overhead cranes) made available for use by the employer and set out conditions relating thereto.

5.8 Site services and facilities

5.8.1 Provided by the *Employer*

This is a mandatory cross reference form clause 25.2 in TSC3. State what the *Employer* will provide in the way of power, water, waste disposal, telecomms, ablutions, fire protection and lighting (etc) on the Affected Property. Give hook up locations and any constraints on how the hook up is to be done. Always conclude by stating that the *Contractor* shall provide everything else necessary for Providing the Service.

5.8.2 Provided by the *Contractor*

Describe what the *Contractor* is to provide in the way of accommodation, laboratories, storage, vehicles and office equipment for the *Service Manager* and any restrictions or minimum requirements concerning the *Contractor's* own facilities. Also state what happens to these facilities upon completion of the contract.

5.9 Control of noise, dust, water and waste

State requirements, if any.

5.10 Hook ups to existing works

State any constraints

5.11 Tests and inspections

5.11.1 Description of tests and inspections

Describe the tests and inspections to be carried out by the *Contractor* and the *Service Manager* and others [40.1].

5.11.2 Materials facilities and samples for tests and inspections

State what materials facilities and samples for tests and inspections the *Contractor* and the *Employer* are to provide, per core clause 40.2.

6 List of drawings

6.1 Drawings issued by the *Employer*

This is the list of drawings issued by the *Employer* at or before the Contract Date and which apply to this contract.

Drawing number	Revision	Title

ESKOM HOLDINGS SOC LTD

CONTRACT NUMBER _____

INVITATION TO TENDER FOR THE PROVISION OF SERVICE & MAINTENANCE ANNUALLY AS WELL AS SUPPLY OF SPARES, REPAIRS, AND CONSUMABLES, FOR METROHM ANALYTICAL INSTRUMENTS TO ALL PETROLEUM AND FLOW SCIENCES (P&FS) LABORATORIES ON AN AS AND WHEN REQUIRED BASIS FOR A PERIOD OF 36 MONTHS (3 YEARS).

--	--	--