



Signed-off Scope Of Work

Title: The supply and delivery of laboratory instruments

Document type:

Signed-off scope of work


Signed-off by:

Signature:  p.p.

Name: Zinhle Mbatha

Senior Supervisor Chemistry


Approved by:

Signature 

Name: Michael Mukwevho

Supervising Manager

Accepted by:

Signature 

Name: Nthabiseng Ntoampe

Senior Chemist Chemistry

CONTENTS

	Page No
1. PURPOSE	3
2. SCOPE OF WORK	3
3. MAN-POWER PLAN	4

SIGNED-OFF SCOPE OF WORK: CONTRACT	Template	
	Revision: 0	Page 3 of 3

1. PURPOSE

Laboratory instruments are the heart of chemical services, as they are required to monitor the processes for the water treatment plants as well as cycle chemistry to ensure that the requirements of Long Term Plant health are achieved and to meet the surveillance requirements set in the chemistry standards. According to the laboratory instruments maintenance plan all the laboratory instruments have a life span of 10 years and the instruments that being procured have exceeded their life span and are therefore due for replacement.

2. SCOPE OF WORK

2.1 Executive overview

The supply and delivery of the following laboratory instrumentation:

Inductively Coupled Plasma OES (ICP)

Ion Chromatography Dual

2.2 Deficiencies and Modifications

- The *Contractor* must inform the *Employer* in writing of the deficiencies in their area of maintenance responsibility.
- The *Contractor* must inform the *Employer* about any factor known to them that may affect plant health performance.

2.3 Clean up

The *Contractor* is responsible for cleaning up the site, free of rubble. After completion of the works and clean up, the *Service Manager* shall be called to inspect the site and ensure that it is in a presentable state. The *Employer's* Environmental Section will do ad-hoc inspections on your work site, as well as a final inspection on the dumping site.

2. PPE.

PPE to be provided by the contractor as per Eskom Regulations

3. MAN-POWER PLAN

1 service technician required

CONTROLLED DISCLOSURE