

	SCOPE OF WORK	DUVHA POWER STATION
---	----------------------	----------------------------

Title: **SERVICE, MAINTENANCE, CALIBRATION, SUPPLY OF SPARES AND CONSUMABLE OF COAL LABORATORY INSTRUMENTS AS AND WHEN REQUIRED FOR 5 YEARS**

Document Identifier: **559-59852433**

Alternative Reference Number: **N/A**

Area of Applicability: **Engineering**


Functional Area: **Chemical Services**

Revision: **1**

Total Pages: **14**

Next Review Date: **N/A**

Disclosure Classification: **CONTROLLED DISCLOSURE**

Compiled by	Functional Responsibility	Authorised by
	pp 	Pp 
Bavuseni Mogola Senior Supervisor Tech Chemistry	Cecil Mngqibisa Senior Chemist	Lefumo Maswanganye Manager Chemical Services
Date: 2026-08-28	Date: 2026-09-17	Date: 2026-09-21

CONTENTS	PAGE
1. INTRODUCTION	3
2. SUPPORTING CLAUSES	3
2.1 Scope.....	3
2.1.1 Purpose	3
2.1.2 Applicability.....	3
2.1.3 Effective date.....	3
2.2 Normative/Informative References	3
2.2.1 Normative	3
2.2.2 Informative	3
2.3 Definitions	4
2.4 Abbreviations	4
2.5 Roles and Responsibilities	4
2.5.1 Chemical Services.....	4
2.5.2 Service Provider	4
2.6 Process for Monitoring.....	4
2.7 Related/Supporting Documents	4
3. REQUIREMENTS FOR WORKS	5
3.1 Certificate of Analysis	5
3.2 Safety Data Sheets.....	5
3.3 Quality of Chemicals and Standards	5
3.4 Product Shelf-Life Requirements	5
3.5 Product Catalogue Submission	5
3.6 Traceability Requirements.....	5
3.7 Quality Assurance and Site Audits	6
3.8 Instrument Service and Technical Support	6
3.9 Technical Competency Requirements	11
3.10 Calibration Certification Requirements.....	11
3.11 Service Reporting	12
3.12 Response Time for Equipment Failures	12
3.13 Availability of Spare Parts.....	12
3.14 Prohibition of Chemical Decanting	12
3.15 Delivery Documentation Requirements.....	12
3.16 Transport Vehicle and Driver Requirements	12
4. ACCEPTANCE	14
5. REVISIONS	14
6. DEVELOPMENT TEAM.....	14
7. ACKNOWLEDGEMENTS.....	14

TABLES

Table 1: Instruments of Scope.....	6
Table 2: Eskom Lifesaving Rules	13

1. INTRODUCTION

Duvha Power Station requires a comprehensive contract service for the maintenance, calibration, technical support, and supply of spares and consumables for its coal laboratory analytical instruments. The coal laboratory plays a critical role in supporting efficient and reliable power generation through the accurate determination of coal quality parameters used for fuel management, process optimization, and operational decision-making. Effective coal quality monitoring is essential for maintaining combustion efficiency, reducing equipment wear, enhancing plant performance, and ensuring compliance with operational and environmental requirements. Daily coal samples collected from the mine coal preparation and sampling plant are analysed to provide accurate and timely data that directly influences plant operation and performance. To ensure the integrity, reliability, and traceability of these analytical results, all laboratory instruments must be maintained, serviced, repaired, and calibrated in accordance with ISO/IEC 17025 requirements. The availability of qualified technical support, preventative maintenance services, calibration standards, critical spare parts, and laboratory consumables is therefore vital to sustaining instrument performance, minimizing downtime, and maintaining continuous laboratory operations.

2. SUPPORTING CLAUSES

2.1 Scope

The scope includes the servicing, maintenance, calibration, repair, technical support, and supply of spares and consumables for coal laboratory analytical instruments at Duvha Power Station.

2.1.1 Purpose

The purpose of this Scope of Work is to define the requirements for the provision of service, maintenance, calibration, technical support, and the supply of spares and consumables for coal laboratory analytical instruments at Duvha Power Station over a period of sixty (60) months.

2.1.2 Applicability

This document shall apply throughout Duvha Power Station Chemical Services.

2.1.3 Effective date

It shall be effective from date of authorisation.

2.2 Normative/Informative References

Parties using this document shall apply the most recent edition of the documents listed in the following paragraphs.

2.2.1 Normative

- [1] ISO 9001: Quality Management Systems.
- [2] Insert normative document references here.

2.2.2 Informative

- [3] ISO/IEC 17025: General Requirements for the Competence of Testing and Calibration Laboratories.

CONTROLLED DISCLOSURE

2.3 Definitions

Supply of Spares and Consumables: The provision, delivery, and availability of all replacement parts, consumable items, accessories, reagents, and service-related materials required for the safe, reliable, accurate, and continuous operation of coal laboratory instruments.

2.4 Abbreviations

Abbreviation	Description
GHS	Globally Harmonized System
GLP	Good Laboratory Practice
ISO	International Organization for Standardization
NIST	National Institute of Standards and Technology

2.5 Roles and Responsibilities

2.5.1 Chemical Services

- Provide the Supplier with safe access to Eskom facilities, laboratory areas, equipment, and relevant personnel required for the execution of the contract.
- Review and approve service reports, calibration certificates, and other contract deliverables, and promptly communicate any non-conformances or corrective action requirements.
- Monitor Supplier performance and compliance with contractual, safety, quality, and regulatory requirements, including participation in audits, inspections, and performance reviews.

2.5.2 Service Provider

- Provide, deliver, store, maintain, calibrate, repair, and support all laboratory equipment, chemicals, standards, consumables, and spare parts to ensure uninterrupted laboratory operations.
- Ensure full compliance with Eskom Safety Requirements, Life-Saving Rules, Good Laboratory Practice (GLP), GHS requirements, and ISO/IEC 17025 quality standards throughout the contract period.
- Supply certified, traceable, fit-for-purpose products and services, including valid Certificates of Analysis (CoA), Safety Data Sheets (SDS), calibration certificates, and supporting technical documentation.
- Maintain qualified and OEM-authorised technical personnel to provide competent servicing, calibration, troubleshooting, repairs, and technical support within agreed response times.
- Implement robust quality assurance, traceability, inventory management, and reporting processes to ensure equipment reliability, regulatory compliance, and continuous laboratory performance.

2.6 Process for Monitoring

This document shall be subjected to internal reviews.

2.7 Related/Supporting Documents

N/A.

CONTROLLED DISCLOSURE

3. REQUIREMENTS FOR WORKS

3.1 Certificate of Analysis

3.1.1 A valid Certificate of Analysis (CoA) shall accompany every chemical, reference standard, certified standard, or reagent supplied for calibration, verification, analytical testing, or laboratory use.

3.1.2 The certificate shall clearly indicate the product identification, batch number, purity, concentration, manufacturing date, expiry date, and any other relevant quality parameters necessary to verify the suitability of the material for its intended purpose.

3.2 Safety Data Sheets

3.2.1 The Supplier shall provide a comprehensive 16-section Safety Data Sheet (SDS) for every chemical and standard supplied.

3.2.2 The SDS shall conform to GHS requirements and contain all information necessary for the safe handling, storage, use, transportation, and disposal of the substance.

3.2.3 The SDS provided shall not be older than five (5) years from the date of issue and shall accompany every delivery.

3.3 Quality of Chemicals and Standards

3.3.1 All chemicals, reagents, standards, and certified reference materials supplied for laboratory calibration, verification, and analytical testing shall be of the highest quality and shall be suitable for laboratory applications.

3.3.2 Unless otherwise specified, chemicals shall be Analytical Reagent (AR), SUPRAPURE, or an equivalent recognized high-purity grade.

3.3.3 The Supplier shall provide evidence of quality and traceability upon request

3.4 Product Shelf-Life Requirements

3.4.1 All chemicals, standards, consumables, and applicable instruments supplied shall have a minimum remaining shelf life of twelve (12) months from the date of delivery unless otherwise approved by the Employer.

3.4.2 Expiry dates and batch information shall be clearly visible and permanently marked on each container, package, or product label. Products delivered with insufficient shelf life may be rejected.

3.5 Product Catalogue Submission

3.5.1 Where available, the Supplier shall provide a comprehensive catalogue of chemicals, laboratory equipment, consumables, and spare parts together with the tender submission.

3.5.2 The catalogue shall contain technical specifications, product descriptions, ordering information, and manufacturer details to facilitate product evaluation and the prompt resolution of technical queries.

3.6 Traceability Requirements

3.6.1 Where specified, chemicals, standards, and certified reference materials shall be supplied with valid National Institute of Standards and Technology (NIST) traceable certificates.

CONTROLLED DISCLOSURE

3.6.2 The Supplier shall ensure that the traceability chain is maintained and documented to support the reliability, accuracy, and integrity of laboratory measurements and calibration activities.

3.7 Quality Assurance and Site Audits

3.7.1 The Employer reserves the right to conduct site visits, inspections, or audits at the Supplier's manufacturing facilities, warehouses, laboratories, or service centres.

3.7.2 Such visits may be undertaken to assess manufacturing practices, quality assurance systems, storage conditions, calibration capabilities, and the overall quality of products and services provided under this contract.

3.8 Instrument Service and Technical Support

3.8.1 The Contractor shall have the capability and capacity to perform service, maintenance, calibration, repairs, troubleshooting, and technical support for all instruments listed in Table A.

3.8.2 The Contractor shall also supply all necessary consumables, replacement parts, accessories, and spare components required to ensure the continued operation and reliability of the equipment throughout the contract period.

Table A: Instruments of Scope

LECO AC600 CALORIFIC VALUE (CV)	Unit of Measure	Quantity /Year	Rate	Cost/ Years	Cost/5 Year
SERVICE, MAINTENANCE AND CALIBRATION KIT OF AC600 CALORIFIC VALUE (CV): PRODUCT CODE YY-KIT-AC600	EA	1			
PLUG BANANA 15A 6-32 SCREW: PART NUMBER 621-605-948	EA	4			
WASHER FLT 0,187 X 0,500 X 0,12: PART NUMBER 776-562	EA	4			
PACKAGE O-RING VESSEL: PART NUMBER 621-605-686	EA	4			
PACKAGE O-RING PREP STATION: PART NUMBER 612-605-687	EA	10			
WIRE IGNITER FUSE THREAD 5/PK: PART NUMBER 502-460	EA	20			
FUSE THREAD COTTON 375/PK: PART NUMBER 502-715	EA	20			
SCREW MPHP 6-32X0,250 SST316: PART NUMBER 191-173	EA	5			
FILTER CARTRIDGE 1 MICRON: PART NUMBER 621-657	EA	5			
O-RING 003 0,056X0,171X0,060V: PART NUMBER 611-543	EA	5			
ASSY COMBUSTION VESSEL: PART NUMBER 621-605-954	EA	2			
ASSY COMBUSTION VESSEL RECERTIFICATION	EA	1			
SHAFT MAGNET AND STAND: PART NUMBER	EA	1			
CV CRUCIBLES: PART NUMBER: 774-204	EA	10			
TWIEZER CURVED FINE POINT: PART NUMBER 502-051					
BENZOIC TABLETS 0,5G; PART NUMBER 502-892	EA	3			
LUBRICATION GREASE 5.3OZ; PART NUMBER 501-241	EA	1			
BOTTLE 250ML LDPE WASH; PART NUMBER 601-980	EA	2			
TUBING FLX POLY BLK 0,166ID X 0,042W PART NUMBER 619*590-821	EA	2			
TUBE PFA 0,125 X 0,0625 ID 10FT; PART NUMBER 615-830-120	EA	2			
BEAKER 2000ML POLYPROP W/HANDLE & SPOUT; PART NUMBER 621-605-408	EA	1			

CONTROLLED DISCLOSURE

**SERVICE, MAINTENANCE, CALIBRATION, SUPPLY OF
SPARES AND CONSUMABLE OF COAL LABORATORY
INSTRUMENTS AS AND WHEN REQUIRED FOR 5 YEARS**

Unique Identifier: **559-59852433**

Revision: **1**

Page: **7 of 14**

CABLE ASSY CAT 5 15' RJ45 SHILD; PART NUMBER 620-677	EA	1			
ASSY TOOL REMOVAL INSERT VESSEL AC600; PART NUMBER 621-605-517	EA	1			
CABLE CONNECT STATION PREP AC600; PART NUMBER 621-605-557	EA	1			
TOOL INSTALL O-RING VALVE CHECK VESSEL; PART NUMBER 621-605-665	EA	1			
STAND REMOVAL O-RING VALVE CHECK VESSEL; PART NUMBER 621-605-666	EA	1			
ADAPTER FLUSH VALVE PREP STATION; PART NUMBER 621-605-744	EA	1			
ALIGNMENT TOOL PROBE VESSEL; PART NUMBER 621-605-750	EA	1			
KIT TOOL VESSEL COMBUSTION AC600; PART NUMBER 621-605-750	EA	1			
ASSY INLET WATER AC60; PART NUMBER 621-778	EA	1			
ASSY TUBE DRAIN; PART NUMBER 621-821	EA	1			
COMBUSTION VESSEL STAND' PART NUMBER 633-101-923	EA	1			
TOOL O-RING REMOVAL; PART NUMBER 762-515	EA	1			
ASSY CLEANING BRUSH; PART NUMBER 774-649	EA	1			
ASSY OXYGEN HOSE HI-PRESSURE; PART NUMBER 776-692	EA	1			
CORD POWER SHILD 16/3GA SJT 115' PART NUMBER 776-997	EA	1			
BAR UPPER ELECTRODE ISOLATED VESSEL; PART NUMBER 621-605-873	EA	1			
BAR ELECTRODE VESSEL; PART NUMBER 621-605-874	EA	1			
BAR LOWER ELECTRODE ISOLATED VESSEL; PART NUMBER 621-605-875	EA	1			
O-RING 00,250 X 0,374 X 0,062V; PART NUMBER 772-738	EA	1			
BRACKET RETAINING CRUCIBLES AC600; PART NUMBER 621-605-314	EA	1			
SHAFT MAGNET AND STAND: PRODUCT CODE	EA	1			
MAIN BOARD REPLACEMENT	EA	1			
TRAVELLING	EA	3			
LABOR	EA	10			
TRAVELING DISTANCE	EA	90			
KIT REPLACEMENT	EA	2			
SAFETY FILE	EA	1			

LECO CARBON AND SULPHURE SC832	Unit of Measure	Quantity /Year	Rate	Cost/ Years	Cost/5 Year
SC832 SERVICE KIT: PRODUCT CODE YY-KISC832	EA	1			
O-RING 121 1,062X1.248X0,093V; PART NUMBER 773-320	EA	1			
TUBE BOROS REAGENT 1,380D 7,0; PART NUMBER 626-001-871	EA	1			
EMITTER KIT	EA	1			
HEATING ELEMENTS ASSAY W/CLAMP AND PLUGS: PART NUMBER 625-525-129	EA	2			
ANHYDRONE 10-16 MESH: PART NUMBER 501-171-HAZ	EA	1			
GLASS WOOL 1LB: PART NUMBER 501-081	EA	1			
SECONDARY FILTER 10MICRON: PART NUMBER 775-306	EA	1			
O-RING 108 0,250X0,436X0,093V: PART NUMBER 772-520	EA	1			
SULPHURE BOAT CRUCIBLES ZIR 250/PK: PART NUMBER 528-203-250	EA	1			
SULFUR CRUCIBLE TONG: PART NUMBER 761-929	EA	1			
TUBE COMBUSTION INNER: PART NUMBER 625-601-555	EA	1			

CONTROLLED DISCLOSURE

**SERVICE, MAINTENANCE, CALIBRATION, SUPPLY OF
SPARES AND CONSUMABLE OF COAL LABORATORY
INSTRUMENTS AS AND WHEN REQUIRED FOR 5 YEARS**

Unique Identifier: **559-59852433**

Revision: **1**

Page: **8 of 14**

TUBE LANCE MANUAL LOAD: PART NUMBER 626-001-691	EA	1			
O-RING 326 1,625X1,999X0,187A: PART NUMBER 611-476	EA	1			
O-RING 330 2,125X2,499X0,187A: PART NUMBER 611-477	EA	1			
SCREW CSKH 8-32X0,50 ST ZP: PART NUMBER 190-737	EA	1			
HEATING ELEMENTS ASSAY W/CLAMP AND PLUGS: PART NUMBER 625-525-129	EA	1			
TUBE PURGE SST MANUAL: PART NUMBER 626-001-692	EA	1			
SREW MFHP 6-32X0,25 ST ZP: PART NUMBER 190-427	EA	1			
PLATE RETAINER O-RING PURGE: PART NUMBER 626-001-328	EA	1			
SCREW CSKH 0,250-20X0,62 SST: PART NUMBER 190-427	EA	1			
O-RING 906 0,468X0,624X0,078B FOR 5/16-18 UN FTG: PART NUMBER 175-001-103	EA	1			
FTG ADAPTOR SS 0,37T-0,56-18UNT: PART NUMBER 606-307	EA	1			
O-RING 0,206X0,309X0,103A: PART NUMBER 625-401-302	EA	1			
O-RING 006 0,114X0,184X0,070A: PART NUMBER 625-401-301	EA	1			
O-RING 121 1,062X1,248X0,093V: PART NUMBER 773-320	EA	1			
SREW CAPT 10-24X0,62 SST: PART NUMBER 606-342	EA	1			
CERAMIC STOP STEPPED: PART NUMBER 606-308	EA	1			
NUT FILTER MICRON PLT: PART NUMBER 619-591-878	EA	1			
COMBUSTION TUBE OUTER: PART NUMBER 625-601-554	EA	1			
SCREW FILTER: PART NUMBER 617-228	EA	1			
FILTER IN-LINE 0,1 MICRON DISPOSAL: PART NUMBER 619-591- 699	EA	1			
ACCELERATOR COM-CAT 250G: PART NUMBER 502-321	EA	1			
LUB GREASE VACUUM 5,3OZ: PART NUMBER 501-241	EA	1			
FUNNEL QUICK DISC TUBE: PART NUMBER 502-023	EA	1			
STD LCRM COAL SULPHUR 1,0% 50G: PART NUMBER 502-671	EA	1			
STD LCRM COAL SULPHUR 2,0% 50G: PART NUMBER 502-672	EA	1			
STD LCRM COAL SULPHUR 5,0% 50G: PART NUMBER 502-672	EA	1			
STD LCRM COAL TM PROX PLUS LOW 50G: PART NUMBER 502- 680	EA	1			
ASSY BLOWER CROSSFLOW AND CONNECTOR: PART NUMBER 625-501-731	EA	1			
HOURLY REGULAR TRAVELING	EA	3			
HOURLY REGULAR LABOR	EA	10			
TRAVELING DISTANCE/ZONE CHARGES	EA	90			

CONTROLLED DISCLOSURE

**SERVICE, MAINTENANCE, CALIBRATION, SUPPLY OF
SPARES AND CONSUMABLE OF COAL LABORATORY
INSTRUMENTS AS AND WHEN REQUIRED FOR 5 YEARS**

Unique Identifier: **559-59852433**

Revision: **1**

Page: **9 of 14**

LECO ASH FUSION TEMPERATURE AFT700	Unit of Measure	Quantity /Year	Rate	Cost/ Years	Cost/5 Year
AF700 SERVICE KIT	EA	1			
SCREW CSKH 10-24X0,62SST: PART NUMBER 190-105	EA	1			
WASHER FLT 0,219X0,500X0,04STZP: PART NUMBER 193-056	EA	1			
ADHESIVE CERAMIC ALUMINA 2OZ: PART NUMBER 501-233-HAZ	EA	1			
DEXTRIN SOLUTION 14,2/GR: PART NUMBER 502-010	EA	1			
STD CALIBRATION TEMP GOLD .02 2PK: PART NUMBER 502-496	EA	2			
STD CALIBRATION TEMP NICKEL: PART NUMBER 501-951	EA	2			
INSULATOR FRONT LOWER: PART NUMBER 625-036	EA	3			
O-RING 332 2,375X2,750X0,187A: PART NUMBER 625-501-149	EA	5			
FLANGE COMPRESSION SEAL FURNACE: PART NUMBER 625-501-152	EA	3			
END BLOCK FRONT FURNACE AF SST: PART NUMBER 625-501-167	EA	2			
BAFFLE LIGHT BACKGROUND FURNACE P: PART NUMBER 652-501-247	EA	2			
COVER END TUBE D FURNACE PLT: PART NUMBER 625-501-248	EA	2			
ASSY CONNECTION GAS REDUCING: PART NUMBER 625-501-325	EA	2			
ASSY CONNECTION GAS PURGE: PART NUMBER 625-501-325	EA	2			
KIT REPL 6 HEATING ELEMENTS AF700: PART NUMBER 625-501-328	EA	2			
CONE TRAY 6/5 PLC DUAL SIDED 100/PK: PART NUMBER 625-501-662	EA	2			
GASKET ASSY TERMINAL ELEMENT 0,75: PART NUMBER 633-103-092	EA	1			
TRAY CONE 6/5 PLC DUAL SIDED 10/PK: PART NUMBER 625-501-661	EA	1			
KEY BYPASS/OVERRIDE SOL AZM170: PART NUMBER 625-501-431	EA	1			
FILTER LIGHT NEUTRAL DEN 29MM: PART NUMBER 625-501-443	EA	1			
INSULATION END TUBE D FURNACE: PART NUMBER 625-501-469	EA	1			
CLAMP ELEMENT HEATING 0,75 IN: PART NUMBER 625-501-470	EA	1			
PIN STOP TRAY CONE SAMPLE: PART NUMBER 625-501-471	EA	1			
INSULATION MOD SIDE FILTER: PART NUMBER 710-496	EA	1			

CONTROLLED DISCLOSURE

**SERVICE, MAINTENANCE, CALIBRATION, SUPPLY OF
SPARES AND CONSUMABLE OF COAL LABORATORY
INSTRUMENTS AS AND WHEN REQUIRED FOR 5 YEARS**

Unique Identifier: **559-59852433**

Revision: **1**

Page: **10 of 14**

KIT REGULATOR PRESS 2 STAGE FLAMMABLE CGA350 QTV: PART NUMBER 633-103-357	EA	1			
KIT REGULATOR PRESS 2 STAGE SCI AIR CGA346 QTV: 633-103-354	EA	1			
MOLD CONE AMERICAN STANDARD: PART NUMBER 776-957	EA	1			
TUBE MULLITE 0,125X0,250X15,0: PART NUMBER 633-101-990	EA	1			
TUBE ALIMINA 2,250X2,500X22,20: PART NUMBER 633-101-997	EA	1			
O-RING 202 0,250X0,500X0,125A: PART NUMBER 625-501-430	EA	1			
CABLE ALUM BRAID 50 AMP 7,50 LG: PART NUMBER 633-101-361	EA	1			
CABLE ALUM BRAID 50 AMP 9,20 LG: PART NUMBER 633-101-362	EA	1			
CAMERA DGTL 1,3MP USB LU170M: PART NUMBER 625-501-682	EA	1			
TOOL EXPANSION FOR T-CLAMP: PART NUMBER 613-564	EA	1			
INSULATOR FRONT LOWER: PART NUMBER 625-036	EA	1			
INSULATOR FRONT UPPER: PART NUMBER 625-037	EA	1			
TRAVELLING	EA	1			
LABOR	EA	6			
TRAVELING DISTANCE	EA	30			
FURNACES					
ISO ASH FURNACE CALIBRATION	EA	2			
ISO ASH FURNACE ELEMENT	EA	1			
ISO ASH FURNACE CONTROLLER PANEL	EA	1			
THERMOCOUPLE TYPE K	EA	1			
SHEATH 10MM OD X 300MM LONG PYTHAGORAS	EA	1			
BOARD 25MM FIRED AND RIGIDIZED 1260°C	EA	1			
BOARD 50MM FIRED AND RIGIDIZED 1260°C	EA	1			
CALIBRATION CERTIFICATE	EA	2			
VOLATILE FURNACE CALIBRATION	EA	2			
VOLATILE FURNACE CONTROLLER PANEL	EA	2			
CALIBRATION CERTIFICATES	EA	2			
TRAVELLING	EA	1			
LABOR	EA	6			
TRAVELING DISTANCE	EA	30			
Ovens					

CONTROLLED DISCLOSURE

When downloaded from the document management system, this document is uncontrolled and the responsibility rests with the user to ensure it is in line with the authorized version on the system. No part of this document may be reproduced in any manner or form by third parties without the written consent of Eskom Holdings SOC Ltd, © copyright Eskom Holdings SOC Ltd, Reg No 2002/015527/30

OVENS CALIBRATIONS	EA	8			
FAN REPLACEMENT	EA	8			
CALIBRATION CERTIFICATES	EA	8			
VERIFICATION INSTRUMENTS					
MASS PIECES SET	EA	2			
MASS PIECES 2KG	EA	1			
MASS PIECES 5KG	EA	1			
HYGROMETER TEMPERATURE CALIBRATION	EA	5			
HYGROMETER HUMIDITY CALIBRATION	EA	5			
DIGITAL THERMOMETER WITH TYP K FLEXI PROBE	EA	2			
DIGITAL THERMOMETER WITH SOLID PROBE	EA	2			
SANPLATECH DRY KEEPER DESICCATOR	EA	1			
STOPWATCHES	EA	2			
ANALYTICAL/TOP PAN BALANCES	EA	6			
CALIBRATION CERTIFICATES	EA	27			
TRAVELLING	EA	1			
LABOR	EA	6			
TRAVELING DISTANCE	EA	30			
DIGITAL THERMOMETER WITH TYP K FLEXI AND SOLID PROBE	EA	1			
COAL PROFICIENCY TESTING					
COAL PT GEN ANALYSIS, AI, HGI	EA	12			
COAL PT TOTAL MOISTURE	EA	6			
COAL PT PARTICLE SIZE DISTRIBUTION	EA	2			
COAL PT COURIE	EA	12			
COAL PSD COURIE	EA	2			
COAL ACCREDITED REFERENCE MATERIAL ARM	EA	5			
BEZOIC ACID TABLETS 100/PK 0,5G, ULTRA PURE 99,5-100%, COMBUSTION VALUE 26,444MJ/Kg	EA	1			
COAL PT CURIER TOTAL MOISTURE	EA	6			
SANAS ACCREDITATION ANNUAL MEMBERSHIP FEE	EA	2			

3.9 Technical Competency Requirements

3.9.1 All technicians, engineers, and specialists performing work on laboratory instruments and equipment shall be appropriately trained, qualified, and authorised to perform such work.

3.9.2 Personnel carrying out maintenance, repairs, calibration, or servicing shall be declared competent by the Original Equipment Manufacturer (OEM) and shall provide proof of competency, training, or certification upon request.

3.10 Calibration Certification Requirements

3.10.1 Calibration certificates issued for ovens, furnaces, stopwatches, hygrometers, digital thermometers, mass pieces, balances, flexible probes, solid probes, and any other applicable measuring instruments shall be authorised and signed by a SANAS ISO/IEC 17025-appointed Technical Signatory.

CONTROLLED DISCLOSURE

3.10.2 Calibration certificates shall clearly indicate measurement results, traceability, uncertainty values, environmental conditions, and equipment identification details.

3.11 Service Reporting

3.11.1 Following any maintenance, servicing, calibration, repair, inspection, or technical intervention, the Supplier shall provide a detailed service report.

3.11.2 The report shall include, as a minimum, equipment identification, work performed, observations, test results, calibration data, replaced parts, corrective actions taken, recommendations, and confirmation of instrument performance.

3.11.3 All reports shall be prepared in accordance with the requirements and principles of ISO/IEC 17025.

3.12 Response Time for Equipment Failures

3.12.1 The Contractor shall respond promptly to all equipment breakdowns and operational failures.

3.12.2 Any equipment failure reported by the Employer shall be attended to without delay, and corrective action shall be initiated immediately.

3.12.3 The Contractor shall rectify equipment failures within twenty-four (24) hours of notification unless otherwise agreed in writing by the Employer.

3.13 Availability of Spare Parts

3.13.1 The Contractor shall maintain sufficient stock levels of critical spare parts, consumables, and replacement components to support all equipment covered under the contract.

3.13.2 The Contractor shall ensure the continuous availability and reliability of such items to minimise operational downtime and prevent disruptions to laboratory activities.

3.14 Prohibition of Chemical Decanting

3.14.1 The decanting, repackaging, transfer, or subdivision of chemicals into secondary containers is strictly prohibited.

3.14.2 All chemicals and standards shall be supplied in the original manufacturer's sealed containers with intact labels, batch numbers, expiry dates, and certification documentation.

3.14.3 Any chemicals found to be decanted or repackaged shall be rejected immediately upon delivery.

3.15 Delivery Documentation Requirements

3.15.1 Each delivery of chemicals, standards, and reagents shall be accompanied by all relevant documentation, including Certificates of Analysis, NIST traceability certificates where applicable, and valid 16-section Safety Data Sheets.

3.15.2 The Supplier shall ensure that all documentation is complete, accurate, legible, and available at the time of delivery.

3.16 Transport Vehicle and Driver Requirements

3.16.1 The Supplier shall ensure that all vehicles used for the transportation of chemicals and hazardous substances are roadworthy, appropriately licensed, and equipped for the safe transport of such materials.

CONTROLLED DISCLOSURE

3.16.2 Drivers shall possess a valid driver's licence and any additional permits required by law. Furthermore, the Supplier shall establish and maintain an emergency response procedure for accidents, spills, leaks, fires, or other incidents that may occur during transportation.

3.16.3 Emergency response information and contact details shall always be readily available to the driver and relevant authorities.

Table 1: Eskom Lifesaving Rules

Rule No: 1	Open, isolate, test, earth, and create an equipotential zone.
Rule No: 2	Hook up at height
Rule No: 3	Buckle up
Rule No: 4	Be sober
Rule No: 5	Permit to work
Rule No: 6	Ensure safe live working

CONTROLLED DISCLOSURE

4. ACCEPTANCE

This document has been seen and accepted by:

Name	Designation
Lefumo Maswanganye	Manager Chemical Services
Cecil Mngqibisa	Senior Chemist
Phakamile Ngxekana	Senior Technician Chemistry
Andre Groenewald.	Senior Supervisor
Simon Tsamago	Senior Supervisor

5. REVISIONS

Date	Rev.	Compiler	Remarks
August 2026	1	Bavuseni Mogola	Original Document

6. DEVELOPMENT TEAM

The following people were involved in the development of this document:

- Phakamile Ngxekana
- Cecil Mngqibisa

7. ACKNOWLEDGEMENTS

N/A.

CONTROLLED DISCLOSURE