 GAUTENG PROVINCE PROVINCIAL TREASURY REPUBLIC OF SOUTH AFRICA		<h1 style="text-align: center;">Provincial Supply Chain Management</h1>							
		Request for Proposal		Page 1 of 4					
RFP NUMBER									
RFP DESCRIPTION									
CUSTOMER DEPARTMENT									
CUSTOMER INSTITUTION									
BRIEFING SESSION	Y		N		SESSION COMPULSORY	Y		N	
					SESSION HIGHLY RECOMMENDED	Y		N	
BRIEFING VENUE					DATE			TIME	
COMPULSORY SITE INSPECTION	Y		N		DATE			TIME	
INSPECTION ADDRESS									
TERM AGREEMENT CALLED FOR?	Y		N		TERM DURATION				
CLOSING DATE					CLOSING TIME				
TENDER BOX LOCATION									
GPT is acting as Common Service Provider or buying organisation on behalf of all Gauteng Provincial Government Customer Departments / Institutions. The goods / services are therefore required by the Customer Department / Institution, as indicated on this form RFP 01.									

Notes:

- All bids / tenders must be deposited in the Tender Box at the following address:
Gauteng Provincial Treasury, Imbumba House, 75 Fox Street, Marshalltown, Johannesburg
- Bids / tenders must be deposited in the Tender Box on or before the closing date and time.
- Bids / tenders submitted by fax will not be accepted.
- The GPT Tender Box is generally open 24 hours a day, 7 days a week.
- This bid is subject to the preferential procurement policy framework act, 2000 and the preferential procurement regulations, 2022, the general conditions of contract (gcc) 2010 and, if applicable, any other special conditions of contract.
- ALL BIDS MUST BE SUBMITTED ON THE OFFICIAL GPG RFP FORMS – (NOT TO BE RE-TYPED)
- ALL REQUIRED INFORMATION MUST BE COMPLETED (FAILURE TO DO SO MAY RESULT IN YOUR BID BEING DISQUALIFIED):

The Tendering System

The RFP Pack consists of two parts namely, Section 1 and Section 2. These two sections must be submitted separately, clearly marked with the Tender Number and the Section Number.

Training sessions

Non-compulsory "How to tender" workshops are held every Wednesday at 75 Fox Street from 10:00-13:00.



Provincial Supply Chain Management

Request for Proposal

Page 2 of 4

PART A INVITATION TO BID

SUPPLIER INFORMATION

NAME OF BIDDER					
POSTAL ADDRESS					
STREET ADDRESS					
TELEPHONE NUMBER	CODE		NUMBER		
CELLPHONE NUMBER					
FACSIMILE NUMBER	CODE		NUMBER		
E-MAIL ADDRESS					
VAT REGISTRATION NUMBER					
SUPPLIER COMPLIANCE STATUS	TAX COMPLIANCE SYSTEM PIN:		OR	CENTRAL SUPPLIER DATABASE No:	MAAA
ARE YOU THE ACCREDITED REPRESENTATIVE IN SOUTH AFRICA FOR THE GOODS /SERVICES OFFERED?	<input type="checkbox"/> Yes <input type="checkbox"/> No [IF YES ENCLOSE PROOF]		ARE YOU A FOREIGN BASED SUPPLIER FOR THE GOODS /SERVICES OFFERED?		<input type="checkbox"/> Yes <input type="checkbox"/> No [IF YES, ANSWER THE QUESTIONNAIRE BELOW]

QUESTIONNAIRE TO BIDDING FOREIGN SUPPLIERS

IS THE ENTITY A RESIDENT OF THE REPUBLIC OF SOUTH AFRICA (RSA)?	<input type="checkbox"/> YES <input type="checkbox"/> NO
DOES THE ENTITY HAVE A BRANCH IN THE RSA?	<input type="checkbox"/> YES <input type="checkbox"/> NO
DOES THE ENTITY HAVE A PERMANENT ESTABLISHMENT IN THE RSA?	<input type="checkbox"/> YES <input type="checkbox"/> NO
DOES THE ENTITY HAVE ANY SOURCE OF INCOME IN THE RSA?	<input type="checkbox"/> YES <input type="checkbox"/> NO
IS THE ENTITY LIABLE IN THE RSA FOR ANY FORM OF TAXATION?	<input type="checkbox"/> YES <input type="checkbox"/> NO
IF THE ANSWER IS "NO" TO ALL OF THE ABOVE, THEN IT IS NOT A REQUIREMENT TO REGISTER FOR A TAX COMPLIANCE STATUS SYSTEM PIN CODE FROM THE SOUTH AFRICAN REVENUE SERVICE (SARS) AND IF NOT REGISTER AS PER 2.3 BELOW.	



Provincial Supply Chain Management

Request for Proposal

Page 3 of 4


Tender documents can be obtained from <http://www.treasury.gpg.gov.za>

ANY ENQUIRIES REGARDING BIDDING PROCEDURE MAY BE DIRECTED TO:

DEPARTMENT	
CONTACT PERSON	
TELEPHONE NUMBER	
FACSIMILE	
E-MAIL ADDRESS	

ANY ENQUIRIES REGARDING TECHNICAL INFORMATION MAY BE DIRECTED TO:

DEPARTMENT	
CONTACT PERSON	
TELEPHONE NUMBER	
FACSIMILIE	
E-MAIL ADDRESS	

 <p>GAUTENG PROVINCE PROVINCIAL TREASURY REPUBLIC OF SOUTH AFRICA</p>	<h1 style="text-align: center;">Provincial Supply Chain Management</h1>	
Request for Proposal		Page 4 of 4

PART B TERMS AND CONDITIONS FOR BIDDING

1. BID SUBMISSION:
<p>1.1. BIDS MUST BE DELIVERED BY THE STIPULATED TIME TO THE CORRECT ADDRESS. LATE BIDS WILL NOT BE ACCEPTED FOR CONSIDERATION.</p> <p>1.2. ALL BIDS MUST BE SUBMITTED ON THE OFFICIAL FORMS PROVIDED (NOT TO BE RE-TYPED) OR IN THE MANNER PRESCRIBED IN THE BID DOCUMENT.</p> <p>1.3. THIS BID IS SUBJECT TO THE PREFERENTIAL PROCUREMENT POLICY FRAMEWORK ACT, 2000 AND THE PREFERENTIAL PROCUREMENT REGULATIONS, THE GENERAL CONDITIONS OF CONTRACT (GCC) AND, IF APPLICABLE, ANY OTHER SPECIAL CONDITIONS OF CONTRACT.</p> <p>1.4. THE SUCCESSFUL BIDDER WILL BE REQUIRED TO FILL IN AND SIGN A WRITTEN CONTRACT FORM (SBD7).</p>
2. TAX COMPLIANCE REQUIREMENTS
<p>2.1 BIDDERS MUST ENSURE COMPLIANCE WITH THEIR TAX OBLIGATIONS.</p> <p>2.2 BIDDERS ARE REQUIRED TO SUBMIT THEIR UNIQUE PERSONAL IDENTIFICATION NUMBER (PIN) ISSUED BY SARS TO ENABLE THE ORGAN OF STATE TO VERIFY THE TAXPAYER'S PROFILE AND TAX STATUS.</p> <p>2.3 APPLICATION FOR TAX COMPLIANCE STATUS (TCS) PIN MAY BE MADE VIA E-FILING THROUGH THE SARS WEBSITE WWW.SARS.GOV.ZA.</p> <p>2.4 BIDDERS MAY ALSO SUBMIT A PRINTED TCS CERTIFICATE TOGETHER WITH THE BID.</p> <p>2.5 IN BIDS WHERE CONSORTIA / JOINT VENTURES / SUB-CONTRACTORS ARE INVOLVED; EACH PARTY MUST SUBMIT A SEPARATE TCS CERTIFICATE / PIN / CSD NUMBER. <i>25 August 2023</i></p> <p>2.6 WHERE NO TCS PIN IS AVAILABLE BUT THE BIDDER IS REGISTERED ON THE CENTRAL SUPPLIER DATABASE (CSD), A CSD NUMBER MUST BE PROVIDED.</p> <p>2.7 NO BIDS WILL BE CONSIDERED FROM PERSONS IN THE SERVICE OF THE STATE, COMPANIES WITH DIRECTORS WHO ARE PERSONS IN THE SERVICE OF THE STATE, OR CLOSE CORPORATIONS WITH MEMBERS PERSONS IN THE SERVICE OF THE STATE."</p>

NB: FAILURE TO PROVIDE / OR COMPLY WITH ANY OF THE ABOVE PARTICULARS MAY RENDER THE BID INVALID.

SIGNATURE OF BIDDER		DATE	
CAPACITY UNDER WHICH THIS BID IS SIGNED (Proof of authority must be submitted e.g. company resolution)			



CONSENT FORM TO PROCESS PERSONAL INFORMATION IN TERMS OF THE PROTECTION OF PERSONAL INFORMATION ACT, NO. 4 OF 2013 (POPIA).

The purpose of the POPIA is to protect personal information of individuals and businesses and to give effect to their right of privacy as provided for in the Constitution.

By signing this form, you consent to your personal information to be processed by the Gauteng Department of Health and consent is effective immediately and will remain effective until such consent is withdrawn.

APPLICATION FOR THE CONSENT OF A DATA SUBJECT FOR THE PROCESSING OF PERSONAL INFORMATION FOR THE PURPOSE OF BIDS

Name & Surname/Company: _____

Residential/Postal or Business Address: _____

Contact number (s): _____

Email address: _____

1. In the furtherance of the Gauteng Department of Health's (**The Department**) operational requirements and for purposes of complying with its policies, procedures and privacy laws, we may be required to disclose, process and/or further process your personal information provided to us and/or made available by virtue of submission of this bid.
2. For purposes contemplated in paragraph 1, the Department, hereby requests your consent and/or authorisation for the disclosure, processing and/or further processing of any and/or all your personal information as may be necessary for reasons provided in paragraph 1.
3. By signing this Personal Information Processing Consent Form, you hereby grant the Department permission, consent and/or authorisation to disclose, process and further process your personal information within our records, as may be required and/or necessary from time to time.

I, the undersigned, _____ (*INSERT FULL NAME AND SURNAME*) with Identity Number _____, in my personal capacity or acting on behalf of _____
_____ (Name of **Company**), confirm that:

4. I have read and understood the contents of this Personal Information Processing Consent form, the details of which have been explained to me and furthermore I understand my right to privacy and the right to have my personal information processed in accordance with the conditions for the lawful processing of personal information.
5. I declare that all my personal information supplied to the Department is accurate, up to date, not misleading and that it is complete in all respects and will be held and/ or stored securely for the purpose for which it was collected and that I will immediately advise the Department of any changes to my Personal Information should any of these details change.
6. I also understand that I have the right to request that my personal information be corrected or deleted, if it is inaccurate, irrelevant, excessive, out of date, incomplete, misleading, or obtained unlawfully or that the personal information or record be destroyed or deleted if the Department is no longer authorised to retain it.
7. I declare that my personal/the Company's information and/or data may be disclosed, processed and/or further processed by the Department (including its employees, agents, contractors and representatives) and such other third parties contracted with the Department involved in the processing, verification and management of my and/or Company's Personal Information in accordance with the requirements set out in paragraph 1;
8. I accept the data security and protection measures adopted and/or applied by the Department in their retention, disclosure, processing, and further processing of my and/or Company's personal information/data.
9. I accept that the Department may retain any of my personal/the Company information/data as may be required for purposes contemplated in paragraph 1.

10. With my signature below, do hereby give my or the Company's irrevocable consent, and/or authorisation for purposes required and/or detailed in this *Personal Information Processing Consent* form.

Signed at this day of20.....

.....

Name of data subject/ designated person

.....

Signature

.....

Name/Surname/Dept of Responsible Party

.....

Signature

Date:



GAUTENG PROVINCE
PROVINCIAL TREASURY
REPUBLIC OF SOUTH AFRICA

Provincial Supply Chain Management

RFP Point System

Page 1 of 1

RFP NUMBER		CLOSING DATE	
VALIDITY OF RFP		CLOSING TIME	

In case of queries, please contact the GPT Contact Centre at tel: 0860 011 000

*GPT is acting as Common Service Provider or buying organisation on behalf of all Gauteng Provincial Government Customer Departments / Institutions.

The goods / services are required by the Customer Department / Institution, as indicated on RFP 01.

The Gauteng Provincial Government requests your bid on the goods and/or services listed on the attached forms. Please furnish all information as requested and return your bid on the date stipulated. Late bids will not be accepted for consideration.

This RFP will be evaluated on the basis of the under noted point system, as stipulated in the Preferential Procurement Policy Framework Act (Act number 5 of 2000).

Point System

Points SHALL be allocated as follows:

Points for

Points for

TYPE OF CONTRACT (COMPLETED BY PROJECT MANAGER)

VALUE BASED

SERVICE BASED	Y		N		SERVICE BASED	Y		N		VALUE BASED	Y		N	
VALUE BASED	Y		N											
QUANTITY BASED	Y		N											
TERM BASED	Y		N											



Provincial Supply Chain Management

Instructions to Bidders

Page 1 of 2

1. The RFP (Request for Proposal) Pack is drawn up so that certain essential information should be furnished in a specific manner. Any additional particulars shall be furnished in a separate annexure.
2. The RFP forms should not be retyped or redrafted, but photocopies may be prepared and used. Additional offers may be made for any item, but only on a photocopy of the page in question or on other forms obtainable from the relevant Department or Institution advertising this RFP. Additional offers made in any other manner may be disregarded.
3. Should the RFP forms not be filled in by means of electronic devices, bidders are encouraged to complete forms in a black ink.
4. Bidders shall check the numbers of the pages and satisfy themselves that none are missing or duplicated. No liability shall be accepted with regards to claims arising from the fact that pages are missing or duplicated.
5. The forms RFP 04 to RFP 09 and PREF documents shall be completed, signed and submitted with the bid. RFP 10 (National Industrial Participation Programme Form) will only be added to the RFP pack to be completed by bidders when an imported component in excess of US \$ 10 million is expected.
6. A separate RFP 06 form (RFP Price Schedule per item) shall be completed in respect of each item. Photocopies of this form may be prepared and used or additional copies, (if required) are obtainable from the relevant Department or Institution advertising this RFP(not applicable for Pre-qualification of Bidders).
7. Firm delivery periods and prices are preferred. Consequently bidders shall clearly state whether delivery periods and prices will remain firm or not for the duration of any contract, which may result from this RFP, by completing RFP 06 (RFP Price Schedule per item) and RFP 07 (Non-Firm Prices per item) (not applicable for Pre-qualification of Bidders). *25 August 2023*
8. If non-firm prices are offered bidders must ensure that a separate RFP 07 (Non-Firm Prices per item) is completed in respect of each item for which a non-firm price is offered. Photocopies of this form may be prepared and used or additional copies, (if required) are obtainable from the relevant Department or Institution advertising this RFP (not applicable for Pre-qualification of Bidders).
9. Where items are specified in detail, the specifications form an integral part of the RFP document (see the attached specification) and bidders shall indicate in the space provided whether the items offered are to specification or not (not applicable for Pre-qualification of Bidders).
10. In respect of the paragraphs where the items offered are strictly to specification, bidders shall insert the words "as specified" (see the attached specification) (not applicable for Pre-qualification of Bidders).
11. In cases where the items are not to specification, the deviations from the specifications shall be indicated (see the attached specification).
12. In instances where the bidder is not the manufacturer of the items offered, the bidder must as per RFP 06 (RFP Price Schedule per item) submit a Letter of Supply from the relevant manufacturer or his supplier (not applicable for Pre-qualification of Bidders).
13. The offered prices shall be given in the units shown in the attached specification, as well as in RFP 06 (RFP Price Schedule per item) (not applicable for Pre-qualification of Bidders).
14. With the exception of imported goods, where required, all prices shall be quoted in South African currency. Where bids are submitted for imported goods, foreign currency information must be supplied by completing the relevant portions of RFP 06 (RFP Price Schedule per item) and RFP 07 (Non-Firm Prices per item) (not applicable for Pre-qualification of Bidders).
15. Unless otherwise indicated, the costs of packaging materials (if applicable) are for the account of the bidder and must be included in the bid price on RFP 06 (RFP Price Schedule per item) (not applicable for Pre-qualification of Bidders).
16. Delivery basis (not applicable for Pre-qualification of Bidders):
 - (a) Supplies which are held in stock or are in transit or on order from South African manufacturers at the date of offer shall be offered on a basis of delivery into consignee's store or on his site within the free delivery area of the bidder's centre, or carriage paid consignee's station, if the goods are required elsewhere.
 - (b) Notwithstanding the provisions of paragraph 16(a), offered prices for supplies in respect of which installation / erection / assembly is a requirement, shall include ALL costs on a "delivered on site" basis, as specified on RFP 06 (RFP Price Schedule per item).
17. Unless specifically provided for in the RFP document, no bids transmitted by facsimile or email shall be considered.
18. Failure on the part of the bidder to sign any of the forms RFP 04 to RFP 10 and PREF documents and thus to acknowledge and accept the conditions in writing or to complete the attached RFP forms, Preference documents, questionnaires and specifications in all respects, may invalidate the bid.
19. Bids should preferably not be qualified by the bidder's own conditions of bid. Failure to comply with these requirements (i.e. full



Provincial Supply Chain Management

Instructions to Bidders

Page 2 of 2

acceptance of the General Conditions of Contract or to renounce specifically the bidder's own conditions of bid, when called upon to do so, may invalidate the bid.

20. In case of samples being called for together with the bid (refer to RFP 05 in this regard), the successful bidder may be required to submit **pre-production samples** to the South African Bureau of Standards (SABS) or such testing authority as designated at the request of the relevant Department concerned. Unless the relevant Department decides otherwise, pre-production samples must be submitted within thirty (30) days of the date on which the successful bidder was requested to do so. Mass production may commence only after both the relevant Department and the successful bidder have been advised by the SABS that the pre-production samples have been approved.
21. Should the pre-production samples pass the inspections / tests at the first attempt, the costs associated with the inspections / tests will be for the account of the relevant Department. If the SABS or such testing authority as designated do not approve the pre-production samples, but requires corrections / improvements, the costs of the inspections / tests must be paid by the successful bidder and samples which are acceptable in all respects must then reach the SABS or such testing authority as designated within twenty-one (21) days of the date on which the findings of the SABS or such testing authority as designated were received by the successful bidder. Failure to deliver samples within the specified time and to the required standards may lead to the cancellation of the intended contract.
22. In case of samples being called for together with the bid (refer to RFP 05 in this regard), the samples must be submitted together with the bid before the closing time and date of the RFP, unless specifically indicated otherwise. Failure to submit the requested sample(s) before the closing time and date of the RFP may invalidate the bid.
23. In cases where large quantities of a product are called for, it may be necessary for the relevant item to be shared among two (2) or more suppliers.
24. In cases where the relevant Department or Institution advertising this RFP may deem it necessary, a formal contract may be entered into with the successful bidder, in addition to a Letter of Acceptance and / or purchase order being issued.
25. If any of the conditions on the RFP forms are in conflict with any special conditions, stipulations or provisions incorporated in the bid invitation, such special conditions, stipulations or provisions shall apply.
26. This RFP is subject to the General Conditions of Contract and re-issues thereof. Copies of these conditions are obtainable from any office of the Gauteng Provincial Government (GPG).
27. Each bid must be submitted in a separate, sealed envelope on which the following must be clearly indicated:
 - NAME AND ADDRESS OF THE BIDDER;
 - THE BID (RFP) NUMBER; AND
 - THE CLOSING DATE.

The bid must be deposited or posted;

 - posted to Gauteng Provincial Treasury and to reach the destination not later than the closing time and date; OR
 - deposited in the tender box of the Gauteng Provincial Treasury before the closing time and date.
28. The Gauteng Provincial Government has become a member and as such a key sponsor of the Proudly South African Campaign. GPG therefore would like to procure local products of a high quality, produced through the practise of sound labour relations and in an environment where high environmental standards are maintained. In terms of the Proudly South African Campaign South African companies are encouraged to submit interesting and innovative achievements in the manufacturing field (if relevant to this RFP) – including information on new products, export achievements, new partnerships and successes and milestones.
29. **Compulsory GPG Contract:** It is a mandatory requirement that successful bidder/s (to whom a tender is awarded) sign a GPG Contract upon award of any given contract.

 GAUTENG PROVINCE PROVINCIAL TREASURY REPUBLIC OF SOUTH AFRICA	<h1>Provincial Supply Chain Management</h1>	
	Bid Commitment and Declaration of Interest	Page 1 of 3

BIDDER'S DISCLOSURE

1. PURPOSE OF THE FORM

Any person (natural or juristic) may make an offer or offers in terms of this invitation to bid. In line with the principles of transparency, accountability, impartiality, and ethics as enshrined in the Constitution of the Republic of South Africa and further expressed in various pieces of legislation, it is required for the bidder to make this declaration in respect of the details required hereunder.

Where a person/s are listed in the Register for Tender Defaulters and / or the List of Restricted Suppliers, that person will automatically be disqualified from the bid process.

2. Bidder's declaration

- 2.1 Is the bidder, or any of its directors / trustees / shareholders / members / partners or any person having a controlling interest¹ in the enterprise, employed by the state?

YES		NO	
------------	--	-----------	--

- 2.1.1 If so, furnish particulars of the names, individual identity numbers, and, if applicable, state employee numbers of sole proprietor/ directors / trustees / shareholders / members/ partners or any person having a controlling interest in the enterprise, in table below.

Full Name	Identity Number	Name of State institution

¹ the power, by one person or a group of persons holding the majority of the equity of an enterprise, alternatively, the person/s having the deciding vote or power to influence or to direct the course and decisions of the enterprise.

 GAUTENG PROVINCE PROVINCIAL TREASURY REPUBLIC OF SOUTH AFRICA	<h1>Provincial Supply Chain Management</h1>	
	Bid Commitment and Declaration of Interest	Page 2 of 3

- 2.2 Do you, or any person connected with the bidder, have a relationship with any person who is employed by the procuring institution?

YES		NO	
------------	--	-----------	--

- 2.2.1 If so, furnish particulars:

- 2.3 Does the bidder or any of its directors / trustees / shareholders / members / partners or any person having a controlling interest in the enterprise have any interest in any other related enterprise whether or not they are bidding for this contract?

YES		NO	
------------	--	-----------	--

- 2.3.1 If so, furnish particulars:

3. DECLARATION

I, the undersigned, (name)..... in submitting the accompanying bid, do hereby make the following statements that I certify to be true and complete in every respect:

- 3.1 I have read and I understand the contents of this disclosure;
- 3.2 I understand that the accompanying bid will be disqualified if this disclosure is found not to be true and complete in every respect;
- 3.3 The bidder has arrived at the accompanying bid independently from, and without consultation, communication, agreement or arrangement with any competitor. However, communication between partners in a joint venture or consortium² will not be construed as collusive bidding.
- 3.4 In addition, there have been no consultations, communications, agreements or arrangements with any competitor regarding the quality, quantity, specifications, prices, including methods, factors or formulas used to calculate prices, market allocation, the intention or decision to submit or not to submit the bid, bidding with the intention not to win the bid and conditions or delivery particulars of the products or services to which this bid invitation relates.
- 3.4 The terms of the accompanying bid have not been, and will not be, disclosed by the bidder, directly or indirectly, to any competitor, prior to the date and time of the official bid opening or of the awarding of the contract.

² Joint venture or Consortium means an association of persons for the purpose of combining their expertise, property, capital, efforts, skill and knowledge in an activity for the execution of a contract.
 Filename: RFP4GPT (SBD4)

 GAUTENG PROVINCE PROVINCIAL TREASURY REPUBLIC OF SOUTH AFRICA	<h1>Provincial Supply Chain Management</h1>	
	Bid Commitment and Declaration of Interest	Page 3 of 3

3.5 There have been no consultations, communications, agreements or arrangements made by the bidder with any official of the procuring institution in relation to this procurement process prior to and during the bidding process except to provide clarification on the bid submitted where so required by the institution; and the bidder was not involved in the drafting of the specifications or terms of reference for this bid.

3.6 I am aware that, in addition and without prejudice to any other remedy provided to combat any restrictive practices related to bids and contracts, bids that are suspicious will be reported to the Competition Commission for investigation and possible imposition of administrative penalties in terms of section 59 of the Competition Act No 89 of 1998 and or may be reported to the National Prosecuting Authority (NPA) for criminal investigation and or may be restricted from conducting business with the public sector for a period not exceeding ten (10) years in terms of the Prevention and Combating of Corrupt Activities Act No 12 of 2004 or any other applicable legislation.

I CERTIFY THAT THE INFORMATION FURNISHED IN PARAGRAPHS 1, 2 and 3 ABOVE IS CORRECT.

I ACCEPT THAT THE STATE MAY REJECT THE BID OR ACT AGAINST ME IN TERMS OF PARAGRAPH 6 OF PFMA SCM INSTRUCTION 03 OF 2021/22 ON PREVENTING AND COMBATING ABUSE IN THE SUPPLY CHAIN MANAGEMENT SYSTEM SHOULD THIS DECLARATION PROVE TO BE FALSE.

Signature		Date	
Position		Name of Bidder	

 GAUTENG PROVINCE PROVINCIAL TREASURY REPUBLIC OF SOUTH AFRICA	<h1>Provincial Supply Chain Management</h1>	
	Briefing Session	Page 1 of 1

BRIEFING SESSION - DECLARATION OF ATTENDANCE

RFP NUMBER			
RFP DESCRIPTION			
RFP CLOSING DATE		CLOSING TIME	

*GPT is acting as Common Service Provider or buying organisation on behalf of all Gauteng Provincial Government Customer Departments / Institutions. The goods / services are therefore required by the Customer Department / Institution, as indicated on form RFP 01.

CUSTOMER DEPARTMENT								
CUSTOMER INSTITUTION								
BRIEFING SESSION	Y		N		DATE		TIME	
VENUE								

I/We hereby declare that I/we attended the compulsory briefing session to understand the requirements of the Gauteng Provincial Government to supply all or any of the supplies and/or to render all or any of the services described in the attached RFP documents, on the terms and conditions and in accordance with the specifications stipulated in the bid documents.

I, THE UNDERSIGNED (NAME)

--

CERTIFY THAT THE INFORMATION FURNISHED AT THE SITE INSPECTION WAS UNDERSTOOD.

BIDDER OR ASSIGNEE(S) NAME		POSITION		SIGN		DATE	
-----------------------------------	--	-----------------	--	-------------	--	-------------	--

FULL COMPANY NAME	
--------------------------	--

GPG OFFICIAL NAME		POSITION		SIGN		DATE	
--------------------------	--	-----------------	--	-------------	--	-------------	--

<p>END USER STAMP</p>



GAUTENG PROVINCE
PROVINCIAL TREASURY
REPUBLIC OF SOUTH AFRICA

Provincial Supply Chain Management


Special Conditions

Page 1 of 3

RFP NUMBER	
RFP DESCRIPTION	
CUSTOMER DEPARTMENT	
CUSTOMER INSTITUTION	

THE FOLLOWING MUST ACCOMPANY YOUR BID, IF INDICATED BY "✓"

Samples	SABS /Equivalent Certificate May not be older than one (1) year, the cost of which will be for the account of the bidder.	Bidders Briefing Session
----------------	---	---------------------------------

 <p>GAUTENG PROVINCE PROVINCIAL TREASURY REPUBLIC OF SOUTH AFRICA</p>	Provincial Supply Chain Management	
	Special Conditions	Page 2 of 3

EVALUATION METHODOLOGY

Bidders must complete Compulsory documents and attach it to their tender document, failing which the tender shall not be considered for Stage 1 evaluation.

Points will be awarded in accordance with the Preferential Procurement Policy Framework Act (PPPFA)

Stage 1

Criteria for Functionality	Points
TOTAL	

NOTE: Bidders who fail to meet the above minimum requirements (Stage 1) shall be automatically eliminated


Stage 2

Criteria for Price and Specific Goals	Points
Bid Price	
Specific Goals	
TOTAL	100

Bidders are required to use the two envelope bidding system, whereby the Technical Proposal (Stage 1); Pricing and Specific Goals(Stage 2) be placed in two separate sealed envelopes marked:

- Stage One–

- Stage Two–

 GAUTENG PROVINCE PROVINCIAL TREASURY REPUBLIC OF SOUTH AFRICA	<h1>Provincial Supply Chain Management</h1>	
	Special Conditions	Page 3 of 3

SUPPLIER JOB CREATION ANALYSIS

Company Name		Date Est.	
--------------	--	-----------	--

	Permanent	Temp	SA Citizens	Other	Comments
Staff compliment at Establishment of Enterprise					
Current staff compliment					
Number of jobs to be created if Bid is successful					

- The successful bidder may be audited during the course of the contract to verify the above information.

Comments to include:

- If Job Creation is direct (by your own company) or indirect (by your supplier)
- Where the jobs created for employees that were in existing positions or unemployed? (Net Job Creation)

NOTE: Job Creation should adhere to all applicable RSA Legislation and Regulations.

=====

THIS SECTION IS FOR OFFICE USE ONLY!						
Observations	Initial Job Count	Job Creation Potential	1 st Quarter	2 nd Quarter	3 rd Quarter	4 th Quarter
Year 1						
Year 2						
Year 3						
Year 4						
Year 5						


TENDER SPECIFICATION FOR THE SUPPLY, DELIVERY, INSTALLATION, COMMISSIONING AND MAINTENANCE OF MRI MACHINES (3T) TO VARIOUS HOSPITALS IN GAUTENG

Note:

a. Weight: Each specification will have a weight of 1 or a hash (#), which indicates the level of importance, as follows:

1 - A specification with a standard level of importance that must be complied with.

- Indicates a specification with essential level of importance, where the bid offer will be disqualified, if the specification is not complied with.

b. Score: The Bid Evaluation Committee will assign a score of 0 or 1 for each specification, as follows:

0 - The specification does not comply. 1 - The specification is fully compliant.

c. Weighted and Total Score: The score will be multiplied by the weight to give the total score for each specification and be added up.

d. Bidders must comply as follows with the specification:

i. The bidders must comply fully with the specifications marked with a # weight.

ii. Bidders must obtain a total score of 76 or higher out of 105 points in respect of specifications (line) with a weight of 1.

A	B	C	D	E
	Tender specification	Weight	Comply (Yes/No)	Details of Bid Offer (Provide your answers in this Column. You are advised to be straight to the point)
A	Modality Details			
A 1	MAGNET			
A1.1	The Magnet Field Strength must be 3 Tesla active shielded	#		
A1.2	The Magnet shall be short bore cylindrical type and design	#		
A1.3	Bore length including all covers shall not exceed 180 cm	1		
A1.4	State internal bore dimensions (L x W x H) not less than 70 cm	1		
A1.4.1	State Maximum Vision Angle	1		
A1.5	Magnet homogeneity shall meet the following specifications using the standard deviation VRMS (volume root-meansquare) using 24 plane plot measurement method:			
A1.5.1	50 cm DSV state	1		
A1.5.2	48 cm DSV state	1		
A1.5.3	45 cm DSV state	1		
A1.5.4	40 cm DSV state	1		
A1.5.5	30 cm DSV state	1		
A1.5.6	20 cm DSV state	1		
A1.5.7	10 cm DSV state	1		
A1.6	Magnetic field stability shall be less than 0.1 ppm/hour or better	1		
A1.7	The 5 Gauss/0.5mT fringe field shall be contained in an area of typically 2.4 m (radial) by 4.6m (axial). State actual area	1		


TENDER SPECIFICATION FOR THE SUPPLY, DELIVERY, INSTALLATION, COMMISSIONING AND MAINTENANCE OF MRI MACHINES (3T) TO VARIOUS HOSPITALS IN GAUTENG

A1.8	State helium boil-off details, where applicable:			
A1.8.1	State typical cryogen refilling interval	1		
A1.8.2	State maximum helium capacity.State latest technology/state helium capacity options /state latest cooling technology	1		
A1.8.3	State the weight of magnet with cryogenes, table and covers (kg) in full operation	1		
A1.8.4	The bore diameter shall be a minimum of 70cm measured at the center in an operational mode	#		
A1.8.5	Min boil off rate should not be more than 0.02 1/hr under normal operating conditions	#		
A1.9	The tenderer must guarantee and ensure that there is no magnet vibration of the system, on the current location, with all possible imaging sequences. Give details.	#		
A1.9.1	State space required (Magnet, Electronics,UPSs and Control room). State type of magnet controllers-must be digital	1		
B	MAGNET SAFETY			
B.1	Magnet must be equipped with quench exhaust leading to the outside of the building in event of Magnet quench to prevent injury to staff and patient/depending on type of cooling system available/state latest technology, vent pipe requirements state,if applicable	1		
B.2	Magnet must be equipped with emergency ramp down unit for fast Magnetic Field reduction.	#		
B.3	Magnet must be equipped with ramp up procedure	#		
B.4	MRI Signage / Floor Sticker Warning Signs/ MRI Symbols Sign/ MRI Zone Signs Complete Set. "Do Not Enter" Floor Cones	#		
B.5	Detailed attachment explaining B.1 to B.4	1		
C	GRADIENT SYSTEM			
C.1	Actively shielded hi-performance non-resonant gradient coil system is supplied.	#		
C.2	Gradient amplitude/peak strength must be at least 45mT/m measured per real axis plateau (100% duty cycle) not effective gradients: Sate options of highest gradients/Higher gradients will be an advantage.	1		
C.3	Gradient duty cycle shall be 100%	#		
C.4	Minimum rise time from 0 to max. gradient amplitude shall be 240 micro-seconds or better	1		


TENDER SPECIFICATION FOR THE SUPPLY, DELIVERY, INSTALLATION, COMMISSIONING AND MAINTENANCE OF MRI MACHINES (3T) TO VARIOUS HOSPITALS IN GAUTENG

C.5	Gradient slew rate must be 200 m/mT/ms or more measured per real axis plateau, not effective values .State options of highest gradient	1		
C.5.1	Shortest TE (SSh DWI, b=1000, 128 matrix)	#		
C.5.2	Shortest TE / TR (2D, 128 matrix) : 2D GRE	#		
C.5.3	Shortest TE / TR (2D, 128 matrix) : 3D GRE	#		
C.5.4	Shortest TE / TR (2D, 128 matrix) : 2D GRE EPI	#		
C.5.5	Shortest TE / TR (2D, 256 matrix) : 2D GRE	#		
C.5.6	Shortest TE / TR (2D, 256 matrix) : 3D GRE	#		
C.5.7	Shortest TE / TR (2D, 256 matrix) : 2D GRE EPI	#		
C.5.8	Maximum scan matrix must be minimum 1024 or better	1		
C.5.9	Maximum recon matrix must be stated	1		
C.5.10	State Increment steps for matrix	1		
C.5.11	State maximum number of slices	1		
C.5.12	State minimum slice thickness (3D)	1		
C.5.13	Gradient linearity at 10 cm DSV	#		
C.5.14	Gradient linearity at 20 cm DSV	#		
C.5.15	Gradient linearity at 30 cm DSV	#		
C.5.16	Gradient linearity at 40 cm DSV	#		
C.5.17	Gradient linearity at 50 cm DSV	#		
C.5.18	The reconstruction speed at 100% FOV must not be less than 11000 reconstructions /sec.	1		
C.6	Gradient upgrades to different levels should be possible without changing gradient coil.	1		
C.6.1	State high-end gradient specifications for optional gradient upgrade.	1		
C.6.2	Max slew rate and amplitude and FOV at max slew rate and amplitude - FOV not less than 50cm.	1		
C.6.3	Cooling for gradient coil; state type of coolant	1		
C.7	The output linearity of the gradient amplifiers should be minimum + 0.1% of peak	1		
C.8	A child friendly scanning environment is required. Paediatric noise reduction accessories.State options	1		
C.9	State advanced ergonomic features that characterize the configuration offered.	1		
C.10	Noise should be kept to a minimum.Best acoustic noise reduction technology to be offered. State details	1		
C.11	Free choice of flip angle while maintaining signal to noise ratio to be supported. Specify.	1		


TENDER SPECIFICATION FOR THE SUPPLY, DELIVERY, INSTALLATION, COMMISSIONING AND MAINTENANCE OF MRI MACHINES (3T) TO VARIOUS HOSPITALS IN GAUTENG

D	RADIOFREQUENCY SYSTEM			
D.1	Resonance frequency shall be 127MHz (3 Tesla)	1		
D.1.1	Direct Radio Frequency (RF) Transmitter System: Digital signal generation and processing.	1		
D.1.2	The transmission system to be integrated in the magnet housing	1		
D.1.3	Power output of transmitter amplifier rating shall be 35kW or better	1		
D.1.4	State the bandwidth of the RF transmitter	1		
D.1.5	The system shall be equipped with RF fault protection limiting RF output in event of malfunction	#		
D.1.6	Frequency resolution of the RF synthesizer shall be 0.07 Hz /bit or more	1		
D.2	The receive components shall be integrated into the magnet housing	#		
D.2.1	Phased array acquisition system with minimum of sixteen independent digital RF receiver channels or more, if applicable to the system. State commercial available channel upgrade path and commercial available coils utilizing the number of channels. Quote as optional items by the vendors.	1		
D.2.2	The system should have at least 48 independent RF receiver channels or better, if applicable to the system	1		
D.2.3	Phase resolution shall be 0.1 degree/bit or better	1		
D.2.4	Noise of the preamplifier to be ≤ 0.5 decibels	1		
D.3	Transmitter amplitude shall be 16 bit control or more	1		
D.3.1	Can all of these coil elements be seamlessly used in one study without coil changes and without repositioning of the patient.	1		
D.4	State maximum number of simultaneously connected coil elements. Preference will be given to highest number. Indicate coil combination.	1		
D.4.1	Phase resolution shall be 0.1 degree/bit or more.	1		
D.5	The Digital receiver should use at least 16 bit 1Mhz analogue to digital converter for each receiver channel. State sampling method, sampling rate and bandwidth per channel. Receiver signal resolution should be 32 bit	1		
D.6	State number of RF cabinets required	1		
E	RF COILS			
E.1	The bidder must supply the latest Integrated Coil Technology . Please state how the Coil Philosophy of the unit contributes and improve the Image Quality and Workflow.	1		
E.2	Integrated circular polarised / quadrature transmit and receive body coil.	#		
E.3	Minimum diameter of 70cm is required . The bore to be measured at centre of the system and not on the flare.	#		
E.4	Connection of two or more phased array coils for simultaneous use.	#		
E.5	Connected coils must be detected automatically	#		
E.6	It must be possible to select active coil or elements from the main console	#		
E.7	Coil pre-amplifiers shall be on the patient table connector and coils shall be interchangeable and be of light construction.	1		
E.8	Maximum coils connection to do full body scan. Give details of coils and connections (short cable coil connections preferred)	#		


TENDER SPECIFICATION FOR THE SUPPLY, DELIVERY, INSTALLATION, COMMISSIONING AND MAINTENANCE OF MRI MACHINES (3T) TO VARIOUS HOSPITALS IN GAUTENG

E.9	Standard coils for the following applications should be available with the system :			
E.9.1	Body (Chest, abdomen, pelvis).	#		
E.9.2	Tempo Mandibula Joint (TMJ) kit.	1		
E.9.3	Head	#		
E.9.4	Head-neck	#		
E.9.5	C/T/L spine (State what coil is used for imaging of the total spine cervical, thoracic and lumbar)	#		
E.9.6	Whole spine.	#		
E.9.7	Neck soft tissue.	#		
E.9.8	Extremities and joints: wrist, hips, shoulder, knee, foot/ankle,	#		
E.9.9	Breast . State solution for Breast MR with optional interventional capabilities. Quote interventional capabilities as optional.	#		
E.9.10	Prostate, colon, cervix.	#		
E.9.11	Whole body vascular (Peripheral vascular array).	#		
E.9.12	Peripheral vascular.	#		
E.9.13	Neurovascular.	#		
E.9.14	Dedicated cardiac.	#		
E.9.15	Dedicated pediatric coils.	#		
E.9.16	Supply a cupboard in the room for storage of all coils provided or alternatively sufficient coil cabinets to house the coils.	#		
E.10	Other features:			
E.10.1	Coils to facilitate optimal facial, calvarial and intracranial imaging in the child from 1.5-60kgs, including visualisation of the orbits, optic nerves, optic tracts, pituitary fossa, hypothalamus, temporal lobe, posterior fossa and cerebral cortex. Coils to provide optimal imaging including, but not limited to the following clinical settings:	1		
F	PATIENT SUPPORT/TABLE and MANAGEMENT/ PATIENT COMFORT			
F.1	Patient table must be dockable/detachable. Dockable patient transport system for simplified patient preparation, handling and transportation from preparation room to the MR scanner, without repositioning the patient.	#		
F.2	Two dockable tables must be included in the total price	#		
F.2.1	Two comfortable patient mattresses to be included	#		
F.3	Dual table control panels shall be located at either side of aperture/gantry for easy access. State if controls are available on the rear of the magnet	#		
F.4	Centering laser light beams for anatomical references.	1		
F.5	Table movement shall be controlled from both gantry and operator console	#		
F.6	Patient table shall be equipped with manual override for quick removal of patient from the magnet-bore in case of emergency	#		
F.7	Maximum load must be minimum of 200kg or better. State actual maximum load.	#		


TENDER SPECIFICATION FOR THE SUPPLY, DELIVERY, INSTALLATION, COMMISSIONING AND MAINTENANCE OF MRI MACHINES (3T) TO VARIOUS HOSPITALS IN GAUTENG

F.8	State maximum weight for vertical and horizontal table movement:	1		
F.8.1	Vertical table movement:			
F.8.1.1	Minimum table height from floor	1		
F.8.1.2	Maximum table height from floor	1		
F.8.1.3	Maximum vertical speed	1		
F.8.2	Horizontal table movement:			
F.8.2.1	Maximum horizontal range of table movement	1		
F.8.2.2	Maximum speed:	1		
F.8.2.3	Accuracy of repositioning	1		
F.9	Physiologic measurement unit essential with display of ECG, respiration and pulse at the main console and gantry.	#		
F.10	Two way in-bore intercom system shall allow communication with patient while gradient is running	#		
F.11	Hand held alarm button for patient signalling	#		
F.12	Table restraining/immobilization straps are required per table	#		
F.13	Additional height adjustable patient transport stretcher –MR compatible max load 250kg to be supplied.	1		
F.14	Paediatric friendly environment (Bidder to provide details).	1		
F.15	Integrated music for patient shall be included.	1		
F.16	Variable patient lighting to be included.	1		
F.17	State of the art technology is required to allow young patients to see outside the magnet during the examination. Give details	1		
F.18	MRI compatible sand bags and sponges in various sizes and shapes.	#		
F.19	Ventilation in the gantry /Fresh air supply	#		
F.20	Auto-voice	#		
G	WORKSTATION:DOCTORS CONSOLE			
G.1	Two dedicated advanced post processing workstations will be supplied, with extended memory.	#		
G.2	The dedicated workstation will be used for all post processing. The system must include licenses for all 3D reconstruction and post processing requirements. Please give details and options where applicable.	#		
G.3	Post processing on workstation to be compatible with software requested on console.	#		
G.4	High quality read/write CD/DVD storage device required.	#		
G.5	DICOM compatible and configured to connect to a PACS system. Vendor must connect to the PACS system at no extra cost.	#		
H	OPERATOR USER INTERFACE			
H.1	One acquisition console with additional post processing workstation to be supplied.	#		
H.2	Two 23 inch or larger hi-resolution colour LCD flat-panel flicker-free monitor with undistorted image display required.	#		
H.3	Monitor 1024 x 1024 pixel resolution required or better	#		
H.4	User interface providing flexible multi-tasking in foreground or background (scanning, filming, reconstruction).	#		
H.5	Four TB external hard drive as storage devices required	1		


TENDER SPECIFICATION FOR THE SUPPLY, DELIVERY, INSTALLATION, COMMISSIONING AND MAINTENANCE OF MRI MACHINES (3T) TO VARIOUS HOSPITALS IN GAUTENG

H.6	All standard image processing features are required in addition the following advanced features must be available:			
H.6.1	Multi-planner reformatting (MPR).	#		
H.6.2	Multi-projection volume rendering (MPVR).	#		
H.6.3	Maximum intensity projection (MIP).	#		
H.6.4	MR Angiography processing.	#		
H.6.5	3D surface rendering.	#		
H.6.6	MR Hydrography MRCP/Urography/Myelography.	#		
H.6.7	Image add/subtract.	#		
H.6.8	CT image display.	#		
H.6.9	CT image integrating must be offered.	#		
H.6.10	State any special features to improve productivity and consistency (e.g. automated scanning procedures) i.e. qualitative assessment.	1		
I	ACQUISITION PARAMATERS, APPLICATION SOFTWARE AND IMAGING TECHNIQUES			
I.1	Please explain the different techniques available to reduce motion artefacts to facilitate paediatric imaging.	1		
I.2	Display slice and slab thickness and left and right designation:			
I.2.1	2D	#		
I.2.2	3D	#		
I.3	State number of slices:			
I.3.1	2D	#		
I.3.2	3D	#		
I.4	Variable field of view required to maximum of 48cm or better. State details	1		
I.5	Standard/Conventional and fast imaging techniques/ Sequences required for all the following applications with dedicated post processing:			
a	State standard sequences in offered software package	#		
b	State maximum parallel imaging factor possible for offered software package (and compatible coils)	#		
c	State sequences that are compatible with parallel imaging	#		
d	List optional sequences and software packages, in addition to what is listed below. State details and costs.	1		
I.5.1	Neuro:			
i	Brain	#		
ii	Conventional imaging sequences	#		
iii	Spectroscopy (Single voxel and multi voxel). State if compatible with parallel imaging 2D / 3D	#		
iv	Diffusion weighted	#		
v	Perfusion weighted. Sate sequences used.	#		
vi	Fast MRI sequences. Sate sequences used.	#		
vii	Tractography / Fibre tracking	1		
viii	Spine	#		


TENDER SPECIFICATION FOR THE SUPPLY, DELIVERY, INSTALLATION, COMMISSIONING AND MAINTENANCE OF MRI MACHINES (3T) TO VARIOUS HOSPITALS IN GAUTENG


I.5.2	Cardiac:			
i	Functional evaluation	#		
ii	Morphology	#		
iii	Valvular analysis	#		
iv	Coronary artery imaging	#		
v	Viability studies	#		
vi	ECG gating	#		
vii	Myocardial tagging	#		
I.5.3	Abdomen:			
i	Conventional sequences	#		
ii	Virtual endoscopy	#		
iii	MRCP	#		
iv	Small bowel studies	#		
v	Dynamic liver contrast studies	#		
I.5.4	Chest:			
i	Conventional sequences	#		
ii	Mammography	#		
iii	Dynamic (3D FSGRE etc.)	#		
iv	Biopsy capability	#		
v	Pulmonary ventilation studies	#		
I.5.5	Pelvis:			
i	Conventional sequences	#		
ii	Prostate imaging (including spectroscopy)	#		
iii	Dynamic pelvic floor imaging	#		
I.5.6	Vascular:			
i	TOF (2D/3D)	#		
ii	PC (2D/3D)	#		
iii	Contrast enhancement for:			
iv	Neck vessel studies	#		
v	Intracranial vessel studies	#		
vi	Aortic arch and branches	#		
vii	Abdominal aorta and outflow	#		
viii	Upper limbs	#		
ix	Pulmonary arteries	#		
x	Renal arteries	#		
xi	Multistep MR angiography	#		


TENDER SPECIFICATION FOR THE SUPPLY, DELIVERY, INSTALLATION, COMMISSIONING AND MAINTENANCE OF MRI MACHINES (3T) TO VARIOUS HOSPITALS IN GAUTENG

I.5.7	Head and neck:			
i	Conventional sequences	#		
ii	IAM's	#		
iii	TMJ's	#		
J	PERFORMANCE EVALUATION PHANTOMS REQUIRED / QUALITY CONTROL			
J.1	Performance evaluation software and appropriate phantoms and quality assurance kits required for evaluation of image quality as per licensing conditions.State details	#		
K	SAFETY			
K.1	The Magnet system shall include an emergency ramp down unit (RDU) for fast reduction of the magnetic field.	#		
K.2	The Magnet system shall include an emergency ramp down unit (RDU) for fast reduction of the magnetic field.	1		
K.3	Real-time SAR calculation shall be performed by software to ensure that RF power levels comply with regulatory guidelines and be displayed on each image.	#		
K.4	The system shall have manual override of the motor drive for quick removal of the patient from the magnet bore.	#		
K.5	The system shall have RF fault protection hardware (RF transmit enable limit and RF power and duty cycle limit) to limit The RF output in The event of equipment malfunction.	#		
K.6	Pinpoint walk-through metal detector. Dedicated Ferro-magnetic detector with a pre-warning on approach well outside MRI room, not a conventional metal detector. Highest sensitivity rating for the MRI environment.	#		
K.6.1	Hand held metal detector with different setting sensitivity.State details	1		
K.6.2	The detector shall detect but not have any effect on heart pacemakers	1		
K.7	A UPS System (Uninterrupted Power Supply) must be provided for each workstation and should be able to run for at least 15 minutes. Must be included in the tender price.	#		
K.7.1	State details of kVA of UPS provided for the workstations.	1		
K.8	A UPS System (Uninterrupted Power Supply) must be provided for the MRI Unit and included in the tender price.State details of kVA of UPS provided for the unit. UPS must be 160kva or higher for the complete MRI system with sealed MF batteries for 30min back-up to support the RF amplifier and the gradients during a power failure. must be compatible with the MRI system to include a 200kVa isolation transformer	#		
K.9	Give details of comprehensive magnet safety and MRI safety /include details of isolation transformer,earthing details,anti-theft cables,UPS and stability of power supply requirements	#		
K.10	Air-conditioning: Water chiller with filter connected to UPS	#		



TENDER SPECIFICATION FOR THE SUPPLY, DELIVERY, INSTALLATION, COMMISSIONING AND MAINTENANCE OF MRI MACHINES (3T) TO VARIOUS HOSPITALS IN GAUTENG

L	ADDITIONAL EQUIPMENT			
L.1	Extra patient monitoring screen at the console.State details	1		
L.2	Please provide the specifications of the unit on offer .	1		
L.3	MRI integrated pulse oximeter for adults and children. This unit must have probes for adult / paediatrics & neonates.	#		
M	MR FLUROSCOPY (OPTIONAL)			
M.1	A MR Fluoroscopy mode with real-time acquisition and display techniques is required as an option. Please State the soft and hardware required if this is optional.	1		
M.2	State AI technologies available	1		
N	STANDARD ACCESSORIES			
N.1	MRI compatible accessories for Neonates,Paediatrics and Adults.State details.	#		
N.2	MRI compatible Incubator for Neonatal Imaging.	#		
N.3	MRI compatible contrast injector interphased with MR unit.	#		
N.4	200 syringes and 200 matching syringe lines.	#		
N.5	Drug Infusion pump compatible to MRI.	#		
N.6	List of standard accessories included but not specified in the tender.	1		
N.7	Patient Monitoring Video Camera System. Direct connection to monitor. All parts in the MRI room must be MRI compatible	#		
O	TECHNOLOGY			
O.1	Innovation. Give explanation	1		
O.2	Unique. Give explanation	1		
O.3	Clinical benefits. Give detailed explanation	1		
P	OPTIONAL ACCESSORIES			
P.1	The tenderer to give full description and pricing of optional accessories available for the system. Each item must be priced separately.Quote prices	1		
P.2	MRI compatible laryngoscopes, a fully equipped MRI compatible emergency trolley and MRI compatible Stethoscope.	1		
P.3	Scale weighing w/large dial type 2 150kg	1		
P.4	Patient slide for patient transfer 200g. (PVC Body Transfer Board and PVC Body Transfer Board Storage Brackets)	1		
P.5	MRI compatible wheelchair with footrests	1		
P.6	MRI Heavy Duty / Extra Wide Folding Walker.	1		
P.7	MRI Transfer / Gait Belt.	1		
P.8	MRI compatible steps / MRI Double Step Stool with Handrail	1		

				
TENDER SPECIFICATION FOR THE SUPPLY, DELIVERY, INSTALLATION, COMMISSIONING AND MAINTENANCE OF MRI MACHINES (3T) TO VARIOUS HOSPITALS IN GAUTENG				
P.9	I.V. Pole, MRI compatible	1		
P.10	Plastic chairs in magnet room	1		
P.11	Chair operator high back swivel with arm rests.	1		
P.12	Chair draughtsman mid armrests	1		
P.13	Shelve loose storage and display, wood.	#		
P.14	MRI Compatible Fire Extinguisher	1		
P.15	Hoist mobile electric hydraulic.	#		
P.16	One scanner cum copier and color printer. (Fax/printer/copier/scanner)	#		
P.17	Dicom color printer / camera for 8"x10" paper/transparency color printing, 500 transparencies to be included.	#		
P.18	Modem for remote service support.	1		
Q	MRI COMPATIBLE ANAESTHETIC UNIT			
Q.1	MRI Compatible Anaesthetic unit for General Anaesthetics of the paediatric patients is required. Please state the specifications of the unit on offer. Please also note the warranty and customer support agreement terms.	#		
Q.2	MRI compatible anaesthetic machine with ECG, NIBP, SPO2 & Gas capnography to suit adults, pediatric and neonates. ECG: Echo Cardiogram / NIBP: Non invasive BP / SPO2: Oxygen saturation / Gas Capnography: CO2 measurement	#		
Q.3	Anaesthesia machine must have a scavenger and the room must have an outlet for the scavenger for anaesthetic gases to be removed from the magnet room	#		


TENDER SPECIFICATION FOR THE SUPPLY, DELIVERY, INSTALLATION, COMMISSIONING AND MAINTENANCE OF MRI MACHINES (3T) TO VARIOUS HOSPITALS IN GAUTENG

R	LASER CAMERA			
R.1	A state of the art dry laser camera to be included. Specify: Make Model	1		
R.2	UPS for the dry laser camera to be included	#		
S	GENERAL ISSUES			
S.1	IT INTEGRATION			
S.1.1	The system must have the following DICOM 3 compatibility:			
S.1.1.1	DICOM Send / Receive	#		
S.1.1.2	DICOM Storage Commitment (SC):	#		
S.1.1.3	DICOM query/retrieve	#		
S.1.1.4	DICOM worklist	#		
S.1.1.5	DICOM Basic Print	#		
S.1.1.6	DICOM MPPS (Modality Performed Procedure Steps)	#		
S.2	The system must comply to the following IHE profiles (Integrating the Healthcare Enterprise)	#		
S.2.1	Scheduled Workflow (SWF)	1		
S.2.2	Patient Information Reconciliation (PIR)	1		
S.2.3	Simple Image and Numeric Reports (SINR)	1		
S.2.4	System to be integrated to the existing PACS	#		
	TOTAL:	105		



GAUTENG PROVINCE
HEALTH
REPUBLIC OF SOUTH AFRICA

TENDER SPECIFICATION FOR THE SUPPLY, DELIVERY, INSTALLATION, COMMISSIONING AND MAINTENANCE OF MRI MACHINES (1.5 T) AT VARIOUS HOSPITALS IN GAUTENG

Note:

a. Weight: Each specification will have a weight of 1 or a hash (#), which indicates the level of importance, as follows:

1 - A specification with a standard level of importance that must be complied with.

- Indicates a specification with essential level of importance, where the bid offer will be disqualified, if the specification is not complied with.

b. Score: The Bid Evaluation Committee will assign a score of 0 or 1, for each specification, as follows:

0 - The specification does not comply. 1 - The specification is fully compliant.

c. Weighted and Total Score: The score will be multiplied by the weight to give the total score for each specification and be added up.

d. The bidders must comply as follows with the specification:

i. Bidders must comply fully with the specifications marked with a # weight.

ii. Bidders must obtain a total score of 50 or higher out of 102 points in respect of specifications (line) with a weight of 1.

A	B	C	D	E	F
	Tender Specification	Weight	Comply (Yes/No)	Details of Bid Offer (Provide your answers in this Column. You are advised to be straight to the point)	Reference to Brochure / Technical Data Sheet and provide page numbers for verification
A	Modality Details				
A 1	MAGNET				
A1.1	The Magnet Field Strength shall be at least 1.5 Tesla active shielded.	#			
A1.2	The Magnet must be short closed bore cylindrical type and design.	#			
A1.3	Bore length including all covers shall not exceed 180 cm	1			
A1.4	State internal bore dimensions (L x W x H) not less than 70cm inner magnet.	1			
A1.5	Magnet homogeneity should meet the following specifications using the standard deviation VRMS (volume root-meansquare) using 24 plane plot measurement method:				
A1.5.1	50 cm DSV (Diameter Spherical Volume) state (at most 5ppm)	1			
A1.5.2	45 cm DSV state (at most 1.2 ppm)	1			
A1.5.3	40 cm DSV state (at most 1 ppm)	1			
A1.5.4	30 cm DSV state (at most 0.5 ppm)	1			
A1.5.5	20 cm DSV state (at most 0.1 ppm)	1			
A1.5.6	10 cm DSV state (at most 0.02 ppm)	1			


TENDER SPECIFICATION FOR THE SUPPLY, DELIVERY, INSTALLATION, COMMISSIONING AND MAINTENANCE OF MRI MACHINES (1.5 T) AT VARIOUS HOSPITALS IN GAUTENG

A 1.6	Magnetic field stability shall be less than 0.1 ppm/hour or better	1			
A 1.7	The 5 Gauss/0.5mT fringe field must be contained in an area of typically 2.4 m (radial) by 4.0m (axial). State actual area.	1			
A 1.8	State helium boil-off details, where applicable:				
A1.8.1	State typical cryogen refilling interval	1			
A1.8.2	State maximum helium capacity.State latest technology/state helium capacity options /state latest cooling technology	1			
A1.8.3	State the weight of magnet with cryogenes, table and covers (kg) in full operation	1			
A1.8.4	The bore diameter shall be a minimum of 70cm measured at the center in an operational mode	#			
A1.8.5	Min boil off rate should not be more than 0.02 1/hr under normal operating conditions	#			
A1.9	The tenderer must guarantee and ensure that there is no magnet vibration of the system, on the current location, with all possible imaging sequences. Give details.	#			
A1.9.1	State space required (Magnet, Electronics,UPSs and Control room). State type of magnet controllers-must be digital	1			
B.	MAGNET SAFETY				
B.1	Magnet must be equipped with quench exhaust leading to the outside of the building in event of Magnet quench to prevent injury to staff and patient/depending on type of cooling system available/state latest technology, vent pipe requirements state,If applicable	1			
B.2	Magnet must be equipped with emergency ramp down unit for fast Magnetic reduction, state details.	#			
B.3	Magnet must be equipped with ramp up procedure	#			
B.4	MRI Signage / Floor Sticker Warning Signs/ MRI Symbols Sign/ MRI Zone Signs Complete Set. "Do Not Enter" Floor Cones	#			
B.5	Detailed attachment explaining B.1 to B.4	1			
C	GRADIENT SYSTEM (Provide the state of the art and most recent technology details)				
C.1	Actively shielded hi-performance non-resonant gradient coil system is supplied.	#			
C.2	Gradient amplitude/peak strength must be at least 33mT/m measured per real axis plateau (100% duty cycle) not effective gradients: Sate options of highest gradients/Higher gradients will be an advantage.	1			
C.3	Gradient duty cycle shall be 100%.	#			
C.4	Gradient rise time from 0 to max gradient amplitude must be 240 micro-seconds or more.	1			


TENDER SPECIFICATION FOR THE SUPPLY, DELIVERY, INSTALLATION, COMMISSIONING AND MAINTENANCE OF MRI MACHINES (1.5 T) AT VARIOUS HOSPITALS IN GAUTENG

C.5	Gradient slew rate must be 120 m/mT/ms or more measured per real axis plateau, not effective values .State options of highest gradient	1			
C.5.1	Shortest TE (SSh DWI, b=1000, 128 matrix)	#			
C.5.2	Shortest TE / TR (2D, 128 matrix) : 2D GRE	#			
C.5.3	Shortest TE / TR (2D, 128 matrix) : 3D GRE	#			
C.5.4	Shortest TE / TR (2D, 128 matrix) : 2D GRE EPI	#			
C.5.5	Shortest TE / TR (2D, 256 matrix) : 2D GRE	#			
C.5.6	Shortest TE / TR (2D, 256 matrix) : 3D GRE	#			
C.5.7	Shortest TE / TR (2D, 256 matrix) : 2D GRE EPI	#			
C.5.8	Maximum scan matrix must be minimum 1024 or better	1			
C.5.9	Maximum recon matrix must be stated	1			
C.5.10	State Increment steps for matrix	#			
C.5.11	State Maximum number of slices	1			
C.5.12	State Minimum slice thickness (3D)	1			
C.5.13	Gradient linearity at 20 cm DSV	#			
C.5.14	Gradient linearity at 50 cm DSV	#			
C.5.15	Gradient linearity at 30 cm DSV	#			
C.5.16	Gradient linearity at 40 cm DSV	#			
C.5.17	Gradient linearity at 50 cm DSV	#			
C.5.18	The reconstruction speed at 100% FOV must not be less than 11000 reconstructions /sec.	1			
C.6	Gradient upgrades to different levels should be possible without changing gradient coil.	1			
C.6.1	State high-end gradient specifications for optional gradient upgrade.	1			
C.6.2	Max slew rate and amplitude and FOV at max slew rate and amplitude - FOV not less than 50cm.	1			
C.6.3	Cooling for gradient coil; state type of coolant.	1			
C.7	The output linearity of the gradient amplifiers should be minimum of + 0.1% of peak.	1			
C.8	A child friendly scanning environment is required. Paediatric noise reduction accessories. State options	1			
C.9	State advanced ergonomic features that characterize the configuration offered.	1			
C.10	Noise should be kept to a minimum.	1			
C.11	Free choice of flip angle while maintaining signal to noise ratio to be supported. Specify.	1			


TENDER SPECIFICATION FOR THE SUPPLY, DELIVERY, INSTALLATION, COMMISSIONING AND MAINTENANCE OF MRI MACHINES (1.5 T) AT VARIOUS HOSPITALS IN GAUTENG

D.	RADIOFREQUENCY SYSTEM				
D.1	Resonance frequency must be 63.86MHz (1.5 Tesla).	1			
D.1.1	Direct Radio Frequency (RF) Transmitter System: Digital signal generation and processing.	1			
D.1.2	The transmission system to be integrated in the magnet housing.	1			
D.1.3	Power output of transmitter amplifier rating must be 18kW or better	1			
D.1.4	State the bandwidth of the RF transmitter.	1			
D.1.5	The system must be equipped with RF fault protection limiting RF output in event of malfunction.	#			
D.1.6	Frequency resolution of the RF synthesizer shall be 0.07 Hz /bit or more	1			
D.2	The receive components shall be integrated into the magnet housing	#			
D.2.1	Phased array acquisition system with minimum of sixteen independent digital RF receiver channels or more, if applicable to the system. State commercial available channel upgrade path and commercial available coils utilizing the number of channels. Quote as optional items by the vendors.	1			
D.2.2	The system should have at least 48 independent RF receiver channels or better, if applicable to the system	1			
D.2.3	Phase resolution shall be 0.1 degree/bit or better	1			
D.2.4	Noise of the preamplifier to be ≤ 0.5 decibels	1			
D.3	Transmitter amplitude shall be 16 bit control or more	1			
D.3.1	Can all of these coil elements be seamlessly used in one study without coil changes and without repositioning of the patient.	1			
D.4	State maximum number of simultaneously connected coil elements. Preference will be given to highest number. Indicate coil combination.	1			
D.4.1	Phase resolution shall be 0.1 degree/bit or more.	1			
D.5	The Digital receiver should use at least 16 bit 1Mhz analogue to digital converter for each receiver channel. State sampling method, sampling rate and bandwidth per channel.	1			
D.6	State number of RF cabinets required	1			


TENDER SPECIFICATION FOR THE SUPPLY, DELIVERY, INSTALLATION, COMMISSIONING AND MAINTENANCE OF MRI MACHINES (1.5 T) AT VARIOUS HOSPITALS IN GAUTENG


E.	RF COILS				
E.1	The bidder must supply the latest Integrated Coil Technology. Please state how the Coil Philosophy of the unit contributes and improve the Image Quality and Workflow.	1			
E.2	Integrated circular polarised / quadrature transmit and receive body coil.	#			
E.3	Minimum diameter of 70cm is required . The bore to be measured at centre of the system and not on the flare.	#			
E.4	Connection of two or more phased array coils for simultaneous use.	#			
E.5	Connected coils must be detected automatically.	#			
E.6	It must be possible to select active coil or elements from the main console.	#			
E.7	Coil pre-amplifiers shall be on the patient table connector and coils shall be interchangeable and be of light construction.	1			
E.8	Maximum coils connection to do full body scan. Give details of coils and connections (short cable coil connections preferred)	#			
E.9	Standard coils for the following applications should be available with the system :				
E.9.1	Body (Chest, abdomen, pelvis).	#			
E.9.2	Tempo Mandibula Joint (TMJ) kit.	1			
E.9.3	Head	#			
E.9.4	Head-neck	#			
E.9.5	C/T/L spine (State what coil is used for imaging of the total spine cervical, thoracic and lumbar)	#			
E.9.6	Whole spine.	#			
E.9.7	Neck soft tissue.	#			
E.9.8	Extremities and joints: wrist, hips, shoulder, knee, foot/ankle,	#			
E.9.9	Breast . State solution for Breast MR with optional interventional capabilities. Quote interventional capabilities as optional.	#			
E.9.10	Prostate, colon, cervix.	#			
E.9.11	Whole body vascular (Peripheral vascular array).	#			
E.9.12	Peripheral vascular.	#			
E.9.13	Neurovascular.	#			
E.9.14	Dedicated cardiac.	#			
E.9.15	Dedicated pediatric coils.	#			
E.9.16	Supply a cupboard in the room for storage of all coils provided or alternatively sufficient coil cabinets to house the coils.	#			


TENDER SPECIFICATION FOR THE SUPPLY, DELIVERY, INSTALLATION, COMMISSIONING AND MAINTENANCE OF MRI MACHINES (1.5 T) AT VARIOUS HOSPITALS IN GAUTENG

E.10	Other features:				
E.10.1	Coils to facilitate optimal facial, calvarial and intracranial imaging in the child from 1.5-60kgs, including visualisation of the orbits, optic nerves, optic tracts, pituitary fossa, hypothalamus, temporal lobe, posterior fossa and cerebral cortex. Coils to provide optimal imaging including, but not limited to the clinical settings.	#			
F	PATIENT SUPPORT/TABLE and MANAGEMENT/ PATIENT COMFORT				
F.1	Patient table must be dockable/detachable. Dockable patient transport system for simplified patient preparation, handling and transportation from preparation room to the MR scanner, without repositioning the patient.	#			
F.2	Two dockable tables must be included in the total price	#			
F.3	Dual table control panels shall be located at either side of aperture/gantry for easy access. State if controls are available on the rear of the magnet.	#			
F.4	Centering laser light beams for anatomical references.	1			
F.5	Table movement shall be controlled from both gantry and operator console.	#			
F.6	Patient table shall be equipped with manual override for quick removal of patient from the magnet-bore in case of emergency.	#			
F.7	Maximum load must be minimum of 200kg or better. State actual maximum load.	#			
F.8	maximum weight for vertical and horizontal table movement:				
F.8.1	Vertical table movement:				
F.8.1.1	Minimum table height from floor	1			
F.8.1.2	Maximum table height from floor	1			
F.8.1.3	Maximum vertical speed	1			
F.8.2	Horizontal table movement:				
F.8.2.1	Maximum horizontal range of table movement	1			
F.8.2.2	Maximum speed:	1			
F.8.2.3	Accuracy of repositioning	1			



TENDER SPECIFICATION FOR THE SUPPLY, DELIVERY, INSTALLATION, COMMISSIONING AND MAINTENANCE OF MRI MACHINES (1.5 T) AT VARIOUS HOSPITALS IN GAUTENG

F.9	Physiologic measurement unit essential with display of ECG, respiration and pulse at the main console and gantry.	#			
F.10	Two way in-bore intercom system shall allow communication with patient while gradient is running	#			
F.11	Hand held alarm button for patient signalling	#			
F.12	Table restraining/immobilization straps are required per table	#			
F.13	Additional height adjustable patient transport stretcher –MR compatible max load 250kg to be supplied.	1			
F.14	Paediatric friendly environment (Bidder to provide details).	1			
F.15	Integrated music for patient shall be included.	1			
F.16	Variable patient lighting to be included.	1			
F.17	State of the art technology is required to allow young patients to see outside the magnet during the examination. Give details	1			
F.18	MRI compatible sand bags and sponges in various sizes and shapes.	#			
F.19	Ventilation in the gantry /Fresh air supply	#			
F.20	Auto-voice	#			
G	WORKSTATION: DOCTORS CONSOLE				
G.1	Two dedicated advanced post processing workstations will be supplied, with extended memory.	#			
G.2	The dedicated workstation will be used for all post processing. The system must include licenses for all 3D reconstruction and post processing requirements. Please give details and options where applicable.	#			
G.3	Post processing on workstation to be compatible with software requested on console.	#			
G.4	High quality read/write CD/DVD storage device required.	#			
G.5	DICOM compatible and configured to connect to a PACS system. Vendor must connect to the PACS system at no extra cost.	#			
G.6	DICOM compatible and configured to connect to a PACS system. Vendor must connect to the PACS system at no extra cost.	#			
H	OPERATOR USER INTERFACE				
H.1	One acquisition console with additional post processing workstation to be supplied.	#			
H.2	Two 23 inch or larger hi-resolution colour LCD flat-panel flicker-free monitor with undistorted image display required.	#			
H.3	Monitor 1024 x 1024 pixel resolution required or better	#			
H.4	User interface providing flexible multi-tasking in foreground or background (scanning, filming, reconstruction).	#			
H.5	Four TB external hard drive as storage devices required	1			

 GAUTENG PROVINCE <small>HEALTH REPUBLIC OF SOUTH AFRICA</small>					
TENDER SPECIFICATION FOR THE SUPPLY, DELIVERY, INSTALLATION, COMMISSIONING AND MAINTENANCE OF MRI MACHINES (1.5 T) AT VARIOUS HOSPITALS IN GAUTENG					
H.6	All standard image processing features are required in addition the following advanced features must be available:				
H.6.1	Multi-planner reformatting (MPR).	#			
H.6.2	Multi-projection volume rendering (MPVR).	#			
H.6.3	Maximum intensity projection (MIP).	#			
H.6.4	MR Angiography processing.	#			
H.6.5	3D surface rendering.	#			
H.6.6	MR Hydrography MRCP/Urography/Myelography.	#			
H.6.7	Image add/subtract.	#			
H.6.8	CT image display.	#			
H.6.9	CT image integrating must be offered.	#			
H.6.10	State any special features to improve productivity and consistency (e.g. automated scanning procedures) i.e. qualitative assessment.	1			
I	ACQUISITION PARAMATERS, APPLICATION SOFTWARE AND IMAGING TECHNIQUES				
I.1	Please explain the different techniques available to reduce motion artefacts to facilitate paediatric imaging.	1			
I.2	Display slice and slab thickness and left and right designation:				
I.2.1	2D.	#			
I.2.2	3D.	#			
I.3	Display number of slices:				
I.3.1	2D.	#			
I.3.2	3D.	#			
I.4	Variable field of view required to maximum of 48cm or better. State.	#			
I.5	Standard/Conventional and fast imaging techniques/ Sequences required for all the following applications with dedicated post processing:				
a	List standard sequences in offered software package.	#			
b	List maximum parallel imaging factor possible for offered software package (and compatible coils).	#			
c	List sequences that are compatible with parallel imaging.	#			
d	List optional sequences and software packages, in addition to what is listed below. State details and costs.	#			


TENDER SPECIFICATION FOR THE SUPPLY, DELIVERY, INSTALLATION, COMMISSIONING AND MAINTENANCE OF MRI MACHINES (1.5 T) AT VARIOUS HOSPITALS IN GAUTENG

I.5.1	Neuro:				
i	Brain.	#			
ii	Conventional imaging sequences.	#			
iii	Spectroscopy (Single voxel and multi voxel). State if compatible with parallel imaging 2D / 3D.	#			
iv	Diffusion weighted.	#			
v	Perfusion weighted. Sate sequences used.	#			
vi	Fast MRI sequences. Sate sequences used.	#			
vii	Tractography / Fibre tracking .	#			
viii	Spine.	#			
I.5.2	Cardiac:				
i	Functional evaluation.	#			
ii	Morphology.	#			
iii	Valvular analysis.	#			
iv	Coronary artery imaging.	#			
v	Viability studies.	#			
vi	ECG gating.	#			
vii	Myocardial tagging.	#			
I.5.3	Abdomen:				
i	Conventional sequences.	#			
ii	Virtual endoscopy.	#			
iii	MRCP.	#			
iv	Small bowel studies.	#			
v	Dynamic liver contrast studies & Liver DWI.	#			
I.5.4	Chest:				
i	Conventional sequences.	#			
ii	Mammography .	#			
iii	Dynamic (3D FSGRE etc.).	#			
iv	Biopsy capability.	#			
v	Pulmonary ventilation studies.	#			
I.5.5	Pelvis:				
i	Conventional sequences.	#			
ii	Prostate imaging (including spectroscopy).	#			
iii	Dynamic pelvic floor imaging.	#			

<div><div>GAUTENG PROVINCE HEALTH REPUBLIC OF SOUTH AFRICA</div></div>					
TENDER SPECIFICATION FOR THE SUPPLY, DELIVERY, INSTALLATION, COMMISSIONING AND MAINTENANCE OF MRI MACHINES (1.5 T) AT VARIOUS HOSPITALS IN GAUTENG					
I.5.6	Vascular:				
i	TOF (2D/3D).	#			
ii	PC (2D/3D).	#			
iii	Contrast enhancement for:				
iv	Neck vessel studies.	#			
v	Intracranial vessel studies.	#			
vi	Aortic arch and branches.	#			
vii	Abdominal aorta and outflow.	#			
viii	Upper & Lower limbs.	#			
ix	Pulmonary arteries.	#			
x	Renal arteries.	#			
xi	Multistep MR angiography.	#			


TENDER SPECIFICATION FOR THE SUPPLY, DELIVERY, INSTALLATION, COMMISSIONING AND MAINTENANCE OF MRI MACHINES (1.5 T) AT VARIOUS HOSPITALS IN GAUTENG


I.5.7	Head and neck:				
i	Conventional sequences.	#			
ii	IAM's (Internal Auditory Meatus).	#			
iii	TMJ's.	#			
I.6	Explain what techniques are available to reduce breathhold time.	1			
J	PERFORMANCE EVALUATION PHANTOMS REQUIRED / QUALITY CONTROL				
J.1	Performance evaluation software and appropriate phantoms and quality assurance kits required for evaluation of image quality as per licensing conditions.State details	#			
K	SAFETY				
K.1	The Magnet system shall include an emergency ramp down unit (RDU) for fast reduction of the magnetic field.	#			
K.2	The Magnet system shall include an emergency ramp down unit (RDU) for fast reduction of the magnetic field.	1			
K.3	Real-time SAR calculation shall be performed by software to ensure that RF power levels comply with regulatory guidelines and be displayed on each image.	#			
K.4	The system shall have manual override of the motor drive for quick removal of the patient from the magnet bore.	#			
K.5	The system shall have RF fault protection hardware (RF transmit enable limit and RF power and duty cycle limit) to limit The RF output in The event of equipment malfunction.	#			
K.6	Pinpoint walk-through metal detector. Dedicated Ferro-magnetic detector with a pre-warning on approach well outside MRI room, not a conventional metal detector. Highest sensitivity rating for the MRI environment.	#			
K.6.1	Hand held metal detector with different setting sensitivity.State details	1			
K.6.2	The detector shall detect but not have any effect on heart pacemakers	1			
K.7	A UPS System (Uninterrupted Power Supply) must be provided for each workstation provided and included in the tender price.	#			
K.7.1	State details of kVA of UPS provided for the workstations.	1			
K.8	A UPS System (Uninterrupted Power Supply) must be provided for the MRI Unit and included in the tender price.State details of kVA of UPS provided for the unit. UPS must be 160kva or higher for the complete MRI system with sealed MF batteries for 30min back-up to support the RF amplifier and the gradients during a power failure. must be compatible with the MRI system,to include	#			
K.9	Give details of comprehensive magnet safety and MRI safety /include details of isolation transformer,earthing details,anti-theft cables,UPS and stability of power supply requirements	#			
K.10	Air-conditioning: Water chiller with filter connected to UPS	#			


TENDER SPECIFICATION FOR THE SUPPLY, DELIVERY, INSTALLATION, COMMISSIONING AND MAINTENANCE OF MRI MACHINES (1.5 T) AT VARIOUS HOSPITALS IN GAUTENG

L	ADDITIONAL EQUIPMENT				
L.1	Extra patient monitoring screen at the console.State details	1			
L.2	Please provide the specifications of the unit on offer .	1			
L.3	MRI integrated pulse oximeter for adults and children. This unit must have probes for adult / paediatrics & neonates.	#			
M	MR FLUROSCOPY (OPTIONAL)				
M.1	a MR Fluoroscopy mode with real-time acquisition and display techniques is required as an option. Please State the soft and hardware required if this is optional.	1			
M.2	State AI technologies available	1			
N	STANDARD ACCESSORIES				
N.1	MRI compatible accessories for Neonates,Paediatrics and Adults.State details.	#			
N.2	MRI compatible Incubator for Neonatal Imaging.	#			
N.3	MRI compatible contrast injector interphased with MR unit.	#			
N.4	200 syringes and 200 matching syringe lines.	#			
N.5	Drug Infusion pump compatible to MRI.	#			
N.6	List of standard accessories included but not specified in the tender.	1			
N.7	Patient Monitoring Video Camera System. Direct connection to monitor. All parts in the MRI room must be MRI compatible	#			
O	TECHNOLOGY				
O.1	Innovation. Give explanation	1			
O.2	Unique. Give explanation	1			
O.3	Clinical benefits. Give detailed explanation	1			


TENDER SPECIFICATION FOR THE SUPPLY, DELIVERY, INSTALLATION, COMMISSIONING AND MAINTENANCE OF MRI MACHINES (1.5 T) AT VARIOUS HOSPITALS IN GAUTENG

P	OPTIONAL ACCESSORIES				
P.1	The tenderer to give full description and pricing of optional accessories available for the system. Each item must be priced separately. Quote prices	1			
P.2	MRI compatible laryngoscopes, a fully equipped MRI compatible emergency trolley and MRI compatible Stethoscope.	1			
P.3	Scale weighing w/large dial type 2 150kg	1			
P.4	Patient slide for patient transfer 200g. (PVC Body Transfer Board and PVC Body Transfer Board Storage Brackets)	1			
P.5	MRI compatible wheelchair with footrests	1			
P.6	MRI Heavy Duty / Extra Wide Folding Walker.	1			
P.7	MRI Transfer / Gait Belt.	1			
P.8	MRI compatible steps / MRI Double Step Stool with Handrail	1			
P.9	I.V. Pole, MRI compatible	1			
P.10	Plastic chairs in magnet room	1			
P.11	Chair operator high back swivel with arm rests.	1			
P.12	Chair draughtsman mid armrests	1			
P.13	Shelve loose storage and display, wood.	1			
P.14	MRI Compatible Fire Extinguisher	1			
P.15	Hoist mobile electric hydraulic.	1			
P.16	One scanner cum copier and color printer. (Fax/printer/copier/scanner)	1			
P.17	Dicom color printer / camera for 8"x10" paper/transparency color printing, 500 transparencies to be included.	1			
P.18	Modem for remote service support.	1			

 GAUTENG PROVINCE <small>HEALTH REPUBLIC OF SOUTH AFRICA</small>					
TENDER SPECIFICATION FOR THE SUPPLY, DELIVERY, INSTALLATION, COMMISSIONING AND MAINTENANCE OF MRI MACHINES (1.5 T) AT VARIOUS HOSPITALS IN GAUTENG					
Q	MRI COMPATIBLE ANAESTHETIC UNIT				
Q.1	MRI Compatible Anaesthetic unit for General Anaesthetics of the paediatric patients is required. Please state the specifications of the unit on offer. Please also note the warranty and customer support agreement terms.	#			
Q.2	MRI compatible anaesthetic machine with ECG, NIBP, SPO2 & Gas capnography to suit adults, pediatric and neonates. ECG: Echo Cardiogram / NIBP: Non invasive BP / SPO2: Oxygen saturation / Gas Capnography: CO2 measurement	#			
Q.3	Anaesthesia machine must have a scavenger and the room must have an outlet for the scavenger for anaesthetic gases to be removed from the magnet room	#			
R	LASER CAMERA				
R.1	A state of the art dry laser camera to be included. Specify: Make Model	1			
R.2	UPS for the dry laser camera to be included	#			
S	GENERAL ISSUES				
S.1	IT INTEGRATION				
S.1.1	The system must have the following DICOM 3 compatibility:				
S.1.1.1	DICOM Send / Receive	#			
S.1.1.2	DICOM Storage Commitment (SC):	#			
S.1.1.3	DICOM query/retrieve	#			
S.1.1.4	DICOM worklist	#			
S.1.1.5	DICOM Basic Print	#			
S.1.1.6	DICOM MPPS (Modality Performed Procedure Steps)	#			
S.2	The system must comply to the following IHE profiles (Integrating the Healthcare Enterprise)	#			
S.2.1	Scheduled Workflow (SWF)	1			
S.2.2	Patient Information Reconciliation (PIR)	1			
S.2.3	Simple Image and Numeric Reports (SINR)	1			
S.2.4	System to be integrated to the existing PACS	#			
	TOTAL:	102			



GAUTENG PROVINCE
HEALTH
REPUBLIC OF SOUTH AFRICA

SPECIAL CONDITIONS OF CONTRACT GT/GDH/063/2023 THE SUPPLY, DELIVERY, INSTALLATION, COMMISSIONING AND MAINTENANCE OF MAGNETIC RESONANCE IMAGING (MRI) TO GAUTENG HEALTH INSTITUTIONS FOR A PERIOD OF THREE YEARS

ABBREVIATIONS

B-BBEE:	Broad Based Black Economic Empowerment
B-BBEE Controlled:	A juristic person, having shareholding or similar members interest, in which black Company participants, enjoy a right to Exercisable Voting Rights that is at least 51% of the total such rights, measured using the Flow-Through Principle.
B-BBEE Owned:	A juristic person having shareholding or similar member's interest that is BEE original company controlled, in which black participants enjoy a right to Economic interest that is at least 51% of the total of such rights, measured using the Flow-Through Principle.
BEC:	Bid Evaluation Committee
BSC:	Bid Specification Committee
DICOM:	Digital Imaging and Communication in Medicine
GCC:	General Conditions of Contract
GPG:	Gauteng Provincial Government
GPT:	Gauteng Provincial Treasury
GDoH:	Gauteng Department of Health
HIS:	Hospital Information System
PACS:	Picture Archiving and Communication System
POPI:	Protection of Personal Information Act 4 of 2013
PPPFA:	Preferential Procurement Policy Framework Act 5 of 2000
QC:	Quality Control
RIS:	Radiology Information System
RFP:	Request for Proposal
RDP:	Reconstruction Development Programme
SABS:	South African Bureau of Standards
SANAS:	South African National Accreditation System
SANS:	South African National Standard
SAHPRA:	South African Health Product Regulatory Authority
SCC:	Special Conditions of Contract
VAT:	Value- Added Tax



SPECIAL CONDITIONS OF CONTRACT GT/GDH/063/2023 THE SUPPLY, DELIVERY, INSTALLATION, COMMISSIONING AND MAINTENANCE OF MAGNETIC RESONANCE IMAGING (MRI) TO GAUTENG HEALTH INSTITUTIONS FOR A PERIOD OF THREE YEARS

1. COPYRIGHT

This document may be reproduced and distributed under the strict condition that the content hereof is not altered, unless the alteration has been done by authorized personnel stipulated the GDOH document control procedures are followed.

2. THE PURPOSE

The primary objective of the Gauteng Department of Health is the provision of Health services at the various Health institutions. An integral part thereof is to invite an open tender for the supply, delivery, installation, commissioning and maintenance of Magnetic Resonance Imaging (MRI) Machines to various Gauteng Department of Health institutions and to conclude a contract for the period of three years

3. THE BACKGROUND

The various Gauteng Department of Health institutions need to provide diagnostic X-ray services. The Gauteng Department of Health requires Magnetic Resonance Imaging (MRI) Machines based on an end-user need analysis identifying Magnetic Resonance Imaging (MRI) Machines. The items have been in demand but have not yet been procured.

The range of the Magnetic Resonance Imaging (MRI) Machines included in this tender, accommodates the different levels of care being rendered at Health facilities.

The procurement Plan of 2023/2024 at all institutions also has the MRI as a planned item.

Department of health is looking to acquire the MRI machines to comply with its mandate of delivering health care services to the communities. The acquiring of these equipment will include full service, repair and maintenance plan contract throughout the lifespan of the equipment.

4. LEGISLATIVE AND REGULATORY FRAMEWORK

4.1. The General Conditions of Contract (GCC):

This bid and all contracts emanating from this tender will be subject to the General Conditions of Contract (GCC), as issued by National Treasury in accordance with Treasury Regulation 16A published in terms of the Public Finance Management Act, 1999 (Act 1 of 1999). The general conditions are available on the National Treasury website (www.treasury.gov.za).



SPECIAL CONDITIONS OF CONTRACT GT/GDH/063/2023 THE SUPPLY, DELIVERY, INSTALLATION, COMMISSIONING AND MAINTENANCE OF MAGNETIC RESONANCE IMAGING (MRI) TO GAUTENG HEALTH INSTITUTIONS FOR A PERIOD OF THREE YEARS

4.2. The Special Conditions of Contract:

The Special Conditions of Contract are supplementary to that of the General Conditions of Contract. Where the Special Conditions of Contract conflict with the General Conditions of Contract, the Special Conditions of Contract shall prevail.

4.3. Other legal prescripts:

- a. The Constitution of the Republic of South Africa, 1996, Section 217
- b. Broad-Based Black Economic Empowerment Act 53 of 2003
- c. Public Finance Management Act 1 of 1999
- d. Preferential Procurement Policy Framework Act 5 of 2000
- e. Preferential Procurement Regulations, 2017
- f. Open Tender Framework, 2019
- g. Gauteng Finance Management Supplementary Amendment Act 6 of 2019
- h. Protection of Information Act 84 of 1982
- i. Promotion of Access to Information Act 2 of 2000
- j. Promotion of Administrative Justice Act 3 of 2000
- k. Occupational Health and Safety Act 85 of 1993
- l. Hazardous Substance Act 15 of 1973
- m. National Health Act 61 of 2003
- n. Health Professions Act 56 of 1974
- o. Protection of Personal Information Act 4 of 2013

4.4. Applicable National Standards

- a. IEC 60601 Compliance for electrical safety and essential performance.
- b. ISO 13485 Compliance for Quality Management Systems.
- c. Compliance with South African Health Products Regulatory Authority

5. THE FORMAT OF THE BID DOCUMENT

The bidders must submit the bid in a lever arch file in the format, as per Table 1 below.

Table 1: The Bid Format

Part of Bid Submission	Required documents
Part 1	<p>Section 1: Technical Proposal of the tender</p> <p>All the documents included in Section 1 must be read, completed, signed where applicable and submitted. Product information documents (e.g., catalogues, operating manuals, instruction leaflets, etc.), in English language.</p> <ol style="list-style-type: none"> 1. SBD 01: Invitation to Bid 2. SBD 4: Bidder's Disclosure



GAUTENG PROVINCE
HEALTH
REPUBLIC OF SOUTH AFRICA

SPECIAL CONDITIONS OF CONTRACT GT/GDH/063/2023 THE SUPPLY, DELIVERY, INSTALLATION, COMMISSIONING AND MAINTENANCE OF MAGNETIC RESONANCE IMAGING (MRI) TO GAUTENG HEALTH INSTITUTIONS FOR A PERIOD OF THREE YEARS

	<p>3. Quality Standards Certifications: Product/s supplied to the Gauteng Department of Health must conform to the quality standards and international device regulations. The bidders must submit the following certifications together with the bid documents before the closing date and time of the bid:</p> <p>a. A valid SAHPRA license or Letter of Compliance for the utilization of imported medical devices or locally manufactured medical devices.</p> <p>4. Product Brochure and Technical Data Sheets: Bidders must submit the fully comprehensive product brochures in and a copy of the technical data sheets that includes the technical specifications of the items tendered for. ALL BROCHURES MUST BE CLEARLY MARKED:</p> <ul style="list-style-type: none"> ▪ Brochures for (Item description) ▪ Item number: to be indicated on brochure ▪ Name of the company <p>5. Manufacturing Certificate: A valid copy of the Product Manufacturer Certificate, if the bidder is the original product manufacturer; or If the bidder is not the original product manufacturer, copy of the valid letter from the original product manufacturer, reseller or wholesale supplier that authorises the bidder to resell the product must be submitted together with the bid documents</p> <p>6. Tax Compliance Requirements: A printout via SARS e-Filing of the valid Tax Compliance Status (TCS) PIN, must be submitted with the bid documents at the closing date and time of the bid. In bids where consortia, joint ventures and sub-contractors are involved, each party must submit a separate PIN. The PIN, which is issued by the South African Revenue Services, can be used by third parties to verify the compliance status of the bidder online via SARS e-Filing.</p> <p>7. Copy of Central Supplier Database (CSD) Registration Summary Report Bidder must be registered with CSD and provide the Supplier Master Registration Number (MAAA number).</p>
Part 2	<p>All the supporting documents required for the Technical Evaluation:</p> <p>a. Organizational structure showing the technicians and the CVs of the technicians.</p> <p>b. Company profile.</p>



GAUTENG PROVINCE
HEALTH
REPUBLIC OF SOUTH AFRICA

SPECIAL CONDITIONS OF CONTRACT GT/GDH/063/2023 THE SUPPLY, DELIVERY, INSTALLATION, COMMISSIONING AND MAINTENANCE OF MAGNETIC RESONANCE IMAGING (MRI) TO GAUTENG HEALTH INSTITUTIONS FOR A PERIOD OF THREE YEARS

	<p>c. An original certified copy of the accreditation issued by the South African National Accreditation System (SANAS) to the Inspection Body affiliated with the bidders for the Acceptance Quality Assurance Test and commissioning of the equipment offered.</p> <p>d. The bidders are required to submit a certified copy of the Registration as an importer with the South African Revenue Service (SARS) together with the Import Permit obtained from to the International Trade Administration Commission (ITAC).</p>
Part 3	<p>Section 2: Financial Proposal of the tender. Completed Price Schedule document, referred to as Annexure B of the tender pack as well as an electronic copy in Excel format (not PDF), captured and saved on a memory stick.</p> <ol style="list-style-type: none"> 1. SBD 3.2: Pricing Schedule – Non-Firm Prices (Purchases) 2. Annexure B: Price schedule (total cost of ownership for the MRI machines for all Gauteng Department of Health institutions for a period of three years (36 months). 3. SBD 6.1: For Preference point claim

6. THE PRODUCTS SPECIFICATIONS

The bidders must refer to the attached **Annexure A** -

The tender specification for the supply, delivery, installation, de-commissioning, commissioning and maintenance of Magnetic Resonance Imaging (MRI) for all Gauteng Department of Health institutions for a period of three years (36 months).

The bidders must complete the Tender Specification as follows:

1. The original Technical Specification must be submitted.
2. The Technical Specification in MS Excel format that is attached must be completed in order to submit it in original.
3. Each response to the line item will be verified in the brochures, technical data sheets, user and technical manuals submitted by the bidders. Should the information required not be stated in the brochures, technical data sheets, user and technical manuals the bidder should then supply a letter from the manufacturer to verify the response.

7. EVALUATION METHODOLOGY

The evaluation of the bids will be done in accordance with the requirements of the Preferential Procurement Policy Framework Act (Act 5 of 2000) and the Preferential Procurement Regulations, 2022 in two stages:

Stage 1A: Mandatory Administrative Responsiveness

Stage 1B: Technical Evaluation

Stage 2: Price and Preference Points (Specific goals) Evaluation



SPECIAL CONDITIONS OF CONTRACT GT/GDH/063/2023 THE SUPPLY, DELIVERY, INSTALLATION, COMMISSIONING AND MAINTENANCE OF MAGNETIC RESONANCE IMAGING (MRI) TO GAUTENG HEALTH INSTITUTIONS FOR A PERIOD OF THREE YEARS

The 90/10 preference point system, which is applicable to bids with a Rand value equal to or above R50 million (all applicable taxes included)., shall be applied, where a maximum of 90 points will be allocated for price and maximum of 10 for specific goals (RDP goals: e, i and j) in terms of the requirements of the Preferential Procurement Policy Framework Act (Act 5 of 2000), Preferential Procurement Regulations 4 of 2022 and the Gauteng Department of Health Preferential Procurement Policy of 2022.

STAGE 1A: MANDATORY ADMINISTRATIVE COMPLIANCE

Bidders will be evaluated for the Mandatory Administrative compliance stage 1A as follows:

Table.2 Mandatory requirements:

ITEM NO:	MANDATORY DOCUMENTS (REQUIREMENTS)
1.	SBD 1: Invitation to Bid
2.	SBD 3.2: Non - Firm Prices (Purchases)
3.	Annexure B: Price Schedule
4.	SBD 4: Bidder's Disclosure
5.	<p>Quality Standards Certifications: Products supplied to the Gauteng Department of Health must conform to the quality standards and international device regulations. The bidders must submit the following certificates together with the bid documents:</p> <p>5.1. South African Health Product Regulatory Authority (SAHPRA) Licence/Certificate. A valid copy of the Licence/certificate of compliance as proof of registration with South African Health Product Regulatory Authority.</p>
6.	<p>Product Brochure and Technical Data Sheets: Bidders must submit the fully comprehensive copy of product brochures and a copy of the technical data sheets that includes the technical specifications of the items tendered for together with the bid documents.</p>
7.	<p>Compulsory Briefing Session certificate The bidders are requested to attend a compulsory briefing session to address and clarify any misunderstanding or ambiguity prior to the proposal submission closing date. If a bidder does not attend the compulsory briefing session the bidder shall be regarded as non-responsive and will be disqualified. The bidders must sign the Briefing Session Attendance Register. Minutes of all proceedings during the compulsory briefing session shall be recorded and be binding. Bidders must submit the signed briefing session certificate issued at the briefing session with the bid documents before the closing date.</p>



GAUTENG PROVINCE
HEALTH
REPUBLIC OF SOUTH AFRICA

SPECIAL CONDITIONS OF CONTRACT GT/GDH/063/2023 THE SUPPLY, DELIVERY, INSTALLATION, COMMISSIONING AND MAINTENANCE OF MAGNETIC RESONANCE IMAGING (MRI) TO GAUTENG HEALTH INSTITUTIONS FOR A PERIOD OF THREE YEARS

8.	<p>Manufacturing Certificate: A valid copy of the Product Manufacturer Certificate, if the bidder is the original product manufacturer; or If the bidder is not the original product manufacturer, copy of the valid letter from the original product manufacturer, reseller or wholesale supplier that authorises the bidder to resell the product must be submitted together with the bid documents.</p>
----	---

If a bidder does not meet all the requirements stated above the bid will be disqualified.

STAGE 1B: TECHNICAL EVALUATION

Only bidders who have complied with the Mandatory Administrative Compliance will be evaluated for the Technical Evaluation.

Bidders must observe and refer to the technical specifications of supply, delivery, installation, de-commissioning, commissioning and maintenance of Magnetic Resonance Imaging (MRI) for all Gauteng Department of Health institutions for a period of three years (36 months). See attached **Annexure A**.

- The Product Specification in MS Excel format that is attached below must be completed in order to submit it in original.
- Each response to the line item will be verified in the brochures, technical data sheets, user and technical manuals submitted by the prospective service providers. *25 August 2023*
- Bidders must indicate each line item on the brochure by stating the page number and paragraph on column F (Reference to Brochure / Technical Data Sheet (verification), of the technical specification annexure A.
- Should the information required not be stated in the brochures, technical data sheets, user and technical manuals the bidder should then supply a letter from the manufacturer to verify the response.
- The bidders must refer to in the Microsoft Excel to indicate the compliance with the product specification.

If a bidder does not meet the technical specifications requirements, the bid will be disqualified.



SPECIAL CONDITIONS OF CONTRACT GT/GDH/063/2023 THE SUPPLY, DELIVERY, INSTALLATION, COMMISSIONING AND MAINTENANCE OF MAGNETIC RESONANCE IMAGING (MRI) TO GAUTENG HEALTH INSTITUTIONS FOR A PERIOD OF THREE YEARS

STAGE 2: PRICE AND SPECIFIC GOALS

The 90/10 preference point system, which is applicable to bids with a Rand value equal to or above R50 million (all applicable taxes included)., shall be applied, where a maximum of 90 points will be allocated for price and maximum of 10 for specific goals (RDP goals: e, i and j) in terms of the requirements of the Preferential Procurement Policy Framework Act (Act 5 of 2000), Preferential Procurement Regulations 4 of 2022 and the Gauteng Department of Health Preferential Procurement Policy of 2022.

Bidders are referred to:

- The SBD 3.2 Non-firm prices (Purchases)
- Annexure-B for pricing schedule and
- The SBD 6.1 for price and specific goals preference point claim.

For this tender the Gauteng department of health will promote the enterprises/companies based in Gauteng in support of the RDP goals (3), the specific goal (i) The empowerment of the work force by standardising the level of skill and knowledge of workers(3) and (j) the development of human of human resources, including by assisting in tertiary and other advanced training programmes, in line with key indicators such as percentage of wage bill spent on education and training and improvement of management skills (4). These services will be rendered in Gauteng province for the creation of jobs or the intensification of labour absorption of the GDoH Preferential Procurement Policy of 2022.

Table 3. The maximum points for this tender are allocated as follows:

PRICE AND SPECIFIC GOAL REQUIREMENTS	POINTS	DOCUMENTARY PROOF
POINTS FOR PRICE	90	SBD 3.2 and Annexure B (Pricing Schedule)
POINTS FOR SPECIFIC GOAL (1): RDP goal (2)(e) Promotion of enterprise/s/companies located in Gauteng Province	3	Bidder must submit certified copy of the lease agreement or Municipal bill or a valid copy of title deed in their name
POINTS FOR SPECIFIC GOAL (1): RDP goal (2) (i) The empowerment of the work force by standardising the level of skill and knowledge of workers	3	Evidence of staff employment or absorption contract funded by the bidder. (Bidder to submit copy of contracts indicating commencement date, signed by the employer and employee)



[illegible][illegible]

— — — — —

© 2006 The Authors
Journal compilation © 2006 Blackwell Publishing Ltd

1. *Journal of Management Studies*, 1990, 27, 1, 1-14.



SPECIAL CONDITIONS OF CONTRACT GT/GDH/063/2023 THE SUPPLY, DELIVERY, INSTALLATION, COMMISSIONING AND MAINTENANCE OF MAGNETIC RESONANCE IMAGING (MRI) TO GAUTENG HEALTH INSTITUTIONS FOR A PERIOD OF THREE YEARS

8.3. Electrical power supply:

The successful bidder must -

- a. Verify that adequate electrical power supply is available for the optimal functionality of the equipment.
- b. If electrical power supply is found to not be adequate, then an upgrade that meets our Certificate of Compliance must be provided by the supplier at their costs.
- c. The power input must be 380V, 50Hz AC offered, where applicable. Bidders must ensure that the product quoted for is fitted appropriately with approved mains cable.
- d. A 3 phase in / out power supply must be offered, where applicable also stating the operating Amps and Watts of the equipment.

8.4. Building alterations and installation:

The successful bidder must -

- a. Provide a separate quotation for building alterations, where applicable.
- b. Be responsible for all the building, air conditioning, electrical, mechanical and plumbing alterations, which can only be executed through the approval of the Department of Infrastructure Development (DID) and the Facility Management Unit (FMU) at the Institution, where applicable.
- c. Consult the Facility Managing Unit (FMU) at the health institution and the regional representative of the Gauteng Department of Infrastructure Development (DID) in respect of the building alterations in order to establish minimum standards.
- d. Inspect the site in order to quote for any building alterations that need to be made to accommodate the equipment tendered for. A building plan should be presented to the Facility Management Unit at the Health institution who will in turn obtain the necessary approvals prior to commencement of actual work.

8.4.1. Specify the following on the quotation for building alterations:

- a. Construction alterations:
- b. Air conditioning, where applicable (specify), as follows:
 1. Include any air conditioning that is considered necessary for optimal functioning of the unit.
 2. Any such air conditioning must be covered by the three-year guarantee period.
 3. The air conditioner must be 18 000 BTU or above.
- c. Electrical alterations, where applicable (specify).
- d. Mechanical and plumbing alterations, where applicable (must be specified). Including other related work

8.4.2. Complete the following within 6-12 weeks after receiving purchase order:

- a. The delivery of the equipment to the institution.
- b. The finalization of the building alterations.



SPECIAL CONDITIONS OF CONTRACT GT/GDH/063/2023 THE SUPPLY, DELIVERY, INSTALLATION, COMMISSIONING AND MAINTENANCE OF MAGNETIC RESONANCE IMAGING (MRI) TO GAUTENG HEALTH INSTITUTIONS FOR A PERIOD OF THREE YEARS

- c. Time required for installation and commissioning from the date of delivery of the equipment to the institution.
- d. Total time required from the placement of the order to the commissioning of the equipment.
- e. In the instance of any delay in the installation, a communication must be done with the relevant institution in writing.
- f. Should there be existing equipment in the room, the successful bidder must in consultation and co-operation with the Asset Management Unit at the Health institution, remove it for official disposal. Please note that the existing equipment remains the property of the health institution, where applicable.
- g. Provide equipment layout and electrical drawings in brochure form and on a memory stick.
- h. The staff of the successful bidder shall be available for consultation and to attend installation planning meetings throughout the entire installation process of the system.
- i. Do a pre-installation survey and notify the health institution with recommendation of any gaps affecting the installation of the equipment.

8.5. Radiation Protection

The successful bidder must -

Provide all Quality Control Test Tools for basic testing of the applicable Systems.

8.6. Technology

- a. No product or part thereof shall be second hand or refurbished.
- b. The bidder must be able to upgrade the software at no cost during the contract period. The system must be upgradeable to enhance functionality.

8.7. Manuals and documentation

The bidders must submit a copy of all the brochures that includes the brand name offered and complete technical specification as follows:

- a. Full printout of the brochure file.
- b. The marketing brochure as well as the technical product data sheets.

8.7.1. The successful bidders must submit the complete service / repair and user manual book as follows:

- a. original manuals must be supplied.
- b. The service / repair and user manual book must include the following information:
 - 1. Fault finding guide
 - 2. Circuit diagrams / schematics
 - 3. Circuit descriptions and PCB layouts
 - 4. Calibration guide
 - 5. Part numbers and exploded diagram of mechanical parts / panels
- c. An electronic copy of the manuals and documentation in English must be supplied on a memory stick.



SPECIAL CONDITIONS OF CONTRACT GT/GDH/063/2023 THE SUPPLY, DELIVERY, INSTALLATION, COMMISSIONING AND MAINTENANCE OF MAGNETIC RESONANCE IMAGING (MRI) TO GAUTENG HEALTH INSTITUTIONS FOR A PERIOD OF THREE YEARS

- d. The following must be submitted in English:
 1. Complete operator /user manuals
 2. DICOM conformance statements
 3. HIS and RIS/PACS conformance statements
 4. Quality assurance manuals
 5. Service manuals with full maintenance procedures, parts lists, system diagrams and electrical, mechanical and pneumatic schematics.
 6. Log book with instructions for daily, weekly, monthly and quarterly maintenance checklists.
 7. Certificate of calibration and inspection from factory.
 8. The vendor shall provide updates and revisions of the manuals at no extra charge for the lifetime of the equipment

8.8. Warranty, Maintenance and Quality Control Testing

- a. The three-year warranty will commence after formal acceptance and handover of the equipment.
- b. The three-year warranty must be provided for all the equipment in the tender document.
- c. The annual QC tests are mandatory on X-Ray units.
- d. Scheduled Services by Technical Staff and Annual QC tests by Inspection Bodies, must be performed during normal working hours, from 08h00 to 17h00 during week days or as per arrangement.
- e. Bidders must supply a three-year warranty against poor workmanship and latent defects and parts, as follows:
 1. The warranty must include all materials used and all workmanship.
 2. All software updates must be included in the warranty at no extra costs.
 3. Spares and traveling time cost to be included in the warranty. Spares should be available within three to five working days.
 4. A callout and backup service for urgent service requests must be available daily for 24 hours and be included in the warranty. The response time must be within two hours.
- f. The successful bidder must arrange with both the respective Hospital / Institution and the Health Technology Services - HOD of Nuclear Medicine /Assistant Director Radiography / Medical Physicist before commissioning the equipment at the respective hospital / institution.
- g. The equipment will only be accepted after the commissioning and approval for use by SAHPRA of the equipment to be ready for the first patient.
- h. Software changes to the equipment that are corrective in nature and initiated due to software errors, regulatory requirements or safety reasons, shall be delivered and installed at no charge for the life span of the equipment.
- i. If movable equipment is taken away to the service and repair facility for repair and maintenance during the warranty and maintenance periods, a loan set must be supplied for use by the institution.
- j. If the equipment bidden for cannot be repaired during the warranty period, it must be replaced and the institution be furnished with new equipment
- k. The up-time of the unit must comply as follows:



SPECIAL CONDITIONS OF CONTRACT GT/GDH/063/2023 THE SUPPLY, DELIVERY, INSTALLATION, COMMISSIONING AND MAINTENANCE OF MAGNETIC RESONANCE IMAGING (MRI) TO GAUTENG HEALTH INSTITUTIONS FOR A PERIOD OF THREE YEARS

1. Up-Time is defined as follows: 24/7; i.e., 365 days times 24 hours = 8760 Hours. A down time of 2% relates to 175 hours per annum.
2. The up-time of the unit must be better than 98%, excluding scheduled preventative maintenance and software upgrades, measured on a quarterly basis. The percentage lower than 98% will be added to the warranty period.
3. A sliding scale penalty clause will form part of the service contract. This will result in the maintenance payment being reduced by a pro rata amount that the up-time is less than 98%.
- l. The seven year all inclusive, fully comprehensive preventative maintenance, service and repair contract will commence after the three-year warranty period has expired, as follows.
 1. Specify the total cost per year and the cumulative cost for the seven-year all-inclusive service contract.
 2. The contract must cover, but not be limited to the following: All parts (including, where appropriate, X-Ray tubes and other glassware), labour, travelling, accommodation, service and maintenance.
 3. The seven-year maintenance plan must also include all quality checks and quality assurance requirements (including Annual QA-tests by SANAS accredited inspection bodies) and also all required calibrations.
 4. All software updates must be included in the seven-year maintenance plan at no extra costs.
 5. A Technical Report outlining the functional status of each unit must be submitted to the Head of the Department on an annual basis. (not to be confused with Annual QC Test Results issued by the Inspection Bodies)
 6. For each and every repair, service and maintenance and all other services (upgrades, updates, etcetera) a job card must be furnished.
- m. The lifespan and end of support date of the equipment offered must be indicated.
 1. Spare parts must be guaranteed available for the specified life of the equipment, with a minimum of ten years.
 2. Spare part kits should specify all spare parts that will be included and stored on-site.
- n. Bidders must note that dedicated test equipment, spare parts and any special tooling required for the maintenance of equipment quoted on must be made available and a dedicated workshop must be arranged. The Department has the right to visit and validate the workshop.

8.9. Commissioning of the equipment

The equipment and accessories ordered shall be delivered, installed, tested and commissioned at the expense of the bidders prior to acceptance.

8.10. User and Technical Training

- a. Upon commissioning the system, the application specialist must be immediately available to provide the training of staff.
- b. On site user training sessions on the operation of the equipment must be provided at no additional cost to the final bid price to ensure the correct



SPECIAL CONDITIONS OF CONTRACT GT/GDH/063/2023 THE SUPPLY, DELIVERY, INSTALLATION, COMMISSIONING AND MAINTENANCE OF MAGNETIC RESONANCE IMAGING (MRI) TO GAUTENG HEALTH INSTITUTIONS FOR A PERIOD OF THREE YEARS

application of the unit. Sessions to be conducted five days per week for the first 2 weeks.

- c. Optional training for two days for clinical engineers, medical physicist and IT specialists must be provided.
- d. The application specialist must provide repetitive and refresher training for one day every six months for the duration of the warranty and maintenance, as requested by the users in order to demonstrate and train all staff on all aspects of the equipment.
- e. Quality Control training for two days must be provided by the supplier's technician / Inspection Body is mandatory within the first year and must be included in the tender price (not included in other training).
- f. All other further training must be available on request.

9. PAYMENT TERMS

Section 38(1) (f) of the PFMA and Treasury Regulation 8.2.3 regulates the payment to suppliers within 30 days of invoice receipt. In support of this it is compulsory for the successful bidder/s, on award, to register for GPT Electronic Invoice Submission and Tracking.

10. THE CONDITIONS OF THE BID AWARD

- a. The Gauteng Department of Health reserves the right to award or not to award the tender.
- b. The bid will be awarded to the bidder/s whose bid scored the highest combined price and preference points out of 100.
- c. The Gauteng Department of Health reserves the right to award the same item to more than one bidder to address product availability and compatibility and this will be at the discretion of the Bid Evaluation Committee".
- d. The successful bidder must be tax compliant at the awarding of the tender.
- e. Bidder must be registered with CSD and provide the Supplier Master Registration Number (MAAA number)
- f. The successful bidder will be required to submit valid copy/s of competency certificates of each technician/s employed by the bidder relevant to the equipment offered on the installation, commissioning, calibration and maintenance of the item/s tendered for.

11. TRAVEL

The Gauteng Department of Health will not be liable for any other travel costs incurred by the bidder. Prices quoted must be furnished based on "rendered on site" at Gauteng Health Institutions.

12. AUTHORISATION DECLARATION

- a. Any bidder who is sourcing goods or services from a third party must compile an "Authorization Declaration" in full for all relevant goods or services, sign it and submit it together with the bid documents at the closing date and time of the bid.



SPECIAL CONDITIONS OF CONTRACT GT/GDH/063/2023 THE SUPPLY, DELIVERY, INSTALLATION, COMMISSIONING AND MAINTENANCE OF MAGNETIC RESONANCE IMAGING (MRI) TO GAUTENG HEALTH INSTITUTIONS FOR A PERIOD OF THREE YEARS

- b. The Gauteng Department of Health reserves the right to verify any information supplied by the bidder in the Authorization Declaration and should the information be found to be false or incorrect, the Gauteng Department of Health will exercise any of the remedies available to it in the bid documents.
- c. The bidder must ensure that all financial and supply arrangements for goods or services have been mutually agreed upon between the bidder and the third party. No agreement between the bidder and the third party will be binding on the Gauteng Department of Health.
- d. The bidders must submit a duly completed and signed Authorization Declaration, with the required annexure(s), in accordance with the above provisions in order not to delay the evaluation of the bid.

13. COUNTER CONDITIONS

Bidders' attention is drawn to the fact that amendments to any of the Bid Conditions or setting of counter conditions by bidders may result in the invalidation of such bids.

14. FRONTING

- a. The Gauteng Department of Health supports the spirit of broad based black economic empowerment and recognizes that real empowerment can only be achieved through individuals and businesses conducting themselves in accordance with the Constitution and in an honest, fair, equitable, transparent and legally compliant manner. Against this background the National Treasury condemns any form of fronting.
- b. The Gauteng Department of Health, in ensuring that bidders conduct themselves in an honest manner will, as part of the bid evaluation processes, conduct or initiate the necessary enquiries/investigations to determine the accuracy of the representation made in bid documents.
- c. Should any of the fronting indicators as contained in the Guidelines on Complex Structures and Transactions and Fronting, issued by the Department of Trade and Industry, be established during such enquiry/investigation, the onus will be on the bidder / contractor to prove that fronting does not exist.
- d. Failure to do so within a period of 14 days from date of notification may invalidate the bid / contract and may also result in the restriction of the bidder/contractor to conduct business with the public sector for a period not exceeding ten years, in addition to any other remedies the National Treasury may have against the bidder/contractor concerned.

15. CONTRACT PERIOD

The contract shall be for a period of three years.



SPECIAL CONDITIONS OF CONTRACT GT/GDH/063/2023 THE SUPPLY, DELIVERY, INSTALLATION, COMMISSIONING AND MAINTENANCE OF MAGNETIC RESONANCE IMAGING (MRI) TO GAUTENG HEALTH INSTITUTIONS FOR A PERIOD OF THREE YEARS

16. VALIDITY PERIOD

Bids are held to be valid for a period of ninety (90) days after the closing date. Should a bidder retract his/her offer without good reason, in the opinion of the Department, they may be held responsible for the cost of a possible re-tender.

17. POST AWARD PARTICIPATION

In terms of Treasury Regulation 16A6.6, an Accounting Officer/Accounting Authority may, on behalf of the department, constitutional institution, public entity, request to participate in transversal term contracts arranged by means of a competitive bidding process by Gauteng Department of Health, subject to written approval by the BAC for Transversal Contracts and relevant contractors.

18. MERGERS, TAKE OVERS AND CHANGES IN SUPPLIER DETAIL

- a. Where a contracted supplier merges with or is taken over by another, the contracted supplier must inform the Department of Health in writing immediately (within seven days) of relevant details.
- b. The Department of Health reserves the right to agree to the transfer of contractual obligations to the new supplier under the prevailing conditions of contract or to cancel the contract.
- c. A contracted supplier must inform the Department of Health within seven days of any changes of address, name or banking details.

19. THIRD PARTIES

- a. Participating authorities will not make a payment to or consult regarding orders with a third party.
- b. No third party is entitled to put an account on hold.

20. USE OF FLUID CORRECTING SUBSTANCES

The use of any corrective fluid is strictly prohibited and will result into disqualification of bidder from the process.

21. TENDER PRICE

The tender price must be clearly broken down into all the items that are included and the prices per item. All bidders must indicate whether an optional item in a brochure is required to meet the specification and if it is included in the tender price. The breakdown of the prices also assists if part payments have to be made. The following prices must be submitted and indicated separately:

- a. A detailed quotation that indicates what items are included and showing the breakdown of prices.
- b. The three-year warranty must be included in the unit price of the equipment. The purchase pricing schedule must be completed in full.



SPECIAL CONDITIONS OF CONTRACT GT/GDH/063/2023 THE SUPPLY, DELIVERY, INSTALLATION, COMMISSIONING AND MAINTENANCE OF MAGNETIC RESONANCE IMAGING (MRI) TO GAUTENG HEALTH INSTITUTIONS FOR A PERIOD OF THREE YEARS

- c. The tender price must specify the standard items included in the equipment offered.
- d. The bidder must clearly distinguish the cost of the standard and the optional items (e.g. consumables) in the pricing schedule, as follows.
 - 1. The price of the unit.
 - 2. Warranty of three years.
 - 3. Maintenance contract.
 - 4. Optional items separately.

22. THE COMPLETION OF THE PRICING SCHEDULE (ANNEXURE B)

Bidders are required to complete all the mandatory response fields in the tender pricing schedule for the equipment. The column for the bid price and the other columns in the Price Schedule as well as item questionnaires.

- a. Annexure B must be submitted as well as an electronic copy be in Excel format (not PDF), captured and saved on a memory stick.
- b. Hard Copy Format:

The original (hard copy) must be written clearly and legibly.
- c. Soft Copy Format:
- d. The electronic (soft copy) must be submitted on memory stick to the Gauteng Provincial Treasury Tender Office. The memory stick must be clearly marked with the company name and tender number. The electronic copy in Excel format will form part of the evaluation worksheets of the bids.
- e. The price schedule in MS Excel format that is attached must be completed in order to submit it in original. The bidders must also complete the Price Schedule in MS Excel format and complete it electronically in order to submit it electronically.
- f. The bidders must ensure that there are no discrepancies between the electronic (soft-copy) saved on a memory stick and the original hard copy submissions of the Price Schedule. If any discrepancies are detected, the hard copy document will take precedence over the electronic copy. The Gauteng Department of Health may contact the bidder, but shall not be obliged to do so, for clarification regarding any discrepancies found.
- g. Each original bid with the memory stick must be submitted in a separate, sealed envelope to Gauteng Provincial Treasury, Tender Box before the closing date



GAUTENG PROVINCE
HEALTH
REPUBLIC OF SOUTH AFRICA

SPECIAL CONDITIONS OF CONTRACT GT/GDH/063/2023 THE SUPPLY, DELIVERY, INSTALLATION, COMMISSIONING AND MAINTENANCE OF MAGNETIC RESONANCE IMAGING (MRI) TO GAUTENG HEALTH INSTITUTIONS FOR A PERIOD OF THREE YEARS

and time. The name and address of the bidder, the bid number and the closing date must be clearly endorsed on the sealed envelope

23. ENQUIRIES

Technical Enquiries:

Prof Nasreen Mahomed

E-mail: Nasreen.Mahomed@wits.ac.za

Admin Enquiries:

Acquisition Management:

E-mail: mzimkhulu.gunundu@gauteng.gov.za jerry.phukuje@gauteng.gov.za



GAUTENG PROVINCE
PROVINCIAL TREASURY
REPUBLIC OF SOUTH AFRICA

Provincial Supply Chain Management

Registered Supplier Confirmation

Page 1 of 1

THIS FORM IS TO BE COMPLETED BY REGISTERED SUPPLIERS ONLY

PLEASE NOTE:

SUPPLIERS ARE REQUIRED TO PROVIDE THEIR REGISTERED CENTRAL SUPPLIER DATABASE (CSD) NUMBER _____

For confirmation of your supplier number and/or any assistance please call the GPT Call Centre on **0860 011 000**.

Registered Suppliers to ensure that all details completed below are CURRENT.

MANDATORY SUPPLIER DETAILS			
GPT Supplier number			
Company name (Legal & Trade as)			
Company registration No.			
Tax Number			
VAT number (If applicable)			
COIDA certificate No.			
UIF reference No.			
Street Address		Postal Address	
CONTACT DETAILS			
Contact Person		Telephone Number	
Fax Number		Cell Number	
e-mail address		Principal's Id number	
BANKING DETAILS (in the name of the Company)			
Bank Name		Branch Code	
Account Number		Type of Account	

I HEREBY CERTIFY THAT THIS INFORMATION IS CORRECT.

Name(s) & Signature(s) of Bidder(s)

DATE:



Provincial Supply Chain Management

Financial Statements

Page 1 of 1

Submission of Financial Statements

The latest financial statements for the last two years are required (except if it is a new or a dormant entity)

- a) Financial statements must be signed by the auditor (in the case of companies) or the accounting officer (in the case of close corporations) the owner (in case of sole proprietors). Signatures must be on the accounting officer's / auditors report on the auditor's /accounting officer's letterhead.
- b) Financial statements must be signed by the member/s (in the case of close corporations) or by the director/s (in the case of companies.)
- c) In bids where consortia/joint ventures/sub-contractors and partnerships are involved, all bidders must submit their financial statements.
- d) If it is a new or dormant entity an opening set of financial statements must be submitted with the tender document. A letter from the auditor (in the case of companies) or the accounting officer (in the case of close corporations) stating that the entity has not yet traded must be attached.
- e) In cases where an entity has operated for a period less than a year the Management Accounts Report for the period in operation must be submitted signed accordingly as stated in paragraph (a) and (b) of this document.
- f) In cases where the entity has operated for a period more than a year but less than two years, then the financial statement for the first year of operation signed accordingly as per paragraph (a) and (b) of this document must be submitted.



SERVICE DESCRIPTION

AGREEMENT ENTERED INTO BY AND BETWEEN THE GAUTENG PROVINCIAL GOVERNMENT IN ITS _____

AND HEREIN REPRESENTED BY _____ IN HIS /
HER CAPACITY AS _____ AND AS SUCH DULY
AUTHORISED ("THE END USER")

AND

_____ A COMPANY WITH LIMITED
LIABILITY AND DULY INCORPORATED IN TERMS OF THE COMPANY
LAWS OF THE REPUBLIC SOUTH AFRICA, WITH COMPANY
REGISTRATION NO _____ AND PRINCIPAL PLACE
OF BUSINESS AT _____ AND HEREIN REPRESENTED
BY _____ IN HIS / HER CAPACITY AS
_____ AND AS SUCH DULY AUTHORISED
("THE SUPPLIER").

AND WHEREAS

The Supplier is the preferred supplier for the supply, delivery, installation, commissioning and maintenance of office equipment and labour saving devices ("equipment") in terms of the contract.

AND WHEREAS

The End user is, from time to time, desirous of hiring from the supplier one or more equipment, and the Supplier is in turn desirous of renting such equipment to the End user

Table of Contents

1.	Definition and rules of interpretation.....	3 – 4
2.	Duration and Termination.....	5
3.	Obligations of the Supplier.....	5
3.1	Delivery and Installation.....	5
3.2	Maintenance	5 - 6
3.3	Product Support.....	6
3.4	Training	7
3.5	Indemnity and Insurance.....	7
3.6	Subcontracting.....	8
3.7	Confidentiality.....	8
4.	Obligations of the End-User.....	8
5.	Breach.....	8 - 9
6.	Payment	9 - 12
7.	Notice and Domicilia.....	10
8.	General	10 - 11
8.1	Whole Agreement.....	11
8.2	No Variation.....	11
8.3	Waiver.....	11
8.4	Severability.....	11
8.5	Applicable Law.....	11
8.6	Jurisdiction.....	11
8.7	Survival.....	11

1 NOW THEREFORE THE PARTIES AGREE AS FOLLOWS:

1.1 Rules of interpretation. In this Agreement:-

1.1.1 clause headings are for convenience and are not to be used in its interpretation;

1.1.2 unless the context indicates a contrary intention an expression which denotes:-

1.1.2.1 any gender shall include the other genders;

1.1.2.2 a natural person shall include a juristic person and vice versa;

1.1.2.3 the singular shall include the plural and vice versa; and

1.1.2.4 references to clauses, schedules, parts and sections are, unless otherwise provided, references to clauses, schedules, parts and sections of this Agreement.

1.2 Meanings of expressions and words. In this Agreement the following expressions and words have the meanings assigned to them below and derivative expressions and words will have a corresponding meaning: -

1.2.1 **Agreement** means this agreement read together with the General and Special Conditions of Contract of Contract RT3-2009R which form an integral part of this Agreement

1.2.2 **Copy Charges** means the consideration, where applicable, (or, as the context may require, part thereof) payable by the End-user to the Supplier for the maintenance to be provided by the Supplier in terms hereof, which is the amount payable for each black and white or colour copy (as the case may be) that is produced by the equipment at the rate as set out in Addendum 1 of this Agreement, and which is calculated by multiplying the total black and white or colour copies (as the case may be) so produced during a copy period by the charge payable for each black and white or colour copy (as the case may be) as stipulated in Addendum 1 of this Agreement.

1.2.3 **Copy Period** means a period of one calendar month, each month commencing on the 1st day of each month (except the first period, which will be the period from Commencement Date until the last day of that calendar month). Copy Period means the period during which copies are made, calculated by means of an opening and closing meter reading, on a monthly basis.

1.2.4 **End-user** means the government institution described on

page one hereof.

- 1.2.5 **Equipment** means all or any, as the context may require, of the equipment which is/are or will be the subject matter of this Agreement and which are more fully described in Addendum 1 of this Agreement.
- 1.2.6 **Initial Period** means the period of 36 (thirty six) months from the Commencement Date.
- 1.2.7 **Maintenance** means the obligation assumed by the Supplier to maintain the relevant equipment in proper and efficient operating condition on the terms as set out herein and in accordance with the specifications applicable to the relevant equipment.
- 1.2.8 **Rental** means the consideration payable by the End-user to the Supplier for the use of the equipment in the amounts as stipulated in Addendum 1 of this Agreement.
- 1.2.9 **Working day** means days on which business is generally conducted, i.e. Saturdays, Sundays and official public holidays excluded.
- 1.2.10 **Commencement date** means the date on which the installation and commissioning of equipment is completed.
- 1.2.11 **Material breach** means an event that goes to the root of this agreement.
- 1.2.12 **Month** means calendar month.
- 1.2.14 **Termination date** means 36 months after the commencement date.

2. DURATION AND TERMINATION

- 2.1 This agreement shall commence upon the commencement and shall endure for a period of 36 months and automatically terminate on the termination date by effluxion of time, unless terminated earlier or extended in terms of the provisions of this contract;
- 2.2 This agreement, may at the sole discretion of the End user, be extended in writing for a maximum period of twenty four months on the same terms and conditions except for the rental which shall be reduced by 75% (seventy five percent) of the rental specified herein.

3. OBLIGATIONS OF THE SUPPLIER

3.1 DELIVERY AND INSTALLATION

3.1.1 The Supplier undertakes to:

- 3.1.1.1 deliver the equipment conforming exactly to the description of the equipment as specified in addendum 1 of this agreement;
- 3.1.1.2 deliver and install new and unused equipment at the location selected by the End-user;
- 3.1.1.3 ensure that the equipment is delivered and installed in good condition and working order;

3.2 MAINTENANCE

The Supplier undertakes to:

- 3.2.1 ensure that the equipment performs in accordance with the manufacturer's specifications;
- 3.2.2 keep and maintain the equipment rented by the End-user in good and proper condition and working order and in such manner that the End-user will have the use thereof in an efficient operating condition, and to take such reasonable preventative action as may be necessary or open to it in order to limit the incidence and frequency of breakdowns of equipment to a minimum.
- 3.2.3 for this purpose The Supplier shall ensure that a qualified technician responds promptly to any notification of the End user of a breakdown or malfunction of any equipment.
The response time on such notification shall be as follows:
- 3.2.4 forthwith provide temporary loan equipment to the End-user if the fault in the equipment cannot be repaired, or is not expected to be reasonably repaired, within the period as set out in of this Agreement.

- 3.2.5 the availability of an adequate number of qualified technicians and personnel on a full-time basis to perform the maintenance required under this Agreement;
- 3.2.6 to make available the services of a fully qualified technician from 08h30 to 16h30 each working day to carry out preventative maintenance on the equipment;
- 3.2.7 to supply the quantities of spare parts, toner, developer, fuser oil and other consumables necessary to keep the equipment in proper operating condition;
- 3.2.8 to make available of full coverage maintenance, including preventive maintenance, all service calls and replacement all defectives or worn parts including expandable parts, and all consumable supplies. Should the Supplier fail to provide any of the consumables, or repair or replace with an equivalent unit, any equipment as required, then the Rental Copy Charges for the relevant month in respect of such equipment shall be forfeited by the Supplier and accordingly the End-user shall not be required to pay such rental and copy charges. Should the Supplier not have remedied the failure within 10 (ten) working days of notice from the Enduser then the End-user shall be entitled on written notice to the Supplier to immediately terminate the Agreement in respect of the relevant equipment at no additional cost or penalty to the End-user and the Supplier shall be obliged to remove the relevant equipment listed in the Agreement at is sole cost and expense;
- 3.2.9 remove the equipment from location of the End-user on termination of this Agreement at no additional charge;
- 3.2.10 perform all the services in terms of this contract with due care skill, efficiency and diligence in accordance with the best professional practice;

3.3 **PRODUCT SUPPORT**

- 3.3.1 The Supplier will from time to time and to the extent that is reasonably necessary or required by the End-user for the proper utilisation of the equipment, provide advice and assistance to the End-user and to provide such reports and data relevant to the usage of the equipment as may reasonably be required by the End-user.
- 3.3.2 Without limiting the generality of its obligations under clause
- 3.3.1 The Supplier hereby authorises the End-user to install access key control devices on the relevant equipment and will provide all necessary assistance to ensure the proper integration of the access key control devices with the equipment. The Supplier shall also assist the End-user in the installation of any copy control devices and copy management devices on the equipment as may be reasonably required by the End-user.
- 3.3.3 Where The Supplier or any of its employees, agents or independent contractors ("Representatives") accesses the premises of the End-

user, under or pursuant to, the terms of this Agreement, The Supplier and its representatives shall abide by and comply with the safety, health and environmental policies and procedures and other lawful directions of the End-user.

3.4 TRAINING

- 3.4.1 On installation of the equipment, The Supplier shall provide adequate training to the personnel of the End-user at no additional charge.
- 3.4.2 Instruction manuals shall also be provided by the Supplier free of charge for all equipment rented in terms of this agreement. The instruction manuals shall contain, but not be limited to, the following information:
 - 3.4.2.1 Defining the capabilities of the equipment (specification).
 - 3.4.2.2 Describing the technical operations of the equipment.
 - 3.4.2.3 Describing the use criteria of the equipment.
- 3.4.3 The Supplier shall also provide such further training may be required by the End User from time to time.

3.5 INDEMNITY AND INSURANCE

The Supplier hereby:

- 3.5.1 Undertakes, at its own expense, to indemnify, protect and defend the End User from and against all actions, claims, losses or damages arising from any negligent act or omission by the Supplier including but not limited to all damages or loss which may be payable or arise as a result of any claim or proceedings in respect of the death, injury to any person and the loss or damage to any property which may arise out of or in consequence of the execution of any obligations in terms of this agreement;
- 3.5.2 at its expense take out and keep in force in respect of the indemnity given by it in terms of this agreement a public liability insurance policy providing cover with a limit of not less than R 3 000 000-00 (three million rand) for any one occurrence of an insured peril in any year and unlimited as to cumulative amount in respect of more than one such occurrence in any year;

3.6 SUBCONTRACTING

It is recorded that:

- 3.6.1 The Supplier will be entitled to appoint suitably qualified subcontractors who satisfy the eligibility criteria applicable to the award of the contract to perform all or any of its obligations arising from this Agreement;

- 3.6.2 No sub-contract can create contractual relations between any subcontractor and End- User;
- 3.6.3 The Supplier shall be responsible for all the acts, defaults and negligence of its subcontractors and their experts, agents or employees as if they were acts, defaults or negligence of the Supplier shall not be absolved from its responsibility from under this clause on the basis that such person was acting outside the scope of its engagement by The Supplier.
- 3.6.4 The Supplier will provide the End user with a list (regularly updated for the duration of this agreement) of all the subcontractors that it intends using to perform all or any of its functions in terms of this agreement.

3.7 **CONFIDENTIALITY**

- 3.7.1 The Supplier shall treat all documents and information received in connection with this agreement as private and confidential, and shall not, save in so far as may be necessary for the purposes of performance thereof, publish or disclose any particulars without the prior written consent of the End user.

4. **OBLIGATIONS OF THE END - USER**

- 4.1 The End - user undertakes to:
 - 4.1.1 Use the equipment for the purpose that it is intended and in accordance with any reasonable manufacturers' instructions and user manual as to the use thereof;
 - 4.1.2 Keep the equipment in its possession and custody and control at its premises in accordance with the same policies and procedures that the End user applies in respect of its own assets and equipment;
 - 4.1.3 Advise the supplier prior to relocation equipment.
 - 4.1.4 Allow the supplier or its representatives reasonable access to the inspection of the equipment on prior written notice;
 - 4.1.5 Undertakes to ensure that the installation area, access ways, electrical supply and where relevant, the IT configuration of its premises and other equipment or any network are suitable for the installation, passage and electrical/or electronic connection of the equipment when it is delivered for installation and thereafter.

5. **BREACH**

- 5.1 Either party commits a breach of contract where it fails to discharge any of its obligations in terms of this agreement;
- 5.2 Should either party commit a material breach of this agreement ("the defaulting party") and fail to remedy such breach within ten (10) days of written demand from the other party ("the aggrieved party") then the aggrieved party may, in addition to any other rights and remedies that it may have, including the right to claim damages:-
 - 5.2.1 Claim specific performance ;

- 5.2.2 or Terminate this agreement, such termination to be effective immediately upon receipt by the defaulting party of written notice to that effect
- 5.3 In any case where the End – User is entitled to damages, then the End-user may claim such damages from the Supplier;
- 5.4 This agreement shall automatically and without notice terminate upon occurrence of the following events:
- 5.4.1 a receiver, liquidator or administrator is appointed over any of the property or assets of that the Supplier;
- 5.4.2 the Supplier makes any voluntary arrangement with its creditors by reason of financial difficulty or becomes subject to an administration order, or provisional or final liquidation or insolvency order;
- 5.4.3 the Supplier goes into liquidation or is declared insolvent;
- or
- 5.4.4 that the Supplier ceases, or threatens to cease, to carry on business.

6. PAYMENT

The End – User shall pay the Supplier:

- 6.1 the rental applicable to the contract at the time of signing this agreement which rental shall be fixed for the entire initial rental period of 36 months. In the event of the extension of the contract, the rental shall reduce by 75% of the original rental The first Rental Charge shall be paid after the Commencement Date of the Agreement, within 30 days of the date of the original copy of statement or tax invoice to the Enduser and shall thereafter be payable monthly in arrear within 30 days of the last day of the month in which The Supplier delivers an original copy of statement and tax invoice to the End-user.
- 6.2 Copy Charges, applicable on the contract at the time of signing this Agreement will apply and would thereafter be adjusted on the thirteenth month and twenty-fifth month of the contract period.
- 6.3 The first of the Copy Charges shall be paid within thirty (30) days in which the original copy of statement and tax invoice in respect thereof is rendered, and shall thereafter be payable monthly in arrears on the first day of the month following the month in respect whereof the Copy Charge has arisen or within 30 days of the last day of the month in which the original copy of statement and tax invoice is delivered to the End-user, whichever is the later.
- 6.4 Payment shall be paid by electronic means into bank account :
Name :
Bank :
Branch :
Account number :
- 6.5 No other charges other than those set out herein will be payable for any other service rendered unless specifically agreed to in writing by the parties

7. NOTICES AND DOMICILIA

7.1 The parties select as their respective domicile citandi et executandi the following addresses:

7.1.1 End User
Physical address
Postal Address
Telephone No.
Fax No.
Email
Contact person

7.1.2 Supplier
Physical address
Postal Address
Telephone no
Fax No
Email
Contact person

Or such other address, telefax or telephone number as may be substituted by notice as herein required

7.2 Any notice addressed to a party at its physical or postal address shall be sent by prepaid registered post or delivered by hand or sent by telefax.

7.3 Any notice shall be deemed to have been given:

7.3.1 if posted 14 calendar days after the date of posting;

7.3.2 if hand delivered, on the day of the delivery; The parties may communicate by electronic means.

7.4.3 if sent by telefax, on the date and time of sending, which telefax, is evidenced by a fax confirmation print out.

8. GENERAL

8.1 **Whole Agreement.** This Agreement constitutes the entire Agreement between the Parties in respect of the subject matter hereof and neither Party shall be bound by any undertakings, representations, warranties or promises not recorded in this Agreement.

8.2 **No Variation.** This agreement together with General Conditions of Contract, Special Conditions of the contract and all Standard Bidding Documents constitutes the entire agreement between the parties. No variation or consensual cancellation of this Agreement and no addition to this Agreement shall be of any force or effect unless reduced to writing and signed by the Parties or their duly authorised representatives.

8.3 **Waiver.** No waiver of any of the terms and conditions of this Agreement will be binding or effectual for any purpose unless expressed in writing and signed by the Party hereto giving the same, and any such waiver will be

effective only in the specific instance and for the purpose given. No failure or delay on the part of either Party hereto in exercising any right, power or privilege hereunder will operate as a waiver thereof, nor will any single or partial exercise of any right, power or privilege preclude any other or further exercise thereof or the exercise of any other right, power or privilege.

- 8.4 **Severability.** Should any of the terms and conditions of this Agreement be held to be invalid, unlawful or unenforceable, such terms and conditions will be severable from the remaining terms and conditions which will continue to be valid and enforceable. If any term or condition held to be invalid is capable of amendment to render it valid, the Parties agree to negotiate an amendment to remove the invalidity.
- 8.5 **Applicable Law.** This Agreement will be governed by and construed in accordance with the law of the Republic of South Africa and all disputes, actions and other matters relating thereto will be determined in accordance with such law.
- 8.6 **Jurisdiction.** The Parties hereto hereby consent and submit to the jurisdiction of such High Court of South Africa, in any dispute arising from or in connection with this Agreement.
- 8.7 **Survival.** Notwithstanding termination of this Agreement, any clause which, from the context, contemplates ongoing rights and obligations of the parties, shall survive such termination and continue to be of full force and effect.

SIGNED AT _____ ON THIS _____ DAY OF _____ 2023.

END USER

WITNESSES

1. _____

2. _____

SIGNED AT _____ ON THIS _____ DAY OF _____ 2023.

SUPPLIER

WITNESSES

1. _____

2. _____

Annexure A**GOVERNMENT PROCUREMENT
GENERAL CONDITIONS OF CONTRACT
July 2010****NOTES**

The purpose of this document is to:

- (i) Draw special attention to certain general conditions applicable to government bids, contracts and orders; and
- (ii) To ensure that clients be familiar with regard to the rights and obligations of all parties involved in doing business with government.

In this document words in the singular also mean in the plural and vice versa and words in the masculine also mean in the feminine and neuter.

- The General Conditions of Contract will form part of all bid documents and may not be amended.
- Special Conditions of Contract (SCC) relevant to a specific bid, should be compiled separately for every bid (if applicable) and will supplement the General Conditions of Contract. Whenever there is a conflict, the provisions in the SCC shall prevail.

TABLE OF CLAUSES

1. Definitions
2. Application
3. General
4. Standards
5. Use of contract documents and information; inspection
6. Patent rights
7. Performance security
8. Inspections, tests and analysis
9. Packing
10. Delivery and documents
11. Insurance
12. Transportation
13. Incidental services
14. Spare parts
15. Warranty
16. Payment
17. Prices
18. Contract amendments
19. Assignment
20. Subcontracts
21. Delays in the supplier's performance
22. Penalties
23. Termination for default
24. Dumping and countervailing duties
25. Force Majeure
26. Termination for insolvency
27. Settlement of disputes
28. Limitation of liability
29. Governing language
30. Applicable law
31. Notices
32. Taxes and duties
33. National Industrial Participation Programme (NIPP)
34. Prohibition of restrictive practices

General Conditions of Contract

1. Definitions

1. The following terms shall be interpreted as indicated:
 - 1.1 “Closing time” means the date and hour specified in the bidding documents for the receipt of bids.
 - 1.2 “Contract” means the written agreement entered into between the purchaser and the supplier, as recorded in the contract form signed by the parties, including all attachments and appendices thereto and all documents incorporated by reference therein.
 - 1.3 “Contract price” means the price payable to the supplier under the contract for the full and proper performance of his contractual obligations.
 - 1.4 “Corrupt practice” means the offering, giving, receiving, or soliciting of any thing of value to influence the action of a public official in the procurement process or in contract execution.
 - 1.5 "Countervailing duties" are imposed in cases where an enterprise abroad is subsidized by its government and encouraged to market its products internationally.
 - 1.6 “Country of origin” means the place where the goods were mined, grown or produced or from which the services are supplied. Goods are produced when, through manufacturing, processing or substantial and major assembly of components, a commercially recognized new product results that is substantially different in basic characteristics or in purpose or utility from its components.
 - 1.7 “Day” means calendar day.
 - 1.8 “Delivery” means delivery in compliance of the conditions of the contract or order.
 - 1.9 “Delivery ex stock” means immediate delivery directly from stock actually on hand.
 - 1.10 “Delivery into consignees store or to his site” means delivered and unloaded in the specified store or depot or on the specified site in compliance with the conditions of the contract or order, the supplier bearing all risks and charges involved until the supplies are so delivered and a valid receipt is obtained.
 - 1.11 "Dumping" occurs when a private enterprise abroad market its goods on own initiative in the RSA at lower prices than that of the country of origin and which have the potential to harm the local industries in the RSA.

- 1.12 "Force majeure" means an event beyond the control of the supplier and not involving the supplier's fault or negligence and not foreseeable. Such events may include, but is not restricted to, acts of the purchaser in its sovereign capacity, wars or revolutions, fires, floods, epidemics, quarantine restrictions and freight embargoes.
- 1.13 "Fraudulent practice" means a misrepresentation of facts in order to influence a procurement process or the execution of a contract to the detriment of any bidder, and includes collusive practice among bidders (prior to or after bid submission) designed to establish bid prices at artificial non-competitive levels and to deprive the bidder of the benefits of free and open competition.
- 1.14 "GCC" means the General Conditions of Contract.
- 1.15 "Goods" means all of the equipment, machinery, and/or other materials that the supplier is required to supply to the purchaser under the contract.
- 1.16 "Imported content" means that portion of the bidding price represented by the cost of components, parts or materials which have been or are still to be imported (whether by the supplier or his subcontractors) and which costs are inclusive of the costs abroad, plus freight and other direct importation costs such as landing costs, dock dues, import duty, sales duty or other similar tax or duty at the South African place of entry as well as transportation and handling charges to the factory in the Republic where the supplies covered by the bid will be manufactured.
- 1.17 "Local content" means that portion of the bidding price which is not included in the imported content provided that local manufacture does take place.
- 1.18 "Manufacture" means the production of products in a factory using labour, materials, components and machinery and includes other related value-adding activities.
- 1.19 "Order" means an official written order issued for the supply of goods or works or the rendering of a service.
- 1.20 "Project site," where applicable, means the place indicated in bidding documents.
- 1.21 "Purchaser" means the organization purchasing the goods.
- 1.22 "Republic" means the Republic of South Africa.
- 1.23 "SCC" means the Special Conditions of Contract.
- 1.24 "Services" means those functional services ancillary to the supply of the goods, such as transportation and any other incidental services, such as installation, commissioning, provision of technical assistance, training, catering, gardening, security, maintenance and other such obligations of the supplier covered under the contract.

- 1.25 “Written” or “in writing” means handwritten in ink or any form of electronic or mechanical writing.
- 2. Application**
- 2.1 These general conditions are applicable to all bids, contracts and orders including bids for functional and professional services, sales, hiring, letting and the granting or acquiring of rights, but excluding immovable property, unless otherwise indicated in the bidding documents.
- 2.2 Where applicable, special conditions of contract are also laid down to cover specific supplies, services or works.
- 2.3 Where such special conditions of contract are in conflict with these general conditions, the special conditions shall apply.
- 3. General**
- 3.1 Unless otherwise indicated in the bidding documents, the purchaser shall not be liable for any expense incurred in the preparation and submission of a bid. Where applicable a non-refundable fee for documents may be charged.
- 3.2 With certain exceptions, invitations to bid are only published in the Government Tender Bulletin. The Government Tender Bulletin may be obtained directly from the Government Printer, Private Bag X85, Pretoria 0001, or accessed electronically from www.treasury.gov.za
- 4. Standards**
- 4.1 The goods supplied shall conform to the standards mentioned in the bidding documents and specifications.
- 5. Use of contract documents and information; inspection.**
- 5.1 The supplier shall not, without the purchaser’s prior written consent, disclose the contract, or any provision thereof, or any specification, plan, drawing, pattern, sample, or information furnished by or on behalf of the purchaser in connection therewith, to any person other than a person employed by the supplier in the performance of the contract. Disclosure to any such employed person shall be made in confidence and shall extend only so far as may be necessary for purposes of such performance.
- 5.2 The supplier shall not, without the purchaser’s prior written consent, make use of any document or information mentioned in GCC clause 5.1 except for purposes of performing the contract.
- 5.3 Any document, other than the contract itself mentioned in GCC clause 5.1 shall remain the property of the purchaser and shall be returned (all copies) to the purchaser on completion of the supplier’s performance under the contract if so required by the purchaser.
- 5.4 The supplier shall permit the purchaser to inspect the supplier’s records relating to the performance of the supplier and to have them audited by auditors appointed by the purchaser, if so required by the purchaser.
- 6. Patent rights**
- 6.1 The supplier shall indemnify the purchaser against all third-party claims of infringement of patent, trademark, or industrial design rights arising from use of the goods or any part thereof by the purchaser.
- 7. Performance**
- 7.1 Within thirty (30) days of receipt of the notification of contract award,

security

the successful bidder shall furnish to the purchaser the performance security of the amount specified in SCC.

- 7.2 The proceeds of the performance security shall be payable to the purchaser as compensation for any loss resulting from the supplier's failure to complete his obligations under the contract.
- 7.3 The performance security shall be denominated in the currency of the contract, or in a freely convertible currency acceptable to the purchaser and shall be in one of the following forms:
 - (a) a bank guarantee or an irrevocable letter of credit issued by a reputable bank located in the purchaser's country or abroad, acceptable to the purchaser, in the form provided in the bidding documents or another form acceptable to the purchaser; or
 - (b) a cashier's or certified cheque
- 7.4 The performance security will be discharged by the purchaser and returned to the supplier not later than thirty (30) days following the date of completion of the supplier's performance obligations under the contract, including any warranty obligations, unless otherwise specified in SCC.

8. Inspections, tests and analyses

- 8.1 All pre-bidding testing will be for the account of the bidder.
- 8.2 If it is a bid condition that supplies to be produced or services to be rendered should at any stage during production or execution or on completion be subject to inspection, the premises of the bidder or contractor shall be open, at all reasonable hours, for inspection by a representative of the Department or an organization acting on behalf of the Department.
- 8.3 If there are no inspection requirements indicated in the bidding documents and no mention is made in the contract, but during the contract period it is decided that inspections shall be carried out, the purchaser shall itself make the necessary arrangements, including payment arrangements with the testing authority concerned.
- 8.4 If the inspections, tests and analyses referred to in clauses 8.2 and 8.3 show the supplies to be in accordance with the contract requirements, the cost of the inspections, tests and analyses shall be defrayed by the purchaser.
- 8.5 Where the supplies or services referred to in clauses 8.2 and 8.3 do not comply with the contract requirements, irrespective of whether such supplies or services are accepted or not, the cost in connection with these inspections, tests or analyses shall be defrayed by the supplier.
- 8.6 Supplies and services which are referred to in clauses 8.2 and 8.3 and which do not comply with the contract requirements may be rejected.
- 8.7 Any contract supplies may on or after delivery be inspected, tested or analyzed and may be rejected if found not to comply with the requirements of the contract. Such rejected supplies shall be held at the

cost and risk of the supplier who shall, when called upon, remove them immediately at his own cost and forthwith substitute them with supplies which do comply with the requirements of the contract. Failing such removal the rejected supplies shall be returned at the suppliers cost and risk. Should the supplier fail to provide the substitute supplies forthwith, the purchaser may, without giving the supplier further opportunity to substitute the rejected supplies, purchase such supplies as may be necessary at the expense of the supplier.

- 8.8 The provisions of clauses 8.4 to 8.7 shall not prejudice the right of the purchaser to cancel the contract on account of a breach of the conditions thereof, or to act in terms of Clause 23 of GCC.

9. Packing

- 9.1 The supplier shall provide such packing of the goods as is required to prevent their damage or deterioration during transit to their final destination, as indicated in the contract. The packing shall be sufficient to withstand, without limitation, rough handling during transit and exposure to extreme temperatures, salt and precipitation during transit, and open storage. Packing, case size and weights shall take into consideration, where appropriate, the remoteness of the goods' final destination and the absence of heavy handling facilities at all points in transit.
- 9.2 The packing, marking, and documentation within and outside the packages shall comply strictly with such special requirements as shall be expressly provided for in the contract, including additional requirements, if any, specified in SCC, and in any subsequent instructions ordered by the purchaser.

10. Delivery and documents

- 10.1 Delivery of the goods shall be made by the supplier in accordance with the terms specified in the contract. The details of shipping and/or other documents to be furnished by the supplier are specified in SCC.
- 10.2 Documents to be submitted by the supplier are specified in SCC.

11. Insurance

- 11.1 The goods supplied under the contract shall be fully insured in a freely convertible currency against loss or damage incidental to manufacture or acquisition, transportation, storage and delivery in the manner specified in the SCC.

12. Transportation

- 12.1 Should a price other than an all-inclusive delivered price be required, this shall be specified in the SCC.

13. Incidental services

- 13.1 The supplier may be required to provide any or all of the following services, including additional services, if any, specified in SCC:
- (a) performance or supervision of on-site assembly and/or commissioning of the supplied goods;
 - (b) furnishing of tools required for assembly and/or maintenance of the supplied goods;
 - (c) furnishing of a detailed operations and maintenance manual for each appropriate unit of the supplied goods;
 - (d) performance or supervision or maintenance and/or repair of the supplied goods, for a period of time agreed by the parties,

- provided that this service shall not relieve the supplier of any warranty obligations under this contract; and
- (e) training of the purchaser's personnel, at the supplier's plant and/or on-site, in assembly, start-up, operation, maintenance, and/or repair of the supplied goods.

13.2 Prices charged by the supplier for incidental services, if not included in the contract price for the goods, shall be agreed upon in advance by the parties and shall not exceed the prevailing rates charged to other parties by the supplier for similar services.

14. Spare parts

14.1 As specified in SCC, the supplier may be required to provide any or all of the following materials, notifications, and information pertaining to spare parts manufactured or distributed by the supplier:

- (a) such spare parts as the purchaser may elect to purchase from the supplier, provided that this election shall not relieve the supplier of any warranty obligations under the contract; and
- (b) in the event of termination of production of the spare parts:
 - (i) Advance notification to the purchaser of the pending termination, in sufficient time to permit the purchaser to procure needed requirements; and
 - (ii) following such termination, furnishing at no cost to the purchaser, the blueprints, drawings, and specifications of the spare parts, if requested.

15. Warranty

15.1 The supplier warrants that the goods supplied under the contract are new, unused, of the most recent or current models, and that they incorporate all recent improvements in design and materials unless provided otherwise in the contract. The supplier further warrants that all goods supplied under this contract shall have no defect, arising from design, materials, or workmanship (except when the design and/or material is required by the purchaser's specifications) or from any act or omission of the supplier, that may develop under normal use of the supplied goods in the conditions prevailing in the country of final destination.

15.2 This warranty shall remain valid for twelve (12) months after the goods, or any portion thereof as the case may be, have been delivered to and accepted at the final destination indicated in the contract, or for eighteen (18) months after the date of shipment from the port or place of loading in the source country, whichever period concludes earlier, unless specified otherwise in SCC.

15.3 The purchaser shall promptly notify the supplier in writing of any claims arising under this warranty.

15.4 Upon receipt of such notice, the supplier shall, within the period specified in SCC and with all reasonable speed, repair or replace the defective goods or parts thereof, without costs to the purchaser.

15.5 If the supplier, having been notified, fails to remedy the defect(s) within the period specified in SCC, the purchaser may proceed to take such remedial action as may be necessary, at the supplier's risk and expense and without prejudice to any other rights which the purchaser

may have against the supplier under the contract.

16. Payment

- 16.1 The method and conditions of payment to be made to the supplier under this contract shall be specified in SCC.
- 16.2 The supplier shall furnish the purchaser with an invoice accompanied by a copy of the delivery note and upon fulfillment of other obligations stipulated in the contract.
- 16.3 Payments shall be made promptly by the purchaser, but in no case later than thirty (30) days after submission of an invoice or claim by the supplier.
- 16.4 Payment will be made in Rand unless otherwise stipulated in SCC.

17. Prices

- 17.1 Prices charged by the supplier for goods delivered and services performed under the contract shall not vary from the prices quoted by the supplier in his bid, with the exception of any price adjustments authorized in SCC or in the purchaser's request for bid validity extension, as the case may be.

18. Contract amendments

- 18.1 No variation in or modification of the terms of the contract shall be made except by written amendment signed by the parties concerned.

19. Assignment

- 19.1 The supplier shall not assign, in whole or in part, its obligations to perform under the contract, except with the purchaser's prior written consent.

20. Subcontracts

- 20.1 The supplier shall notify the purchaser in writing of all subcontracts awarded under this contracts if not already specified in the bid. Such notification, in the original bid or later, shall not relieve the supplier from any liability or obligation under the contract.

21. Delays in the supplier's performance

- 21.1 Delivery of the goods and performance of services shall be made by the supplier in accordance with the time schedule prescribed by the purchaser in the contract.
- 21.2 If at any time during performance of the contract, the supplier or its subcontractor(s) should encounter conditions impeding timely delivery of the goods and performance of services, the supplier shall promptly notify the purchaser in writing of the fact of the delay, its likely duration and its cause(s). As soon as practicable after receipt of the supplier's notice, the purchaser shall evaluate the situation and may at his discretion extend the supplier's time for performance, with or without the imposition of penalties, in which case the extension shall be ratified by the parties by amendment of contract.
- 21.3 No provision in a contract shall be deemed to prohibit the obtaining of supplies or services from a national department, provincial department, or a local authority.
- 21.4 The right is reserved to procure outside of the contract small quantities or to have minor essential services executed if an emergency arises, the supplier's point of supply is not situated at or near the place where the supplies are required, or the supplier's services are not readily

available.

21.5 Except as provided under GCC Clause 25, a delay by the supplier in the performance of its delivery obligations shall render the supplier liable to the imposition of penalties, pursuant to GCC Clause 22, unless an extension of time is agreed upon pursuant to GCC Clause 21.2 without the application of penalties.

21.6 Upon any delay beyond the delivery period in the case of a supplies contract, the purchaser shall, without canceling the contract, be entitled to purchase supplies of a similar quality and up to the same quantity in substitution of the goods not supplied in conformity with the contract and to return any goods delivered later at the supplier's expense and risk, or to cancel the contract and buy such goods as may be required to complete the contract and without prejudice to his other rights, be entitled to claim damages from the supplier.

22. Penalties

22.1 Subject to GCC Clause 25, if the supplier fails to deliver any or all of the goods or to perform the services within the period(s) specified in the contract, the purchaser shall, without prejudice to its other remedies under the contract, deduct from the contract price, as a penalty, a sum calculated on the delivered price of the delayed goods or unperformed services using the current prime interest rate calculated for each day of the delay until actual delivery or performance. The purchaser may also consider termination of the contract pursuant to GCC Clause 23.

23. Termination for default

23.1 The purchaser, without prejudice to any other remedy for breach of contract, by written notice of default sent to the supplier, may terminate this contract in whole or in part:

- (a) if the supplier fails to deliver any or all of the goods within the period(s) specified in the contract, or within any extension thereof granted by the purchaser pursuant to GCC Clause 21.2;
- (b) if the Supplier fails to perform any other obligation(s) under the contract; or
- (c) if the supplier, in the judgment of the purchaser, has engaged in corrupt or fraudulent practices in competing for or in executing the contract.

23.2 In the event the purchaser terminates the contract in whole or in part, the purchaser may procure, upon such terms and in such manner as it deems appropriate, goods, works or services similar to those undelivered, and the supplier shall be liable to the purchaser for any excess costs for such similar goods, works or services. However, the supplier shall continue performance of the contract to the extent not terminated.

23.3 Where the purchaser terminates the contract in whole or in part, the purchaser may decide to impose a restriction penalty on the supplier by prohibiting such supplier from doing business with the public sector for a period not exceeding 10 years.

23.4 If a purchaser intends imposing a restriction on a supplier or any person associated with the supplier, the supplier will be allowed a time period of not more than fourteen (14) days to provide reasons why the

envisaged restriction should not be imposed. Should the supplier fail to respond within the stipulated fourteen (14) days the purchaser may regard the intended penalty as not objected against and may impose it on the supplier.

23.5 Any restriction imposed on any person by the Accounting Officer / Authority will, at the discretion of the Accounting Officer / Authority, also be applicable to any other enterprise or any partner, manager, director or other person who wholly or partly exercises or exercised or may exercise control over the enterprise of the first-mentioned person, and with which enterprise or person the first-mentioned person, is or was in the opinion of the Accounting Officer / Authority actively associated.

23.6 If a restriction is imposed, the purchaser must, within five (5) working days of such imposition, furnish the National Treasury, with the following information:

- (i) the name and address of the supplier and / or person restricted by the purchaser;
- (ii) the date of commencement of the restriction
- (iii) the period of restriction; and
- (iv) the reasons for the restriction.

These details will be loaded in the National Treasury's central database of suppliers or persons prohibited from doing business with the public sector.

23.7 If a court of law convicts a person of an offence as contemplated in sections 12 or 13 of the Prevention and Combating of Corrupt Activities Act, No. 12 of 2004, the court may also rule that such person's name be endorsed on the Register for Tender Defaulters. When a person's name has been endorsed on the Register, the person will be prohibited from doing business with the public sector for a period not less than five years and not more than 10 years. The National Treasury is empowered to determine the period of restriction and each case will be dealt with on its own merits. According to section 32 of the Act the Register must be open to the public. The Register can be perused on the National Treasury website.

24. Anti-dumping and countervailing duties and rights

24.1 When, after the date of bid, provisional payments are required, or anti-dumping or countervailing duties are imposed, or the amount of a provisional payment or anti-dumping or countervailing right is increased in respect of any dumped or subsidized import, the State is not liable for any amount so required or imposed, or for the amount of any such increase. When, after the said date, such a provisional payment is no longer required or any such anti-dumping or countervailing right is abolished, or where the amount of such provisional payment or any such right is reduced, any such favourable difference shall on demand be paid forthwith by the contractor to the State or the State may deduct such amounts from moneys (if any) which may otherwise be due to the contractor in regard to supplies or services which he delivered or rendered, or is to deliver or render in terms of the contract or any other contract or any other amount which may be due to him

25. Force Majeure

- 25.1 Notwithstanding the provisions of GCC Clauses 22 and 23, the supplier shall not be liable for forfeiture of its performance security, damages, or termination for default if and to the extent that his delay in performance or other failure to perform his obligations under the contract is the result of an event of force majeure.
- 25.2 If a force majeure situation arises, the supplier shall promptly notify the purchaser in writing of such condition and the cause thereof. Unless otherwise directed by the purchaser in writing, the supplier shall continue to perform its obligations under the contract as far as is reasonably practical, and shall seek all reasonable alternative means for performance not prevented by the force majeure event.

26. Termination for insolvency

- 26.1 The purchaser may at any time terminate the contract by giving written notice to the supplier if the supplier becomes bankrupt or otherwise insolvent. In this event, termination will be without compensation to the supplier, provided that such termination will not prejudice or affect any right of action or remedy which has accrued or will accrue thereafter to the purchaser.

27. Settlement of Disputes

- 27.1 If any dispute or difference of any kind whatsoever arises between the purchaser and the supplier in connection with or arising out of the contract, the parties shall make every effort to resolve amicably such dispute or difference by mutual consultation.
- 27.2 If, after thirty (30) days, the parties have failed to resolve their dispute or difference by such mutual consultation, then either the purchaser or the supplier may give notice to the other party of his intention to commence with mediation. No mediation in respect of this matter may be commenced unless such notice is given to the other party.
- 27.3 Should it not be possible to settle a dispute by means of mediation, it may be settled in a South African court of law.
- 27.4 Mediation proceedings shall be conducted in accordance with the rules of procedure specified in the SCC.
- 27.5 Notwithstanding any reference to mediation and/or court proceedings herein,
- (a) the parties shall continue to perform their respective obligations under the contract unless they otherwise agree; and
 - (b) the purchaser shall pay the supplier any monies due the supplier.

28. Limitation of liability

- 28.1 Except in cases of criminal negligence or willful misconduct, and in the case of infringement pursuant to Clause 6;
- (a) the supplier shall not be liable to the purchaser, whether in contract, tort, or otherwise, for any indirect or consequential loss or damage, loss of use, loss of production, or loss of profits or interest costs, provided that this exclusion shall not apply to any obligation of the supplier to pay penalties and/or damages to the purchaser; and

- (b) the aggregate liability of the supplier to the purchaser, whether under the contract, in tort or otherwise, shall not exceed the total contract price, provided that this limitation shall not apply to the cost of repairing or replacing defective equipment.
- 29. Governing language** 29.1 The contract shall be written in English. All correspondence and other documents pertaining to the contract that is exchanged by the parties shall also be written in English.
- 30. Applicable law** 30.1 The contract shall be interpreted in accordance with South African laws, unless otherwise specified in SCC.
- 31. Notices** 31.1 Every written acceptance of a bid shall be posted to the supplier concerned by registered or certified mail and any other notice to him shall be posted by ordinary mail to the address furnished in his bid or to the address notified later by him in writing and such posting shall be deemed to be proper service of such notice
- 31.2 The time mentioned in the contract documents for performing any act after such aforesaid notice has been given, shall be reckoned from the date of posting of such notice.
- 32. Taxes and duties** 32.1 A foreign supplier shall be entirely responsible for all taxes, stamp duties, license fees, and other such levies imposed outside the purchaser's country.
- 32.2 A local supplier shall be entirely responsible for all taxes, duties, license fees, etc., incurred until delivery of the contracted goods to the purchaser.
- 32.3 No contract shall be concluded with any bidder whose tax matters are not in order. Prior to the award of a bid the Department must be in possession of a tax clearance certificate, submitted by the bidder. This certificate must be an original issued by the South African Revenue Services.
- 33. National Industrial Participation Programme (NIP)** 33.1 The NIP Programme administered by the Department of Trade and Industry shall be applicable to all contracts that are subject to the NIP obligation.
- 34. Prohibition of Restrictive practices** 34.1 In terms of section 4 (1) (b) (iii) of the Competition Act No. 89 of 1998, as amended, an agreement between, or concerted practice by, firms, or a decision by an association of firms, is prohibited if it is between parties in a horizontal relationship and if a bidder (s) is / are or a contractor(s) was / were involved in collusive bidding (or bid rigging).
- 34.2 If a bidder(s) or contractor(s), based on reasonable grounds or evidence obtained by the purchaser, has / have engaged in the restrictive practice referred to above, the purchaser may refer the matter to the Competition Commission for investigation and possible imposition of administrative penalties as contemplated in the Competition Act No. 89 of 1998.

- 34.3 If a bidder(s) or contractor(s), has / have been found guilty by the Competition Commission of the restrictive practice referred to above, the purchaser may, in addition and without prejudice to any other remedy provided for, invalidate the bid(s) for such item(s) offered, and / or terminate the contract in whole or part, and / or restrict the bidder(s) or contractor(s) from conducting business with the public sector for a period not exceeding ten (10) years and / or claim damages from the bidder(s) or contractor(s) concerned.

Js General Conditions of Contract (revised July 2010)