



QUESTIONS AND ANSWERS

APPOINTMENT OF A SERVICE PROVIDER TO SUPPLY, DELIVER, INSTALL AND COMMISSION OF A NEW LASER DIFFRACTION PARTICLE SIZE ANALYZER INCLUDING SERVICE AND MAINTENANCE SUPPORT FOR A PERIOD OF THREE (3) YEARS.

No	Question	Answer
1	I see that you indicated a sizing range of 0.01 – 3000um. Do you need to go down to 0.01um, or would it be a nice to have?	We need to go down to 0.01um
2	Which kind of samples will be measured on this unit?	Mainly solid samples (various matrices)
3	Which kind of samples would fall between 0.01 - 0.1um?	Various sample matrices – We are a research laboratory, we develop nanomaterials in various research projects.

Board Members:

Mr K Dintwe (Chairperson) | Mr X Mvinjelwa (Deputy Chairperson) | Mr M Mabuzza (CEO) | Adv. N Maake | Mr M Malindisa | Dr M Matlala | Dr S Mngadi | Mr A Moatshe | Dr M Muofhe | Ms T Njozela | Ms N Nxumalo | Dr M Singh |

Acting Company Secretary: Ms N Cele