

Transversal Contracting
ENQUIRIES: Ngwato Nkuna
Email: ngwato.nkuna@treasury.gov.za
Ref RT28-1-2024

QUESTIONS AND ANSWERS

RT28-1-2024: SUPPLY AND DELIVERY OF STERILISATION RELATED TO THE STATE FOR THE PERIOD OF ENDING 31 MAY 2027.

NO	QUESTIONS	ANSWERS
1.	What type of Test Report is required?	Only Test Reports from SANAS accredited institutions will be permitted, failure to do so will lead to a disqualification. Over and above that, all components or elements should be measured or evaluated and the results must indicate passed. A Test Report that has partial elements or components measured or evaluated will not be accepted.
2.	Can the bidder be allowed to submit their bid documents after the closing date?	No. Bid document submission will NOT be allowed after the closing date and time of the bid.

NO	QUESTIONS	ANSWERS
3.	Can we do one submission for multiple products, instead of doing multiple submissions for each item bidding for?	Proposals should be packaged according to the phases as outlined in the Special Conditions of Contract. Bidders must ensure that bid documents are packaged correctly and that there are no loose pages.
4.	Can the SAHPRA certificate be in the name of the 3 rd party that I will be sourcing the products from?	No, the SAHPRA License should be in the name of the bidder and NOT in the name of the 3 rd party.
5.	If during the contract period, certain products are out of stock or for some other reason it is no longer available, are bidders allowed to supply alternative products?	Bidders must maintain sufficient stock to meet demand throughout the duration of the contract and inform the National Treasury at first knowledge of any circumstances that may result in interrupted supply. Should the service provider offer an alternative due to stock availability, the alternative will be evaluated, as it must adhere to the technical specifications. Bidders should read paragraphs 16 and 19 of the Special Conditions of Contract.
6	We wanted to know if the double wraps should be bonded together or not bonded?	The wraps consist of different layers which means that all that layers must be bonded together to form the wrap.

NO	QUESTIONS	ANSWERS
7	Please could you advise me the size of item RT28-1-010, will it be 15mm x 50m or 25mm x 50	Bidders must have the 25mm x 50mm



MULATEDZI RAUDZINGANA

FOR CHIEF DIRECTOR: TRANSVERSAL CONTRACTING

DATE: 10 SEPTEMBER 2024