

	BID CLARIFICATION	Form No: RW SCM 00041 F Revision: 02 Effective Date: February 2020
---	----------------------------	--

Bid Number	RW10393438/22
Clarification Number	01

Enquiries: Johannes Sebothoma
Direct Line: 011 682 0860
E-mail: jsebotho@randwater.co.za
Date: 15/03/2022

**BID NUMBER – RW10393438 – 22 SUPPLY, DELIVERY AND INSTALLATION OF 3 X
AUTO TITRATION SYSTEM AT ZUIKERBOSCH WATER PLANT AND 1 X AUTO TITRATION
AT ZWARKOPJES PUMPING STATION**

Clarification for questions raised by the Suppliers

The pages referenced are the pages of the **RW10393438/22** bid document.

This clarification consists of 01 statements in the following pages.

SIGNATURES

Johannes Sebothoma
Buyer

Lebiso Moloi
Sourcing Manager

Emmanuel Nzilo
Supply Chain Manager

	BID CLARIFICATION	Form No: RW SCM 00041 F Revision: 02 Effective Date: February 2020
---	----------------------------	--

RW10393438 – 22 SUPPLY, DELIVERY AND INSTALLATION OF 3 X AUTO TITRATION SYSTEM AT ZUIKERBOSCH WATER PLANT AND 1 X AUTO TITRATION AT ZWARKOPJES PUMPING STATION

Dear Tenderers,

Please see questions and answers below:

RW10393438 – 22 supply, delivery and installation of 3 x auto titration system at Zuikerbosch water plant and 1 x auto titration at Zwarkopjes pumping station

Please see questions and answers below:

Questions	Responses
1. Regarding the 4 auto titrators that Rand water is looking for, what are the parameters that you are looking to test. If auto sampler is required, please advise how many samples would be required to be tested.	Parameters for the Automated Titration Systems are:simultaneous determination of pH, conductivity, hardness and Total Alkalinity. Number of samples : greater or equal to 16.