



Lead dimensions clarification on Annexure B– GT/GDH/021/2026

Annexure B (Medical Physics), Items 18 to 23 – Lead Bricks Dimensions

Bid number	Description
GT/GDH/021/2026	Supply and Delivery of Radiology Consumables for a Period of Three Years

This clarification must be read with the Minutes of the Compulsory Briefing Session of 02 October 2026).

Lead Bricks Dimensions


Items 18 to 23 dimensions as follows. All items remain brochure-evaluation items, priced per brick (unit of measure: Each).

Summary of dimensions


Item	Shape	Size (Length × Width × Thickness)
18	Rectangular, interlocking	200 × 100 × 50 mm or better
19	Square, interlocking	100 × 100 × 50 mm or better
20	Hexagonal, interlocking	100 mm across flats × 50 mm or better
21	Hexagonal, flat	100 mm across flats × 50 mm or better

Lead brick shapes and dimensions · items 18 to 21 (schematic, not to scale)


Lead brick dimensions for Annexure B items 18 to 21



Item 20 · hexagonal, interlocking
100 mm across flats × 50 mm



Item 21 · hexagonal, flat
100 mm across flats × 50 mm



The drawings are schematic and indicate shape and nominal dimensions only.



Requirements for all lead brick items

The following apply to items 18, 19, 20 and 21:

1. Thickness: 50 mm of solid lead.1. Thickness: 50 mm of solid lead, Thinner bricks will not be accepted as equivalent,
2. Joints: interlocking bricks must stack without straight-through gaps between bricks.
3. Finish: epoxy-coated or painted, so that staff do not handle bare lead.
4. Tolerance: ± 1 mm on all stated dimensions.
5. Brochure: the manufacturer's brochure or technical data sheet must state the shape, dimensions, mass, finish of each brick offered.
6. **Equivalent or better products may be offered in columns F and G of Annexure B.** They must still be at least 50 mm thick

Pricing

- Quote one all-inclusive price per brick (VAT inclusive) for each of items 18 to 21, based on the dimensions.
- Leave items 22 and 23 blank.
- In column K, state the number of bricks per box or pack.
- All other pricing instructions in the bid documents and in the briefing minutes still apply, including the rate of exchange, the cost breakdown and the prohibition on tiered prices.

Ms. Portia Mogomotsi

Mr. Tshilidzi Shandukani