

TO: ALL SERVICE PROVIDERS

ADDENDUM OR ERRATUM

BID NO. SCMU3-25/26- 0047-HO

SUPPLY AND DELIVERY OF N95 MASKS FOR TB HOSPITALS AND OTHER HEALTH FACILITIES IN THE EASTERN CAPE FOR A PERIOD OF 60 MONTHS.REASONS FOR ADDENDUM OR ERRATUM

2ND STAGE MANDATORY REQUIREMENT- EVALUATION CRITERIA (page 10 and 11 on the Bid Document)

- 22.2.1 All the products/masks that will be supplied must be Niosh N95 accredited/ certified and proof therefore as evidence of the Niosh certification must be attached with the bid.
- 22.2.2 A respiratory mask, intended for use in a clinical environment, requires preapproval from the National Regulatory for Compulsory Specifications (NRCS), evidence must be attached.
- 22.2.3 The N95 particulate respirator mask must be fluid resistant (120 mmHG).
- 22.2.4 There must be evidence that the manufacturer is an OEM (Original Equipment Manufacturer) must be attached, in case where a distributor has a letter of support by the OEM, the letter of support must be attached.
- 22.2.5 A letter from the OEM confirming the support of the bidder with specific reference to this bid for its products and services must be attached.
- 22.2.6 The bidders must have offices within the Eastern Cape, evidence must be attached in the form of either a municipal account, affidavit from the Local Council or a valid lease agreement in the name of the bidder (signed by all parties).
- 22.2.7 **A sample of NIOSH approved N95 1860 mask that is offered must be submitted on the bid box (One (1) for regular and One (1) for small).**
- 22.2.8 Invitation to Bid (SBD1) must be completed and signed.
- 22.2.9 Bidders must be a legal entity or partnership (registered), must attach evidence.
- 22.2.10 In the case of a Joint Venture/Consortium, a copy of the agreement entered into by the consortium/JV members must be submitted with the bid. Conditions on JV paragraph 11 should all be observed.

Additional Requirement

22.2.11 Bidders must submit proof of licensing from SOUTH AFRICAN HEALTH PRODUCTS REGULATORY AUTHORITY (SAHPRA)