SCOPE OF WORK: CALIBRATION OF THE MEDICAL EQUIPMENT CONTRACT FOR ESKOM GAUTENG CLUSTER CLINICS (EAL, ERIC, PRINCESS, TSG and RIGGI)

1. Background Information

The contractor is required to perform calibration of the medical equipment **(as per attached detailed list of all equipment's)** situated in Distribution Gauteng Cluster Medical Centres and Wellness facilities. The calibration services of medical equipment are required to be performed on an annual basis and when the need to calibrate the specific medical equipment arises. The tests to be performed must cover but not limited to the minimum checks listed below. The services must include the repairing and replacing of the damaged parts of the medical equipment including any apparatus not mentioned below required for the full functioning of the equipment. The supplier is expected to issue the calibration certificates on completion of the ordered services.

	Description of Equipment's and services
1	**Blood Pressure Gauge** • Calibration of the machine to ensure accuracy of results • Checking leaks • Issue calibration certificate
2	**Aneroid Sphygmomanometers** • Calibration of the machine to ensure accuracy of results • Checking leaks • Issue calibration certificate
3	**Takei T.K.K. 5402 Grip Strength Dynamometer** • Calibration of the machine to ensure accuracy of results • Issue calibration certificate
4	**Takei T.K.K. 5401 Back Strength Dynamometer** • Calibration of the machine to ensure accuracy of results

	• Issue calibration certificate
5	**Digital Height Gauge** • Testing the gauge to ensure accuracy of results • Issue calibration certificate
6	**Weight scales** • Testing the scales, to ensure the required level of accuracy. • Issue of Calibration certificate
7	**Welch Allyn Wall Mounted Diagnostic Set** • Replacement of parts as and when necessary

	Requestors Details	Name	Portia Leshabana
		Designation	Occupational Health Cluster Manager
		Signature	