

1. **Question:** Item 095854800 – Spoon; Excavator, Surgical (Qty 80): no size or configuration is specified. Could you confirm the required size(s), or whether a standard surgical excavator set satisfies this line item.

Response: With reference to the PO long text description; namely, UPC: 755015152014 / ASIN: B00FKMP6ZW, this item refers to a spoon excavator with 1mm working ends on both sides

2. **Question:** Forceps Set (Adult and Pedo): the schedule references an Adult and Pedo forceps set without specifying the individual forceps types/patterns required. Could you provide the specific types expected, or confirm that a standard adult/paediatric extraction forceps set at supplier discretion is acceptable?

Response: to proceed as per the supplier's discretion.

3. **Question:** Items 095854861 (Coupland Elevators) and 095854862 (Root Elevator), Qty 40 each: these are listed as single line items, but Coupland elevators and root elevators are typically supplied in size or side variants (for example 1mm/2mm/3mm, or left/right/straight). Could you confirm which variants are required, and whether the quantity of 40 applies to the item as a set or per individual size/side?

Response: Item 095854861: Provided in a kit containing sizes in 1mm, 2mm and 3mm (1 of each). The total number of kits specified on the PO should be supplied.
Item 095854862: Provided in a kit containing left, right and straight (1 of each). The total number of kits specified on the PO should be supplied.

4. **Question:** Scalpel Blade Kit (6 sizes): the schedule calls for a 6-size scalpel blade kit. None of the suppliers we have approached to date sell this as a single combined product, blades are typically supplied per size. Could you confirm whether individual blade sizes supplied separately, to make up the 6 sizes and required quantity, would satisfy this requirement.

Response: Scalpel blade kit (095854876): These are available in kits containing the scalpel blade sizes on the PO (i.e., no.10; no.11; no.12; no.12B; no.15; no. 15C). The total number of kits specified on the PO should be supplied.