

Technical Evaluation Criteria Supply of FLUKE 177 True RMS Multi-meter for NTCSA Northern Grid

CLAUSE	DESCRIPTION	WEIGHT (%)	COMMENT(S)
1	POWER SUPPLY	35%	
1.1	voltage: 1000V DC/AC		
1.2	current: 0.02-10A DC/AC continuously		
2	True RMS: Yes	10%	
3	WEIGHT	10%	
3.1	420g		
4	DIMENSIONS	10%	
4.1	43 x 90 x 185 mm		
5	MODEL	10%	
5.1	Model: Fluke 177		
6	Safety Specification	25%	
6.1	Overvoltage: EN 61010-1 to 1000 V CAT III EN 61010-1 to 600 V CAT IV		
	Total	100%	
	Threshold	100%	

Returnable to be: Photos of items need to be supplied or a write up of the product with its technical specification.

	Compiled By	Reviewed By	Supported By
Name & Surname	Rudzanii Mmakgaha	Ellan Phaahla	Ellan Phaahla
Designation	Snr Tech (acting sup)	Middle Manager	Middle Manager
Unique No	4543059	1291274	1291274
Signature		 10.06.2025	 10.06.2025

